



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : ketankumar

Run Number: LB119059

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	03/09/2022	11:35
2	CAL2	1	Water	NA	NA	20.4	7.00	03/09/2022	11:36
3	CAL3	1	Water	NA	NA	20.4	10.03	03/09/2022	11:37
4	ICV	1	Water	NA	NA	20.3	7.00	03/09/2022	11:39
5	CCV1	1	Water	NA	NA	20.3	2.01	03/09/2022	11:40
6	N1859-01	1	Water	NA	NA	20.2	4.77	03/09/2022	11:42
7	N1859-01DUP	1	Water	NA	NA	20.3	4.78	03/09/2022	11:43
8	N1859-02	1	Water	NA	NA	20.5	5.10	03/09/2022	11:45
9	N1859-03	1	Water	NA	NA	20.6	5.19	03/09/2022	11:46
10	CCV2	1	Water	NA	NA	20.3	12.02	03/09/2022	11:48

2B 119059

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N1859 WorkList ID : 157557 Department : Wet-Chemistry Date : 03-09-2022 11:26:50

Due Date	Matrix	Sample	Test	Raw Sample			Collect Date	Method
				Preservative	Customer	Storage Location		
03/09/2022	Water	N1859-01	pH	Cool 4 deg C	STER07	N22 CELL-A	03/09/2022	9040C
03/09/2022	Water	N1859-02	pH	Cool 4 deg C	STER07	N22 CELL-B/C	03/09/2022	9040C
03/09/2022	Water	N1859-03	pH	Cool 4 deg C	STER07	N22 CELL-D	03/09/2022	9040C

Date/Time 03/09/22 11:30

Raw Sample Received by: SP (P/C)

Raw Sample Relinquished by: SP (P/C)

Date/Time 03/09/22 12:35

Raw Sample Received by: SP (P/C)

Raw Sample Relinquished by: SP (P/C)