

LB 119061


7470A

INSTRUMENT ID: CV2

Linear

 μ Abs.:

39654



A= 0.0000e+000

B= 2.4978e-004

C= 2.0497e-002

Rho= 0.9997556

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0.11)
0.0	0.000	0.032	0.032	47	0.000	47					15
0.2	0.200	0.230	0.030	837	0.0 %	837					-5
2.5	2.500	2.379	-0.121	9442	0.0 %	9442					0
5.0	5.000	5.000	0.000	19937	0.0 %	19937					2
7.5	7.500	7.633	0.133	30478	0.0 %	30478					-1
10.0	10.000	9.925	-0.075	39654	0.0 %	39654					

LB119061 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 09 Mar 2022 12:39:31

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	47	-	0.0000	7470A	PPB	09 Mar 2022 12:51:46	S	Std
0.2 - 1	52	837	-	0.2000	7470A	PPB	09 Mar 2022 12:54:04	S	Std
2.5 - 1	52.5	9442	-	2.5000	7470A	PPB	09 Mar 2022 12:56:22	S	Std
5.0 - 1	55	19937	-	5.0000	7470A	PPB	09 Mar 2022 12:58:39	S	Std
7.5 - 1	57.5	30478	-	7.5000	7470A	PPB	09 Mar 2022 13:00:56	S	Std
10.0 - 1	510	39654	-	10.0000	7470A	PPB	09 Mar 2022 13:03:12	S	Std
ICV25 - 1	ICV25	14677	3.6866	-	7470A	PPB	09 Mar 2022 13:05:50	U	SMPL
ICB25 - 1	ICB25	-3	0.0197	-	7470A	PPB	09 Mar 2022 13:08:05	U	SMPL
CCV25 - 1	CCV25	19027	4.7731	-	7470A	PPB	09 Mar 2022 13:10:20	U	SMPL
CCB25 - 1	CCB25	-13	0.0173	-	7470A	PPB	09 Mar 2022 13:12:35	U	SMPL
CRA - 1	CRA	798	0.2198	-	7470A	PPB	09 Mar 2022 13:14:51	U	SMPL
HighStd - 1	HighStd	37983	9.5080	-	7470A	PPB	09 Mar 2022 13:17:06	U	SMPL
ChkStd - 1	ChkStd	26672	6.6827	-	7470A	PPB	09 Mar 2022 13:19:22	U	SMPL
PB143216BL - 1	PBW	-35	0.0118	-	7470A	PPB	09 Mar 2022 13:21:39	U	SMPL
PB143216BS - 1	LCSW	16004	4.0180	-	7470A	PPB	09 Mar 2022 13:23:55	U	SMPL
N1807-01 - 1	VNJ-223	0	0.0205	-	7470A	PPB	09 Mar 2022 13:26:11	U	SMPL
N1830-02 - 1	11986	32	0.0285	-	7470A	PPB	09 Mar 2022 13:28:28	U	SMPL
N1830-02DUP - 1	11986DUP	35	0.0292	-	7470A	PPB	09 Mar 2022 13:30:45	U	SMPL
N1830-02MS - 1	11986MS	18098	4.5411	-	7470A	PPB	09 Mar 2022 13:33:02	U	SMPL
N1830-02MSD - 1	11986MSD	15616	3.9211	-	7470A	PPB	09 Mar 2022 13:35:17	U	SMPL
CCV26 - 1	CCV26	20137	5.0504	-	7470A	PPB	09 Mar 2022 14:12:16	U	SMPL
CCB26 - 1	CCB26	-38	0.0110	-	7470A	PPB	09 Mar 2022 14:14:32	U	SMPL
N1830-05 - 1	11977	24	0.0265	-	7470A	PPB	09 Mar 2022 14:16:47	U	SMPL
N1830-08 - 1	350	34	0.0290	-	7470A	PPB	09 Mar 2022 14:19:03	U	SMPL
N1834-02 - 1	AUD-1281	36	0.0295	-	7470A	PPB	09 Mar 2022 14:21:19	U	SMPL
N1835-05 - 1	AUD-1272	15	0.0242	-	7470A	PPB	09 Mar 2022 14:23:35	U	SMPL
N1835-08 - 1	AUD-1242	33	0.0287	-	7470A	PPB	09 Mar 2022 14:25:51	U	SMPL
PB143217BL - 1	PBW	3	0.0212	-	7470A	PPB	09 Mar 2022 14:28:07	U	SMPL
PB143217BS - 1	LCSW	14503	3.6431	-	7470A	PPB	09 Mar 2022 14:30:24	U	SMPL
N1836-02 - 1	TRE-1282	-7	0.0187	-	7470A	PPB	09 Mar 2022 14:32:41	U	SMPL
N1836-02DUP - 1	TRE-1282DUP	18	0.0250	-	7470A	PPB	09 Mar 2022 14:34:58	U	SMPL
N1836-02MS - 1	TRE-1282MS	14459	3.6321	-	7470A	PPB	09 Mar 2022 14:37:14	U	SMPL
CCV27 - 1	CCV27	20475	5.1348	-	7470A	PPB	09 Mar 2022 14:39:30	U	SMPL
CCB27 - 1	CCB27	-28	0.0135	-	7470A	PPB	09 Mar 2022 14:41:45	U	SMPL
N1836-02MSD - 1	TRE-1282MSD	16479	4.1367	-	7470A	PPB	09 Mar 2022 14:44:01	U	SMPL
N1852-02 - 1	DRUM	-10	0.0180	-	7470A	PPB	09 Mar 2022 14:46:17	U	SMPL
PB143156TB - 1	PB143156TB	14	0.0240	-	7470A	PPB	09 Mar 2022 14:48:33	U	SMPL
PB143193TB - 1	PB143193TB	31	0.0282	-	7470A	PPB	09 Mar 2022 14:50:50	U	SMPL
N1830-02LX5 - 1		32	0.0285	-	7470A	PPB	09 Mar 2022 14:53:06	U	SMPL
N1830-02A - 1		15480	3.8871	-	7470A	PPB	09 Mar 2022 14:55:22	U	SMPL
N1836-02LX5 - 1		-34	0.0120	-	7470A	PPB	09 Mar 2022 14:57:39	U	SMPL
N1836-02A - 1		15456	3.8812	-	7470A	PPB	09 Mar 2022 14:59:56	U	SMPL
CCV28 - 1	CCV28	20390	5.1136	-	7470A	PPB	09 Mar 2022 15:02:13	U	SMPL
CCB28 - 1	CCB28	-24	0.0145	-	7470A	PPB	09 Mar 2022 15:04:30	U	SMPL