



PERCENT SOLID

Supervisor: mohammad

Analyst: jignesh

Date: 3/10/2022

OVENTEMP IN Celsius(°C): 107

Time IN: 15:40

In Date: 03/09/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 08:30

Out Date: 03/10/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB119066

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
N1817-01	DBSD5	1	1.13	8.71	9.84	7.37	71.6	
N1817-02	DBSD6	2	1.16	8.78	9.94	7.62	73.6	
N1817-03	DBSD7	3	1.13	8.66	9.79	7.58	74.5	
N1817-04	DBSD8	4	1.14	8.60	9.74	7.17	70.1	
N1817-05	DBSD9	5	1.12	8.54	9.66	6.41	61.9	
N1817-06	DBSE0	6	1.15	8.80	9.95	8.84	87.4	
N1817-07	DBSE1	7	1.13	8.84	9.97	9.17	91.0	
N1817-08	DBSE2	8	1.14	8.52	9.66	8.6	87.6	
N1817-09	DBSE3	9	1.13	8.63	9.76	8.43	84.6	
N1817-10	DBSE4	10	1.13	8.83	9.96	8.78	86.6	
N1817-11	DBSE5	11	1.14	8.59	9.73	8.17	81.8	
N1817-12	DBSE6	12	1.13	8.41	9.54	8.72	90.2	
N1817-23	DBRJ5	13	1.00	1.00	2.00	2.00	100.0	P.E. SAMPLE
N1817-24	DBRJ7	14	1.00	1.00	2.00	2.00	100.0	P.E. SAMPLE
N1817-25	DBRX3	15	1.12	8.82	9.94	8.76	86.6	
N1817-26	DBRX4	16	1.17	8.81	9.98	9.4	93.4	
N1817-27	DBRX5	17	1.12	8.81	9.93	8.06	78.8	
N1817-28	DBRX6	18	1.1	8.42	9.52	7.27	73.3	
N1817-29	DBRX6D	19	1.1	8.42	9.52	7.27	73.3	
N1817-30	DBRX6S	20	1.1	8.42	9.52	7.27	73.3	
N1817-31	DBRX7	21	1.16	8.74	9.9	6.66	62.9	
N1817-32	DBRX8	22	1.16	8.80	9.96	8.09	78.7	
N1817-33	DBRX9	23	1.13	8.79	9.92	8.59	84.9	
N1817-34	DBRY0	24	1.14	8.84	9.98	8.62	84.6	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$