

LB119070

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : 12

Instrument ID : Konelab

3/9/2022 16:16

Test: ^{Reactive}~~Total~~ CN

Sample Id	^{µg/L} Result	Dil. 1 +	Response	Errors
ICV1	95.747	0.0	0.088	
ICB1	0.605	0.0	0.001	
CCV1	244.645	0.0	0.225	
CCB1	0.580	0.0	0.001	
PB143222BL	0.456	0.0	0.001	
N1852-02	0.384	0.0	0.001	
N1852-02DUP	0.561	0.0	0.001	
CCV2	258.297	0.0	0.237	
CCB2	0.763	0.0	0.001	

N	9
Mean	66.893
SD	109.2394
CV%	163.30

Aquakem v. 7.2AQ1

Results from time period:

Wed Mar 09 16:11:38 2022

Wed Mar 09 16:15:31 2022

Sample Id	Sam/Ctr/c	Test short name	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	0.3057	µg/l	3/9/2022 11:01:10	
5.OPPBCN	A	Reactive CN	P	5.0527	µg/l	3/9/2022 11:01:11	
10PPBCN	A	Reactive CN	P	9.9846	µg/l	3/9/2022 11:01:12	
50PPBCN	A	Reactive CN	P	49.4327	µg/l	3/9/2022 11:01:13	
100PPBCN	A	Reactive CN	P	99.2105	µg/l	3/9/2022 11:01:14	
250PPBCN	A	Reactive CN	P	251.5988	µg/l	3/9/2022 11:01:15	
500PPBCN	A	Reactive CN	P	499.415	µg/l	3/9/2022 11:01:16	
ICV1	S	Reactive CN	P	95.7472	µg/l	3/9/2022 16:11:39	
ICB1	S	Reactive CN	P	0.6048	µg/l	3/9/2022 16:11:40	
CCV1	S	Reactive CN	P	244.6452	µg/l	3/9/2022 16:11:42	
CCB1	S	Reactive CN	P	0.5799	µg/l	3/9/2022 16:11:44	
PB143222BL	S	Reactive CN	P	0.456	µg/l	3/9/2022 16:11:47	
N1852-02	S	Reactive CN	P	0.3835	µg/l	3/9/2022 16:11:48	
N1852-02DUP	S	Reactive CN	P	0.561	µg/l	3/9/2022 16:15:29	
CCV2	S	Reactive CN	P	258.2973	µg/l	3/9/2022 16:15:30	
CCB2	S	Reactive CN	P	0.7633	µg/l	3/9/2022 16:15:31	

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : 12 Instrument ID : Konelab

3/9/2022 11:01


Test Total CN

Accepted 3/9/2022 11:01

Factor 1091
Bias 0.001

Coeff. of det. 0.999981

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.001	0.3057	0.0000	-
2	5.0PPBCN	0.005	5.0527	5.0000	1.1
3	10PPBCN	0.010	9.9846	10.0000	-0.2
4	50PPBCN	0.046	49.4327	50.0000	-1.1
5	100PPBCN	0.092	99.2105	100.0000	-0.8
6	250PPBCN	0.231	251.5988	250.0000	0.6
7	500PPBCN	0.458	499.4150	500.0000	-0.1

12 3/9/22