



PERCENT SOLID

Supervisor: mohammad

Analyst: JIGNESH

Date: 3/11/2022

OVENTEMP IN Celsius(°C): 107

Time IN: 15:30

In Date: 03/10/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 08:24

Out Date: 03/11/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB119082

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
N1882-01	DBRX3	1	1.12	8.82	9.94	8.76	86.6	
N1882-02	DBRX4	2	1.17	8.81	9.98	9.4	93.4	
N1882-03	DBRX5	3	1.12	8.81	9.93	8.06	78.8	
N1882-04	DBRX6	4	1.1	8.42	9.52	7.27	73.3	
N1882-05	DBRX6MS	5	1.1	8.42	9.52	7.27	73.3	
N1882-06	DBRX6MSD	6	1.1	8.42	9.52	7.27	73.3	
N1882-07	DBRX7	7	1.16	8.74	9.9	6.66	62.9	
N1882-08	DBRX8	8	1.16	8.80	9.96	8.09	78.7	
N1882-09	DBRX9	9	1.13	8.79	9.92	8.59	84.9	
N1882-10	DBRY0	10	1.14	8.84	9.98	8.62	84.6	
N1882-11	DBRY1	11	1.14	8.52	9.66	8.34	84.5	
N1882-12	DBRY2	12	1.18	8.54	9.72	8.37	84.2	
N1882-13	DBRY3	13	1.16	8.81	9.97	8.44	82.6	
N1882-14	DBRY4	14	1.18	8.60	9.78	8.62	86.5	
N1882-15	DBRY5	15	1.13	8.54	9.67	9.00	92.2	
N1882-16	DBRY6	16	1.15	8.81	9.96	8.74	86.2	
N1882-17	DBRY7	17	1.15	8.74	9.89	7.85	76.7	
N1882-18	DBRY8	18	1.16	8.36	9.52	5.89	56.6	
N1882-19	DBRY9	19	1.16	8.68	9.84	7.1	68.4	
N1882-20	DBRZ0	20	1.18	8.58	9.76	7.7	76.0	
N1882-21	DBRZ1	21	1.16	8.80	9.96	7.03	66.7	
N1882-22	DBRZ2	22	1.1	8.62	9.72	8.26	83.1	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$