


LB119086

7471B

INSTRUMENT ID: CV1

 μ Abs.:

53861



Linear

A= 0.0000e+000

B= 1.8680e-004

C= -2.7469e-002

Rho= 0.9998998

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	σ/μ
0.0	0.000	0.007	0.007	186	0.000	186					—
0.2	0.200	0.192	-0.008	1174	0.0 %	1174					-4)
2.5	2.500	2.482	-0.018	13434	0.0 %	13434					-1
5.0	5.000	5.078	0.078	27333	0.0 %	27333					2
7.5	7.500	7.407	-0.093	39799	0.0 %	39799					-1
10.0	10.000	10.034	0.034	53861	0.0 %	53861					0

LB119086 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 10 Mar 2022 16:17:06				Date	Type	Type
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units				
0.0 - 1	50	186	-	0.0000	7471B	PPB	10 Mar 2022	16:17:27	S	Std
0.2 - 1	50.2	1174	-	0.2000	7471B	PPB	10 Mar 2022	16:19:45	S	Std
2.5 - 1	52.5	13434	-	2.5000	7471B	PPB	10 Mar 2022	16:22:02	S	Std
5.0 - 1	55	27333	-	5.0000	7471B	PPB	10 Mar 2022	16:24:20	S	Std
7.5 - 1	57.5	39799	-	7.5000	7471B	PPB	10 Mar 2022	16:26:36	S	Std
10.0 - 1	58	53861	-	10.0000	7471B	PPB	10 Mar 2022	16:28:53	S	Std
ICV28 - 1	ICV28	21083	3.9108	-	7471B	PPB	10 Mar 2022	16:31:33	U	SMPL
ICB28 - 1	ICB28	126	-0.0039	-	7471B	PPB	10 Mar 2022	16:33:48	U	SMPL
CCV37 - 1	CCV37	27247	5.0622	-	7471B	PPB	10 Mar 2022	16:36:04	U	SMPL
CCB37 - 1	CCB37	123	-0.0045	-	7471B	PPB	10 Mar 2022	16:38:19	U	SMPL
CRA - 1	CRA	1092	0.1765	-	7471B	PPB	10 Mar 2022	16:40:35	U	SMPL
HighStd - 1	HighStd	50823	9.4662	-	7471B	PPB	10 Mar 2022	16:42:50	U	SMPL
ChkStd - 1	ChkStd	37389	6.9567	-	7471B	PPB	10 Mar 2022	16:45:06	U	SMPL
PB143243BL - 1	PBS	3	-0.0269	-	7471B	PPB	10 Mar 2022	16:47:24	U	SMPL
PB143243BS - 1	LCSS	21648	4.0163	-	7471B	PPB	10 Mar 2022	16:49:41	U	SMPL
N1847-02 - 1	CG-B3-030722-0-2	17581	3.2566	-	7471B	PPB	10 Mar 2022	16:51:57	U	SMPL
N1847-04 - 1	CG-B3-030722-2-5	19563	3.6269	-	7471B	PPB	10 Mar 2022	16:54:14	U	SMPL
N1847-08 - 1	CG-B1-030722-0-2	836	0.1287	-	7471B	PPB	10 Mar 2022	16:56:31	U	SMPL
N1847-08DUP - 1	CG-B1-030722-0-2DUP	914	0.1433	-	7471B	PPB	10 Mar 2022	16:58:49	U	SMPL
N1847-08MS - 1	CG-B1-030722-0-2MS	22341	4.1458	-	7471B	PPB	10 Mar 2022	17:01:04	U	SMPL
CCV38 - 1	CCV38	25255	4.6901	-	7471B	PPB	10 Mar 2022	17:03:20	U	SMPL
CCB38 - 1	CCB38	109	-0.0071	-	7471B	PPB	10 Mar 2022	17:05:36	U	SMPL
N1847-08MSD - 1	CG-B1-030722-0-2MSD	20176	3.7414	-	7471B	PPB	10 Mar 2022	17:07:53	U	SMPL
N1847-09 - 1	CG-B1-030722-0-2-BD	1269	0.2096	-	7471B	PPB	10 Mar 2022	17:10:08	U	SMPL
N1847-10 - 1	CG-B1-030722-2-5	321	0.0325	-	7471B	PPB	10 Mar 2022	17:12:24	U	SMPL
N1848-02 - 1	CG-B4-030722-0-2	367286	68.5810	-	7471B	PPB	10 Mar 2022	17:14:40	U	SMPL
N1848-04 - 1	CG-B4-030722-2-5	13384	2.4726	-	7471B	PPB	10 Mar 2022	17:16:56	U	SMPL
N1848-08 - 1	CG-B2-030722-0-2	7159	1.3098	-	7471B	PPB	10 Mar 2022	17:19:19	U	SMPL
N1848-10 - 1	CG-B2-030722-2-5	3497	0.6258	-	7471B	PPB	10 Mar 2022	17:21:35	U	SMPL
N1863-01 - 1	CHRT-25720	3988	0.7175	-	7471B	PPB	10 Mar 2022	17:23:53	U	SMPL
N1864-03 - 1	MB-B2-030922-0-2	1516	0.2557	-	7471B	PPB	10 Mar 2022	17:26:10	U	SMPL
N1864-05 - 1	MB-B2-030922-2-4	950	0.1500	-	7471B	PPB	10 Mar 2022	17:28:25	U	SMPL
CCV39 - 1	CCV39	26880	4.9937	-	7471B	PPB	10 Mar 2022	17:30:41	U	SMPL
CCB39 - 1	CCB39	114	-0.0062	-	7471B	PPB	10 Mar 2022	17:32:57	U	SMPL
N1864-07 - 1	MB-B2-030922-4-6	585	0.0818	-	7471B	PPB	10 Mar 2022	17:35:13	U	SMPL
N1864-09 - 1	MB-B2-030922-4-6-DUP	739	0.1106	-	7471B	PPB	10 Mar 2022	17:37:28	U	SMPL
N1864-11 - 1	MB-B1-030922-0-2	2310	0.4040	-	7471B	PPB	10 Mar 2022	17:39:44	U	SMPL
N1864-13 - 1	MB-B1-030922-2-4	8725	1.6023	-	7471B	PPB	10 Mar 2022	17:42:00	U	SMPL
N1864-15 - 1	MB-B1-030922-4-6	2178	0.3794	-	7471B	PPB	10 Mar 2022	17:44:17	U	SMPL
N1867-01 - 1	HD-01-030922	2506	0.4406	-	7471B	PPB	10 Mar 2022	17:46:33	U	SMPL
N1874-01 - 1	EO-03-030922	2556	0.4500	-	7471B	PPB	10 Mar 2022	17:48:50	U	SMPL
N1847-08LX5 - 1		138	-0.0017	-	7471B	PPB	10 Mar 2022	17:51:07	U	SMPL
N1847-08A - 1		21700	4.0261	-	7471B	PPB	10 Mar 2022	17:53:24	U	SMPL
CCV40 - 1	CCV40	27351	5.0817	-	7471B	PPB	10 Mar 2022	17:55:40	U	SMPL
CCB40 - 1	CCB40	126	-0.0039	-	7471B	PPB	10 Mar 2022	17:57:56	U	SMPL