

LB119100


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

46692



A= 0.0000e+000

B= 2.1406e-004

C= 3.8184e-003

Rho= 0.9999856

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/00
0.0	0.000	0.012	0.012	36	0.000	36					
0.2	0.200	0.209	0.009	959	0.0 %	959					4
2.5	2.500	2.486	-0.014	11597	0.0 %	11597					-1
5.0	5.000	4.966	-0.034	23181	0.0 %	23181					-1
7.5	7.500	7.528	0.028	35150	0.0 %	35150					0
10.0	10.000	9.999	-0.001	46692	0.0 %	46692					0

LB119100 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 11 Mar 2022 12:03:50

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	36	-	0.0000	7471B	PPB	11 Mar 2022 12:17:10	S	Std
0.2 - 1	50.2	959	-	0.2000	7471B	PPB	11 Mar 2022 12:19:28	S	Std
2.5 - 1	52.5	11597	-	2.5000	7471B	PPB	11 Mar 2022 12:21:47	S	Std
5.0 - 1	55	23181	-	5.0000	7471B	PPB	11 Mar 2022 12:24:04	S	Std
7.5 - 1	57.5	35150	-	7.5000	7471B	PPB	11 Mar 2022 12:26:20	S	Std
10.0 - 1	510	46692	-	10.0000	7471B	PPB	11 Mar 2022 12:28:38	S	Std
ICV29 - 1	ICV29	17818	3.8180	-	7471B	PPB	11 Mar 2022 12:31:18	U	SMPL
ICB29 - 1	ICB29	-52	-0.0073	-	7471B	PPB	11 Mar 2022 12:33:33	U	SMPL
CCV41 - 1	CCV41	23167	4.9630	-	7471B	PPB	11 Mar 2022 12:35:50	U	SMPL
CCB41 - 1	CCB41	-60	-0.0090	-	7471B	PPB	11 Mar 2022 12:38:05	U	SMPL
CRA - 1	CRA	914	0.1995	-	7471B	PPB	11 Mar 2022 12:40:22	U	SMPL
HighStd - 1	HighStd	45955	9.8411	-	7471B	PPB	11 Mar 2022 12:42:37	U	SMPL
ChkStd - 1	ChkStd	34133	7.3105	-	7471B	PPB	11 Mar 2022 12:44:53	U	SMPL
PB143288BL - 1	PBS	-86	-0.0146	-	7471B	PPB	11 Mar 2022 12:47:11	U	SMPL
PB143288BS - 1	LCSS	20236	4.3356	-	7471B	PPB	11 Mar 2022 12:49:29	U	SMPL
N1884-01 - 1	HR-03-031022	233	0.0537	-	7471B	PPB	11 Mar 2022 12:51:45	U	SMPL
N1884-03 - 1	HR-04-031022	3582	0.7706	-	7471B	PPB	11 Mar 2022 12:54:02	U	SMPL
N1886-01 - 1	2822-D	363	0.0815	-	7471B	PPB	11 Mar 2022 12:56:19	U	SMPL
N1886-02 - 1	2822-S	5922	1.2715	-	7471B	PPB	11 Mar 2022 12:58:36	U	SMPL
CCV42 - 1	CCV42	21455	4.5966	-	7471B	PPB	11 Mar 2022 13:16:00	U	SMPL
CCB42 - 1	CCB42	71	0.0190	-	7471B	PPB	11 Mar 2022 13:18:16	U	SMPL
N1890-01 - 1	VNJ-232	1982	0.4281	-	7471B	PPB	11 Mar 2022 13:20:32	U	SMPL
N1890-01DUP - 1	VNJ-232DUP	2024	0.4371	-	7471B	PPB	11 Mar 2022 13:22:48	U	SMPL
N1890-01MS - 1	VNJ-232MS	22112	4.7372	-	7471B	PPB	11 Mar 2022 13:25:03	U	SMPL
N1890-01MSD - 1	VNJ-232MSD	21968	4.7064	-	7471B	PPB	11 Mar 2022 13:27:19	U	SMPL
N1893-05 - 1	CONCRETE	226	0.0522	-	7471B	PPB	11 Mar 2022 13:29:35	U	SMPL
N1890-01LX5 - 1		375	0.0841	-	7471B	PPB	11 Mar 2022 13:31:52	U	SMPL
N1890-01A - 1		22166	4.7488	-	7471B	PPB	11 Mar 2022 13:34:08	U	SMPL
N1848-02DLX10 - 1		46262	9.9068	-	7471B	PPB	11 Mar 2022 13:36:25	U	SMPL
CCV43 - 1	CCV43	23443	5.0221	-	7471B	PPB	11 Mar 2022 13:38:42	U	SMPL
CCB43 - 1	CCB43	13	0.0066	-	7471B	PPB	11 Mar 2022 13:41:00	U	SMPL