

LB119105


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

35953



A= 0.0000e+000

B= 2.7109e-004

C= -6.3099e-003

Rho= 0.9985016

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/p 1)
0.0	0.000	0.019	0.019	92	0.000	92					-
0.2	0.200	0.181	-0.019	692	0.0 %	692					-10
2.5	2.500	2.478	-0.022	9166	0.0 %	9166					-1
5.0	5.000	4.880	-0.120	18024	0.0 %	18024					-2
7.5	7.500	7.902	0.402	29172	0.0 %	29172					-5
10.0	10.000	9.740	-0.260	35953	0.0 %	35953					-3

LB119105 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 11 Mar 2022 14:20:51

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SO	92	-	0.0000	7470A	PPB	11 Mar 2022 14:27:15	S	Std
0.2 - 1	SO-2	692	-	0.2000	7470A	PPB	11 Mar 2022 14:29:32	S	Std
2.5 - 1	SO-5	9166	-	2.5000	7470A	PPB	11 Mar 2022 14:31:50	S	Std
5.0 - 1	SO-5	18024	-	5.0000	7470A	PPB	11 Mar 2022 14:34:07	S	Std
7.5 - 1	SO-5	29172	-	7.5000	7470A	PPB	11 Mar 2022 14:39:01	S	Std
10.0 - 1	SO-5	35953	-	10.0000	7470A	PPB	11 Mar 2022 14:43:48	S	Std
ICV30 - 1	ICV30	14675	3.9719	-7470A	PPB	PPB	11 Mar 2022 14:51:21	U	SMPL
ICB30 - 1	ICB30	4	-0.0052	-7470A	PPB	PPB	11 Mar 2022 14:53:36	U	SMPL
CCV44 - 1	CCV44	18081	4.8952	-7470A	PPB	PPB	11 Mar 2022 14:55:51	U	SMPL
CCB44 - 1	CCB44	-42	-0.0177	-7470A	PPB	PPB	11 Mar 2022 14:58:07	U	SMPL
CRA - 1	CRA	698	0.1829	-7470A	PPB	PPB	11 Mar 2022 15:00:22	U	SMPL
HighStd - 1	HighStd	36238	9.8173	-7470A	PPB	PPB	11 Mar 2022 15:02:37	U	SMPL
ChkStd - 1	ChkStd	27594	7.4740	-7470A	PPB	PPB	11 Mar 2022 15:04:52	U	SMPL
PB143290BL - 1	PBW	-48	-0.0193	-7470A	PPB	PPB	11 Mar 2022 15:07:09	U	SMPL
PB143290BS - 1	LCSW	14224	3.8496	-7470A	PPB	PPB	11 Mar 2022 15:09:26	U	SMPL
N1848-06 - 1	FIELD-BLANK	148	0.0338	-7470A	PPB	PPB	11 Mar 2022 15:11:43	U	SMPL
N1864-01 - 1	FIELD-BLANK	228	0.0555	-7470A	PPB	PPB	11 Mar 2022 15:13:59	U	SMPL
N1875-06 - 1	TW-10B	29066	7.8731	-7470A	PPB	PPB	11 Mar 2022 15:16:17	U	SMPL
N1886-03 - 1	30969	474	0.1222	-7470A	PPB	PPB	11 Mar 2022 15:23:29	U	SMPL
N1886-03DUP - 1	30969DUP	427	0.1094	-7470A	PPB	PPB	11 Mar 2022 15:25:44	U	SMPL
CCV45 - 1	CCV45	18194	4.9258	-7470A	PPB	PPB	11 Mar 2022 15:28:00	U	SMPL
CCB45 - 1	CCB45	-49	-0.0196	-7470A	PPB	PPB	11 Mar 2022 15:30:15	U	SMPL
N1886-03MS - 1	30969MS	15854	4.2915	-7470A	PPB	PPB	11 Mar 2022 15:32:30	U	SMPL
N1886-03MSD - 1	30969MSD	15747	4.2625	-7470A	PPB	PPB	11 Mar 2022 15:34:45	U	SMPL
N1886-04 - 1	30970	68	0.0121	-7470A	PPB	PPB	11 Mar 2022 15:37:01	U	SMPL
N1886-05 - 1	2822-A	1731	0.4629	-7470A	PPB	PPB	11 Mar 2022 15:39:17	U	SMPL
N1886-06 - 1	2822-B	138	0.0311	-7470A	PPB	PPB	11 Mar 2022 15:41:33	U	SMPL
N1886-03LX5 - 1		551	0.1431	-7470A	PPB	PPB	11 Mar 2022 15:46:34	U	SMPL
N1886-03A - 1		12264	3.3183	-7470A	PPB	PPB	11 Mar 2022 15:48:51	U	SMPL
CCV46 - 1	CCV46	17914	4.8499	-7470A	PPB	PPB	11 Mar 2022 15:53:50	U	SMPL
CCB46 - 1	CCB46	-9	-0.0087	-7470A	PPB	PPB	11 Mar 2022 15:56:06	U	SMPL