



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : RITESHKUMAR

Run Number: LB119124

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	03/12/2022	09:30
2	CAL2	1	Water	NA	NA	20.3	7.00	03/12/2022	09:31
3	CAL3	1	Water	NA	NA	20.4	10.03	03/12/2022	09:33
4	ICV	1	Water	NA	NA	20.4	6.99	03/12/2022	09:35
5	CCV1	1	Water	NA	NA	20.5	2.01	03/12/2022	09:36
6	N1905-11	1	Solid	20.03	20	20.6	10.07	03/12/2022	09:50
7	N1905-11DUP	1	Solid	20.04	20	20.4	10.08	03/12/2022	09:51
8	N1905-13	1	Solid	20.03	20	20.3	10.41	03/12/2022	09:55
9	N1905-15	1	Solid	20.03	20	20.3	9.15	03/12/2022	10:00
10	N1905-17	1	Solid	20.02	20	20.2	8.08	03/12/2022	10:03
11	N1905-19	1	Solid	20.02	20	20.3	8.15	03/12/2022	10:09
12	N1918-01	1	Solid	20.04	20	20.4	12.39	03/12/2022	10:35
13	N1918-04	1	Solid	20.04	20	20.03	12.42	03/12/2022	10:36
14	CCV2	1	Water	NA	NA	20.4	12.03	03/12/2022	10:50

13 119124

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N1894 WorkList ID : 157644 Department : Wet-Chemistry Date : 03-12-2022 09:14:08

Due Date	Matrix	Sample	Test	Raw Sample			Collect Date	Method
				Preservative	Customer	Storage Location		
03/10/2022	Solid	N1905-11	pH	Cool 4 deg C	ROUX02	N23 WC-14A-1C	03/10/2022	9045D
03/10/2022	Solid	N1905-13	pH	Cool 4 deg C	ROUX02	N23 WC-14A-1A-3	03/10/2022	9045D
03/10/2022	Solid	N1905-15	pH	Cool 4 deg C	ROUX02	N23 WC-14A-1A-1	03/10/2022	9045D
03/10/2022	Solid	N1905-17	pH	Cool 4 deg C	ROUX02	N23 WC-14A-1A-4	03/10/2022	9045D
03/10/2022	Solid	N1905-19	pH	Cool 4 deg C	ROUX02	N23 WC-14A-1A-2	03/10/2022	9045D
03/11/2022	Solid	N1918-01	pH	Cool 4 deg C	TULL02	P11 AVENUE-X-GROUT-SAMPLE-A	03/11/2022	9045D
03/11/2022	Solid	N1918-04	pH	Cool 4 deg C	TULL02	P11 AVENUE-X-GROUT-SAMPLE-B	03/11/2022	9045D

Date/Time 03.12.22 09:20
Raw Sample Received by: JH (WC)
Raw Sample Relinquished by: JH (WC)

Date/Time 03.12.22 12:00
Raw Sample Received by: JH (WC)
Raw Sample Relinquished by: JH (WC)