

LB119144


7471B

INSTRUMENT ID: CV2

Linear

 μ Abs.:

50508



A= 0.0000e+000

B= 1.9638e-004

C= 2.6534e-002

Rho= 0.9998217

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.030	0.030	17	0.000	17					1
0.2	0.200	0.223	0.023	1000	0.0 %	1000					12
2.5	2.500	2.373	-0.127	11946	0.0 %	11946					5
5.0	5.000	5.056	0.056	25608	0.0 %	25608					1
7.5	7.500	7.574	0.074	38430	0.0 %	38430					1
10.0	10.000	9.946	-0.054	50508	0.0 %	50508					1

LB119144 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 14 Mar 2022 13:48:10					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	17	-	0.0000	7471B	PPB	14 Mar 2022 13:51:35	S	Std
0.2 - 1	50.2	1000	-	0.2000	7471B	PPB	14 Mar 2022 13:53:52	S	Std
2.5 - 1	52.5	11946	-	2.5000	7471B	PPB	14 Mar 2022 13:56:10	S	Std
5.0 - 1	55	25608	-	5.0000	7471B	PPB	14 Mar 2022 13:58:27	S	Std
7.5 - 1	57.5	38430	-	7.5000	7471B	PPB	14 Mar 2022 14:00:44	S	Std
10.0 - 1	510	50508	-	10.0000	7471B	PPB	14 Mar 2022 14:03:00	S	Std
ICV31 - 1	ICV31	20365	4.0259	-	7471B	PPB	14 Mar 2022 14:06:05	U	SMPL
ICB31 - 1	ICB31	49	0.0362	-	7471B	PPB	14 Mar 2022 14:08:20	U	SMPL
CCV47 - 1	CCV47	26033	5.1390	-	7471B	PPB	14 Mar 2022 14:10:35	U	SMPL
CCB47 - 1	CCB47	36	0.0336	-	7471B	PPB	14 Mar 2022 14:12:51	U	SMPL
CRA - 1	CRA	985	0.2200	-	7471B	PPB	14 Mar 2022 14:15:07	U	SMPL
HighStd - 1	HighStd	50567	9.9571	-	7471B	PPB	14 Mar 2022 14:17:23	U	SMPL
ChkStd - 1	ChkStd	36189	7.1335	-	7471B	PPB	14 Mar 2022 14:19:39	U	SMPL
PB143342BL - 1	PBS	203	0.0664	-	7471B	PPB	14 Mar 2022 14:21:56	U	SMPL
PB143342BS - 1	LCSS	20374	4.0277	-	7471B	PPB	14 Mar 2022 14:24:13	U	SMPL
N1894-01 - 1	TP1(0-10)	16899	3.3452	-	7471B	PPB	14 Mar 2022 14:26:30	U	SMPL
N1894-02 - 1	TP2(0-5)	28058	5.5367	-	7471B	PPB	14 Mar 2022 14:28:47	U	SMPL
N1894-03 - 1	TP3(0-10)	23437	4.6292	-	7471B	PPB	14 Mar 2022 14:31:04	U	SMPL
N1894-04 - 1	TP4(0-7)	27895	5.5047	-	7471B	PPB	14 Mar 2022 14:33:22	U	SMPL
N1894-05 - 1	TP5(0-5)	19747	3.9045	-	7471B	PPB	14 Mar 2022 14:35:38	U	SMPL
CCV48 - 1	CCV48	26344	5.2001	-	7471B	PPB	14 Mar 2022 14:37:53	U	SMPL
CCB48 - 1	CCB48	37	0.0338	-	7471B	PPB	14 Mar 2022 14:40:08	U	SMPL
N1912-01 - 1	VNJ-242	2823	0.5809	-	7471B	PPB	14 Mar 2022 14:42:24	U	SMPL
N1913-01 - 1	BS-SOILP	1333	0.2883	-	7471B	PPB	14 Mar 2022 14:44:39	U	SMPL
N1914-01 - 1	TP-1	3984	0.8089	-	7471B	PPB	14 Mar 2022 14:46:55	U	SMPL
N1914-04 - 1	TP-2	971	0.2172	-	7471B	PPB	14 Mar 2022 14:49:11	U	SMPL
N1916-01 - 1	CONCRETE-COMP	353	0.0959	-	7471B	PPB	14 Mar 2022 14:51:27	U	SMPL
N1918-07 - 1	JDSH-TB-STOCKPILE-A	35428	6.9840	-	7471B	PPB	14 Mar 2022 14:53:43	U	SMPL
N1918-10 - 1	JDSH-TB-STOCKPILE-B	33447	6.5950	-	7471B	PPB	14 Mar 2022 14:56:00	U	SMPL
N1918-13 - 1	JDSH-TB-STOCKPILE-C	16259	3.2196	-	7471B	PPB	14 Mar 2022 15:03:08	U	SMPL
N1918-13DUP - 1	JDSH-TB-STOCKPILE-CDUP	15985	3.1657	-	7471B	PPB	14 Mar 2022 15:05:26	U	SMPL
N1918-13MS - 1	JDSH-TB-STOCKPILE-CMS	33766	6.6577	-	7471B	PPB	14 Mar 2022 15:07:42	U	SMPL
CCV49 - 1	CCV49	26272	5.1859	-	7471B	PPB	14 Mar 2022 15:09:57	U	SMPL
CCB49 - 1	CCB49	46	0.0356	-	7471B	PPB	14 Mar 2022 15:12:14	U	SMPL
N1918-13MSD - 1	JDSH-TB-STOCKPILE-CMSD	37953	7.4799	-	7471B	PPB	14 Mar 2022 15:14:29	U	SMPL
N1918-13LX5 - 1		2278	0.4739	-	7471B	PPB	14 Mar 2022 16:37:23	U	SMPL
N1918-13A - 1		33300	6.5661	-	7471B	PPB	14 Mar 2022 16:39:42	U	SMPL
CCV50 - 1	CCV50	22806	4.5053	-	7471B	PPB	14 Mar 2022 16:51:20	U	SMPL
CCB50 - 1	CCB50	107	0.0475	-	7471B	PPB	14 Mar 2022 16:54:39	U	SMPL