

LB 119158


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

39982



A= 0.0000e+000

B= 2.5318e-004

C= -8.9813e-002

Rho= 0.9998896

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	1/10
0.0	0.000	-0.066	-0.066	94	0.000	94					1
0.2	0.200	0.231	0.031	1267	0.0 %	1267					1
2.5	2.500	2.587	0.087	10571	0.0 %	10571					3
5.0	5.000	4.975	-0.025	20003	0.0 %	20003					1
7.5	7.500	7.441	-0.059	29744	0.0 %	29744					1
10.0	10.000	10.033	0.033	39982	0.0 %	39982					0

LB119158 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 15 Mar 2022 09:14:01

Sample ID	Extended ID	µ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	94	-	0.0000	7470A	PPB	15 Mar 2022 14:00:30	S	Std
0.2 - 1	50-2	1267	-	0.2000	7470A	PPB	15 Mar 2022 14:02:47	S	Std
2.5 - 1	50-5	10571	-	2.5000	7470A	PPB	15 Mar 2022 14:05:05	S	Std
5.0 - 1	55	20003	-	5.0000	7470A	PPB	15 Mar 2022 14:07:22	S	Std
7.5 - 1	57-5	29744	-	7.5000	7470A	PPB	15 Mar 2022 14:14:56	S	Std
10.0 - 1	510	39982	-	10.0000	7470A	PPB	15 Mar 2022 14:19:52	S	Std
ICV32 - 1	ICV32	16573	4.1062	-	7470A	PPB	15 Mar 2022 14:22:35	U	SMPL
ICB32 - 1	ICB32	109	-0.0622	-	7470A	PPB	15 Mar 2022 14:24:50	U	SMPL
CCV51 - 1	CCV51	21619	5.3838	-	7470A	PPB	15 Mar 2022 14:27:05	U	SMPL
CCB51 - 1	CCB51	226	-0.0326	-	7470A	PPB	15 Mar 2022 14:29:21	U	SMPL
CRA - 1	CRA	1151	0.2016	-	7470A	PPB	15 Mar 2022 14:31:37	U	SMPL
HighStd - 1	HighStd	38199	9.5815	-	7470A	PPB	15 Mar 2022 14:33:53	U	SMPL
ChkStd - 1	ChkStd	27329	6.8294	-	7470A	PPB	15 Mar 2022 14:36:09	U	SMPL
PB143320BL - 1	PBW	173	-0.0460	-	7470A	PPB	15 Mar 2022 14:41:00	U	SMPL
PB143320BS - 1	LCSW	15554	3.8482	-	7470A	PPB	15 Mar 2022 14:43:15	U	SMPL
N1890-02 - 1	VNJ-232	-113	-0.1184	-	7470A	PPB	15 Mar 2022 14:45:32	U	SMPL
N1893-01 - 1	COMP-2	1	-0.0896	-	7470A	PPB	15 Mar 2022 14:47:49	U	SMPL
N1893-02 - 1	COMP-3	78	-0.0701	-	7470A	PPB	15 Mar 2022 14:50:07	U	SMPL
N1893-03 - 1	8110	54	-0.0761	-	7470A	PPB	15 Mar 2022 14:52:24	U	SMPL
N1893-04 - 1	8122	81	-0.0693	-	7470A	PPB	15 Mar 2022 14:54:40	U	SMPL
CCV52 - 1	CCV52	21434	5.3369	-	7470A	PPB	15 Mar 2022 14:56:55	U	SMPL
CCB52 - 1	CCB52	38	-0.0802	-	7470A	PPB	15 Mar 2022 14:59:11	U	SMPL
N1894-06 - 1	TP1(0-10)	51	-0.0769	-	7470A	PPB	15 Mar 2022 15:01:27	U	SMPL
N1894-07 - 1	TP2(0-5)	73	-0.0713	-	7470A	PPB	15 Mar 2022 15:03:43	U	SMPL
N1894-08 - 1	TP3(0-10)	90	-0.0670	-	7470A	PPB	15 Mar 2022 15:05:58	U	SMPL
N1894-09 - 1	TP4(0-7)	1	-0.0896	-	7470A	PPB	15 Mar 2022 15:08:14	U	SMPL
N1894-10 - 1	TP5(0-5)	89	-0.0673	-	7470A	PPB	15 Mar 2022 15:10:30	U	SMPL
N1918-02 - 1	AVENUE-X-GROUT-SAMPLE-A	160	-0.0493	-	7470A	PPB	15 Mar 2022 15:12:47	U	SMPL
N1918-05 - 1	AVENUE-X-GROUT-SAMPLE-B	59	-0.0749	-	7470A	PPB	15 Mar 2022 15:15:03	U	SMPL
N1918-08 - 1	JDSH-TB-STOCKPILE-A	67	-0.0728	-	7470A	PPB	15 Mar 2022 15:17:20	U	SMPL
N1918-11 - 1	JDSH-TB-STOCKPILE-B	2756	0.6080	-	7470A	PPB	15 Mar 2022 15:19:37	U	SMPL
N1918-14 - 1	JDSH-TB-STOCKPILE-C	70	-0.0721	-	7470A	PPB	15 Mar 2022 15:21:53	U	SMPL
CCV53 - 1	CCV53	21447	5.3402	-	7470A	PPB	15 Mar 2022 15:24:09	U	SMPL
CCB53 - 1	CCB53	42	-0.0792	-	7470A	PPB	15 Mar 2022 15:26:25	U	SMPL
N1918-14DUP - 1	JDSH-TB-STOCKPILE-CDUP	-2	-0.0903	-	7470A	PPB	15 Mar 2022 15:28:42	U	SMPL
N1918-14MS - 1	JDSH-TB-STOCKPILE-CMS	16178	4.0062	-	7470A	PPB	15 Mar 2022 15:30:58	U	SMPL
N1918-14MSD - 1	JDSH-TB-STOCKPILE-CMSD	14402	3.5565	-	7470A	PPB	15 Mar 2022 15:33:13	U	SMPL
PB143321BL - 1	PBW	189	-0.0420	-	7470A	PPB	15 Mar 2022 15:35:30	U	SMPL
PB143321BS - 1	LCSW	15475	3.8282	-	7470A	PPB	15 Mar 2022 15:37:47	U	SMPL
N1893-07 - 1	COMP-1	59	-0.0749	-	7470A	PPB	15 Mar 2022 15:40:04	U	SMPL
N1893-07DUP - 1	COMP-1DUP	7	-0.0880	-	7470A	PPB	15 Mar 2022 15:42:21	U	SMPL
N1893-07MS - 1	COMP-1MS	14352	3.5439	-	7470A	PPB	15 Mar 2022 15:44:38	U	SMPL
N1893-07MSD - 1	COMP-1MSD	14140	3.4902	-	7470A	PPB	15 Mar 2022 15:46:55	U	SMPL
N1893-08 - 1	COMP-4	-14	-0.0934	-	7470A	PPB	15 Mar 2022 15:49:13	U	SMPL
CCV54 - 1	CCV54	21260	5.2929	-	7470A	PPB	15 Mar 2022 15:51:30	U	SMPL
CCB54 - 1	CCB54	92	-0.0665	-	7470A	PPB	15 Mar 2022 15:53:45	U	SMPL
N1893-09 - 1	8117	19	-0.0850	-	7470A	PPB	15 Mar 2022 15:56:02	U	SMPL
N1893-10 - 1	TOTE	67	-0.0728	-	7470A	PPB	15 Mar 2022 15:58:19	U	SMPL
PB143322BL - 1	PBW	31	-0.0820	-	7470A	PPB	15 Mar 2022 16:00:35	U	SMPL
PB143322BS - 1	LCSW	15352	3.7971	-	7470A	PPB	15 Mar 2022 16:02:50	U	SMPL
N1905-12 - 1	WC-14A-1C	-11	-0.0926	-	7470A	PPB	15 Mar 2022 16:05:06	U	SMPL
N1905-12DUP - 1	WC-14A-1CDUP	37	-0.0804	-	7470A	PPB	15 Mar 2022 16:07:24	U	SMPL
N1905-12MS - 1	WC-14A-1CMS	15113	3.7365	-	7470A	PPB	15 Mar 2022 16:09:41	U	SMPL
N1905-12MSD - 1	WC-14A-1CMSD	15031	3.7158	-	7470A	PPB	15 Mar 2022 16:11:58	U	SMPL
N1905-14 - 1	WC-14A-1A-3	1	-0.0896	-	7470A	PPB	15 Mar 2022 16:14:17	U	SMPL
N1905-16 - 1	WC-14A-1A-1	0	-0.0898	-	7470A	PPB	15 Mar 2022 16:16:35	U	SMPL
CCV55 - 1	CCV55	21435	5.3372	-	7470A	PPB	15 Mar 2022 16:18:51	U	SMPL
CCB55 - 1	CCB55	212	-0.0361	-	7470A	PPB	15 Mar 2022 16:21:07	U	SMPL
N1905-18 - 1	WC-14A-1A-4	1077	0.1829	-	7470A	PPB	15 Mar 2022 16:23:24	U	SMPL
N1905-20 - 1	WC-14A-1A-2	7	-0.0880	-	7470A	PPB	15 Mar 2022 16:25:40	U	SMPL
N1908-02 - 1	WC-8-9-3	55	-0.0759	-	7470A	PPB	15 Mar 2022 16:27:56	U	SMPL
N1908-04 - 1	WC-8-9-2	37	-0.0804	-	7470A	PPB	15 Mar 2022 16:30:12	U	SMPL
N1908-06 - 1	WC-8-9-1	0	-0.0898	-	7470A	PPB	15 Mar 2022 16:32:28	U	SMPL
N1908-08 - 1	WC-14A-2	79	-0.0698	-	7470A	PPB	15 Mar 2022 16:34:45	U	SMPL
N1908-10 - 1	WC-14A-1b	42	-0.0792	-	7470A	PPB	15 Mar 2022 16:37:01	U	SMPL
PB143285TB - 1	PB143285TB	50	-0.0772	-	7470A	PPB	15 Mar 2022 16:39:19	U	SMPL
PB143287TB - 1	PB143287TB	18	-0.0853	-	7470A	PPB	15 Mar 2022 16:41:36	U	SMPL
PB143306TB - 1	PB143306TB	8	-0.0878	-	7470A	PPB	15 Mar 2022 16:43:52	U	SMPL
CCV56 - 1	CCV56	19416	4.8260	-	7470A	PPB	15 Mar 2022 16:51:58	U	SMPL
CCB56 - 1	CCB56	150	-0.0518	-	7470A	PPB	15 Mar 2022 16:54:14	U	SMPL
N1918-14LX5 - 1		6	-0.0883	-	7470A	PPB	15 Mar 2022 16:56:32	U	SMPL
N1918-14A - 1		14229	3.5127	-	7470A	PPB	15 Mar 2022 16:58:51	U	SMPL
N1893-07LX5 - 1		-17	-0.0941	-	7470A	PPB	15 Mar 2022 17:01:08	U	SMPL
N1893-07A - 1		14530	3.5889	-	7470A	PPB	15 Mar 2022 17:03:26	U	SMPL
N1905-12LX5 - 1		-60	-0.1050	-	7470A	PPB	15 Mar 2022 17:05:43	U	SMPL
N1905-12A - 1		14033	3.4631	-	7470A	PPB	15 Mar 2022 17:08:01	U	SMPL
CCV57 - 1	CCV57	20417	5.0794	-	7470A	PPB	15 Mar 2022 17:15:29	U	SMPL
CCB57 - 1	CCB57	96	-0.0655	-	7470A	PPB	15 Mar 2022 17:17:46	U	SMPL