

LB119159


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

42889



A= 0.0000e+000

B= 2.3265e-004

C= 7.1358e-002

Rho= 0.9998689

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	Std
0.0	0.000	0.069	0.069	-11	0.000	-11					-
0.2	0.200	0.235	0.035	702	0.0 %	702					18
2.5	2.500	2.410	-0.090	10054	0.0 %	10054					-4
5.0	5.000	4.929	-0.071	20882	0.0 %	20882					-1
7.5	7.500	7.507	0.007	31963	0.0 %	31963					0
10.0	10.000	10.049	0.049	42889	0.0 %	42889					0

LB119159 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 15 Mar 2022 10:56:57

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SD	-11	-	0.0000	7471B	PPB	15 Mar 2022 11:06:22	S	Std
0.2 - 1	50-2	702	-	0.2000	7471B	PPB	15 Mar 2022 11:08:39	S	Std
2.5 - 1	52-5	10054	-	2.5000	7471B	PPB	15 Mar 2022 11:10:56	S	Std
5.0 - 1	55	20882	-	5.0000	7471B	PPB	15 Mar 2022 11:13:14	S	Std
7.5 - 1	57-5	31963	-	7.5000	7471B	PPB	15 Mar 2022 11:15:34	S	Std
10.0 - 1	510	42889	-	10.0000	7471B	PPB	15 Mar 2022 11:20:31	S	Std
ICV32 - 1	ICV32	18014	4.2622	-7471B	PPB	15 Mar 2022 11:23:12	U	SMPL	
ICB32 - 1	ICB32	-7	0.0697	-7471B	PPB	15 Mar 2022 11:28:03	U	SMPL	
CCV51 - 1	CCV51	22353	5.2717	-7471B	PPB	15 Mar 2022 11:30:18	U	SMPL	
CCB51 - 1	CCB51	-130	0.0411	-7471B	PPB	15 Mar 2022 11:32:34	U	SMPL	
CRA - 1	CRA	738	0.2431	-7471B	PPB	15 Mar 2022 11:37:45	U	SMPL	
HighStd - 1	HighStd	45196	10.5860	-7471B	PPB	15 Mar 2022 11:40:00	U	SMPL	
ChkStd - 1	ChkStd	33003	7.7494	-7471B	PPB	15 Mar 2022 11:42:16	U	SMPL	
PB143367BL - 1	PBS	-185	0.0283	-7471B	PPB	15 Mar 2022 11:44:37	U	SMPL	
PB143367BS - 1	LCSS	16981	4.0219	-7471B	PPB	15 Mar 2022 11:49:41	U	SMPL	
N1888-01 - 1	SS-1(PILE-1)	3863	0.9701	-7471B	PPB	15 Mar 2022 11:51:57	U	SMPL	
N1888-02 - 1	SS-2(PILE-2)	2836	0.7311	-7471B	PPB	15 Mar 2022 11:54:17	U	SMPL	
N1888-03 - 1	SS-3(PILE-3)	2885	0.7425	-7471B	PPB	15 Mar 2022 11:56:35	U	SMPL	
N1905-11 - 1	WC-14A-1C	25129	5.9175	-7471B	PPB	15 Mar 2022 11:58:54	U	SMPL	
N1905-13 - 1	WC-14A-1A-3	6877	1.6713	-7471B	PPB	15 Mar 2022 12:01:12	U	SMPL	
CCV52 - 1	CCV52	22250	5.2477	-7471B	PPB	15 Mar 2022 12:03:31	U	SMPL	
CCB52 - 1	CCB52	-92	0.0500	-7471B	PPB	15 Mar 2022 12:05:48	U	SMPL	
N1905-15 - 1	WC-14A-1A-1	26351	6.2018	-7471B	PPB	15 Mar 2022 12:08:07	U	SMPL	
N1905-17 - 1	WC-14A-1A-4	723229	HIGH	-7471B	PPB	15 Mar 2022 12:10:22	X		
N1905-19 - 1	WC-14A-1A-2	13877	3.2998	-7471B	PPB	15 Mar 2022 12:12:13	U	SMPL	
N1908-01 - 1	WC-8-9-3	61733	14.4333	-7471B	PPB	15 Mar 2022 12:14:49	U	SMPL	
N1908-03 - 1	WC-8-9-2	12402	2.9566	-7471B	PPB	15 Mar 2022 12:17:08	U	SMPL	
N1908-05 - 1	WC-8-9-1	17891	4.2336	-7471B	PPB	15 Mar 2022 12:19:36	U	SMPL	
N1908-07 - 1	WC-14A-2	39604	9.2850	-7471B	PPB	15 Mar 2022 12:21:55	U	SMPL	
N1908-09 - 1	WC-14A-1b	12166	2.9017	-7471B	PPB	15 Mar 2022 12:24:15	U	SMPL	
N1922-01 - 1	7PH-1-7.5-8	1736	0.4752	-7471B	PPB	15 Mar 2022 12:26:39	U	SMPL	
N1922-01DUP - 1	7PH-1-7.5-8DUP	1816	0.4938	-7471B	PPB	15 Mar 2022 12:28:57	U	SMPL	
CCV53 - 1	CCV53	22840	5.3850	-7471B	PPB	15 Mar 2022 12:31:13	U	SMPL	
CCB53 - 1	CCB53	-194	0.0262	-7471B	PPB	15 Mar 2022 12:33:29	U	SMPL	
N1922-01MS - 1	7PH-1-7.5-8MS	19140	4.5242	-7471B	PPB	15 Mar 2022 12:35:47	U	SMPL	
N1922-01MSD - 1	7PH-1-7.5-8MSD	18755	4.4346	-7471B	PPB	15 Mar 2022 12:38:03	U	SMPL	
N1922-02 - 1	7PH-2-8.5-9	926	0.2868	-7471B	PPB	15 Mar 2022 12:40:22	U	SMPL	
N1944-01 - 1	CHRT-27154	2155	0.5727	-7471B	PPB	15 Mar 2022 12:42:41	U	SMPL	
N1944-03 - 1	CHRT-24449	3863	0.9701	-7471B	PPB	15 Mar 2022 12:44:57	U	SMPL	
N1944-05 - 1	CHRT-26215	2898	0.7456	-7471B	PPB	15 Mar 2022 12:47:14	U	SMPL	
N1922-01LX5 - 1		348	0.1523	-7471B	PPB	15 Mar 2022 12:49:32	U	SMPL	
N1922-01A - 1		18274	4.3227	-7471B	PPB	15 Mar 2022 12:51:50	U	SMPL	
N1905-17DLX10 - 1		151492	35.3153	-7471B	PPB	15 Mar 2022 12:54:07	U	SMPL	
N1908-01DLX5 - 1		12431	2.9634	-7471B	PPB	15 Mar 2022 12:56:25	U	SMPL	
CCV54 - 1	CCV54	20075	4.7417	-7471B	PPB	15 Mar 2022 13:07:55	U	SMPL	
CCB54 - 1	CCB54	-50	0.0597	-7471B	PPB	15 Mar 2022 13:10:12	U	SMPL	
N1905-17DL2X100 - 1		18064	4.2739	-7471B	PPB	15 Mar 2022 13:12:31	U	SMPL	
CCV55 - 1	CCV55	22372	5.2761	-7471B	PPB	15 Mar 2022 13:17:46	U	SMPL	
CCB55 - 1	CCB55	-147	0.0372	-7471B	PPB	15 Mar 2022 13:20:01	U	SMPL	