

LB119175


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

44561



A= 0.0000e+000

B= 2.2376e-004

C= 2.9105e-002

Rho= 0.9998708

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/01
0.0	0.000	0.036	0.036	30	0.000	30					1
0.2	0.200	0.204	0.004	781	0.0 %	781					2
2.5	2.500	2.504	0.004	11062	0.0 %	11062					0
5.0	5.000	4.881	-0.119	21684	0.0 %	21684					-2
7.5	7.500	7.575	0.075	33724	0.0 %	33724					-5
10.0	10.000	10.000	-0.000	44561	0.0 %	44561					0

LB119175 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 16 Mar 2022 10:02:39				Date	Type	Type
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units				
0.0 - 1	50	30	-	0.0000	7471B	PPB		16 Mar 2022 10:11:07	S	Std
0.2 - 1	50.2	781	-	0.2000	7471B	PPB		16 Mar 2022 10:13:25	S	Std
2.5 - 1	52.5	11062	-	2.5000	7471B	PPB		16 Mar 2022 10:15:42	S	Std
5.0 - 1	55	21684	-	5.0000	7471B	PPB		16 Mar 2022 10:17:59	S	Std
7.5 - 1	57.5	33724	-	7.5000	7471B	PPB		16 Mar 2022 10:20:18	S	Std
10.0 - 1	810	44561	-	10.0000	7471B	PPB		16 Mar 2022 10:22:38	S	Std
ICV33 - 1	ICV33	16448	3.7095	-	7471B	PPB		16 Mar 2022 10:25:27	U	SMPL
ICB33 - 1	ICB33	-180	-0.0112	-	7471B	PPB		16 Mar 2022 10:27:43	U	SMPL
CCV58 - 1	CCV58	21462	4.8314	-	7471B	PPB		16 Mar 2022 10:30:01	U	SMPL
CCB58 - 1	CCB58	-176	-0.0103	-	7471B	PPB		16 Mar 2022 10:32:16	U	SMPL
CRA - 1	CRA	815	0.2115	-	7471B	PPB		16 Mar 2022 10:34:35	U	SMPL
HighStd - 1	HighStd	45603	10.2331	-	7471B	PPB		16 Mar 2022 10:36:50	U	SMPL
ChkStd - 1	ChkStd	31869	7.1600	-	7471B	PPB		16 Mar 2022 10:39:06	U	SMPL
PB143383BL - 1	PBS	-184	-0.0121	-	7471B	PPB		16 Mar 2022 10:41:29	U	SMPL
PB143383BS - 1	LCSS	18457	4.1590	-	7471B	PPB		16 Mar 2022 10:43:50	U	SMPL
N1942-05 - 1	S-DRUM-C	4182	0.9649	-	7471B	PPB		16 Mar 2022 10:46:07	U	SMPL
N1943-01 - 1	B1	290	0.0940	-	7471B	PPB		16 Mar 2022 10:48:26	U	SMPL
N1943-01DUP - 1	B1DUP	278	0.0913	-	7471B	PPB		16 Mar 2022 10:50:45	U	SMPL
N1943-01MS - 1	B1MS	20656	4.6510	-	7471B	PPB		16 Mar 2022 10:53:02	U	SMPL
N1943-01MSD - 1	B1MSD	20447	4.6043	-	7471B	PPB		16 Mar 2022 10:55:17	U	SMPL
CCV59 - 1	CCV59	22482	5.0596	-	7471B	PPB		16 Mar 2022 10:57:36	U	SMPL
CCB59 - 1	CCB59	-133	-0.0007	-	7471B	PPB		16 Mar 2022 10:59:55	U	SMPL
N1943-02 - 1	B2	370	0.1119	-	7471B	PPB		16 Mar 2022 11:02:13	U	SMPL
N1943-03 - 1	TA1	1505	0.3659	-	7471B	PPB		16 Mar 2022 11:04:29	U	SMPL
N1943-04 - 1	TB1	3137	0.7310	-	7471B	PPB		16 Mar 2022 11:06:44	U	SMPL
N1943-05 - 1	TB2	3692	0.8552	-	7471B	PPB		16 Mar 2022 11:09:00	U	SMPL
N1943-06 - 1	DF1	1454	0.3544	-	7471B	PPB		16 Mar 2022 11:11:17	U	SMPL
N1948-01 - 1	COMP-4	786	0.2050	-	7471B	PPB		16 Mar 2022 11:13:34	U	SMPL
N1948-02 - 1	COMP-5	1303	0.3207	-	7471B	PPB		16 Mar 2022 11:15:51	U	SMPL
N1948-03 - 1	COMP-6	847	0.2186	-	7471B	PPB		16 Mar 2022 11:18:08	U	SMPL
N1949-09 - 1	CP-BH2-031422-0-2	1775	0.4263	-	7471B	PPB		16 Mar 2022 11:20:25	U	SMPL
N1949-10 - 1	CP-BH2-031422-2-4	364	0.1106	-	7471B	PPB		16 Mar 2022 11:22:41	U	SMPL
CCV60 - 1	CCV60	22278	5.0140	-	7471B	PPB		16 Mar 2022 11:24:57	U	SMPL
CCB60 - 1	CCB60	-124	0.0014	-	7471B	PPB		16 Mar 2022 11:27:13	U	SMPL
N1949-11 - 1	CP-BH2-031422-4-6	141	0.0607	-	7471B	PPB		16 Mar 2022 11:29:33	U	SMPL
N1949-12 - 1	CP-BH1-031422-0-2	1394	0.3410	-	7471B	PPB		16 Mar 2022 11:31:48	U	SMPL
N1949-13 - 1	CP-BH1-031422-2-4	870	0.2238	-	7471B	PPB		16 Mar 2022 11:34:04	U	SMPL
N1949-14 - 1	CP-BH1-031422-4-6	63	0.0432	-	7471B	PPB		16 Mar 2022 11:36:20	U	SMPL
N1949-15 - 1	CP-BH1-031422-4-6-DUP	264	0.0882	-	7471B	PPB		16 Mar 2022 11:38:36	U	SMPL
N1953-01 - 1	RO-RBR-200033	7965	1.8113	-	7471B	PPB		16 Mar 2022 11:40:52	U	SMPL
N1943-01LX5 - 1		116	0.0551	-	7471B	PPB		16 Mar 2022 11:47:55	U	SMPL
N1943-01A - 1		19749	4.4481	-	7471B	PPB		16 Mar 2022 11:50:13	U	SMPL
CCV61 - 1	CCV61	22598	5.0856	-	7471B	PPB		16 Mar 2022 11:52:30	U	SMPL
CCB61 - 1	CCB61	-135	-0.0011	-	7471B	PPB		16 Mar 2022 11:54:49	U	SMPL