



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : mohammad

Run Number: LB119178

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	03/16/2022	11:20
2	CAL2	1	Water	NA	NA	20.3	7.00	03/16/2022	11:21
3	CAL3	1	Water	NA	NA	20.4	10.03	03/16/2022	11:22
4	ICV	1	Water	NA	NA	20.4	6.98	03/16/2022	11:25
5	CCV1	1	Water	NA	NA	20.4	2.01	03/16/2022	11:26
6	N1978-02	1	Water	NA	NA	20.2	4.64	03/16/2022	11:30
7	N1978-02DUP	1	Water	NA	NA	20.2	4.65	03/16/2022	11:31
8	CCV2	1	Water	NA	NA	20.3	12.03	03/16/2022	11:33

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N1979

WorkList ID : 157778

Department : Wet-Chemistry

Date : 03-16-2022 11:12:39

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/15/2022	Water	N1978-02	pH	Cool 4 deg C	TETR06	N22	TT-RW7A-IDWGW-20220315	03/15/2022	9040C

Date/Time 03-16-22 11:15

Raw Sample Received by: 7016001

Raw Sample Relinquished by: 7016001

Date/Time 03-16-22 12:00

Raw Sample Received by: 7016001

Raw Sample Relinquished by: 7016001