

LB119182


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

36075



A= 0.0000e+000

B= 2.7768e-004

C= -5.1082e-002

Rho= 0.9999205

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	U/D
0.0	0.000	-0.046	-0.046	17	0.000	17					
0.2	0.200	0.182	-0.018	840	0.0 %	840					9
2.5	2.500	2.539	0.039	9327	0.0 %	9327					2
5.0	5.000	5.085	0.085	18496	0.0 %	18496					2
7.5	7.500	7.474	-0.026	27100	0.0 %	27100					0
10.0	10.000	9.966	-0.034	36075	0.0 %	36075					0

LB119182 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 16 Mar 2022 12:32:13

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	17	-	0.0000	7470A	PPB	16 Mar 2022 13:24:40	S	Std
0.2 - 1	50.2	840	-	0.2000	7470A	PPB	16 Mar 2022 13:26:57	S	Std
2.5 - 1	52.5	9327	-	2.5000	7470A	PPB	16 Mar 2022 13:29:15	S	Std
5.0 - 1	55	18496	-	5.0000	7470A	PPB	16 Mar 2022 13:31:32	S	Std
7.5 - 1	57.5	27100	-	7.5000	7470A	PPB	16 Mar 2022 13:36:26	S	Std
10.0 - 1	510	36075	-	10.0000	7470A	PPB	16 Mar 2022 13:41:18	S	Std
ICV34 - 1	ICV34	13449	3.6835	-	7470A	PPB	16 Mar 2022 13:46:03	U	SMPL
ICB34 - 1	ICB34	16	-0.0466	-	7470A	PPB	16 Mar 2022 13:50:51	U	SMPL
CCV62 - 1	CCV62	18497	5.0852	-	7470A	PPB	16 Mar 2022 13:53:06	U	SMPL
CCB62 - 1	CCB62	13	-0.0475	-	7470A	PPB	16 Mar 2022 13:55:22	U	SMPL
CRA - 1	CRA	796	0.1700	-	7470A	PPB	16 Mar 2022 13:57:38	U	SMPL
HighStd - 1	HighStd	35437	9.7891	-	7470A	PPB	16 Mar 2022 13:59:54	U	SMPL
ChkStd - 1	ChkStd	25262	6.9637	-	7470A	PPB	16 Mar 2022 14:02:09	U	SMPL
PB143401BL - 1	PBW	44	-0.0389	-	7470A	PPB	16 Mar 2022 14:07:00	U	SMPL
PB143401BS - 1	LCSW	12800	3.5032	-	7470A	PPB	16 Mar 2022 14:09:15	U	SMPL
N1921-01 - 1	WATER	3959	1.0483	-	7470A	PPB	16 Mar 2022 14:11:32	U	SMPL
N1949-07 - 1	FIELD-BLANK	-25	-0.0580	-	7470A	PPB	16 Mar 2022 14:13:49	U	SMPL
N1950-01 - 1	31339	698	0.1427	-	7470A	PPB	16 Mar 2022 14:20:55	U	SMPL
N1950-01DUP - 1	31339DUP	706	0.1450	-	7470A	PPB	16 Mar 2022 14:24:54	U	SMPL
N1950-01MS - 1	31339MS	13332	3.6510	-	7470A	PPB	16 Mar 2022 14:30:17	U	SMPL
CCV63 - 1	CCV63	16702	4.5868	-	7470A	PPB	16 Mar 2022 14:32:32	U	SMPL
CCB63 - 1	CCB63	64	-0.0333	-	7470A	PPB	16 Mar 2022 14:36:45	U	SMPL
N1950-01MSD - 1	31339MSD	14512	3.9786	-	7470A	PPB	16 Mar 2022 14:39:00	U	SMPL
N1950-02 - 1	31340	4405	1.1721	-	7470A	PPB	16 Mar 2022 14:41:15	U	SMPL
N1950-08 - 1	31424	2472	0.6353	-	7470A	PPB	16 Mar 2022 14:43:33	U	SMPL
N1950-10 - 1	3222	179	-0.0014	-	7470A	PPB	16 Mar 2022 14:45:51	U	SMPL
N1952-04 - 1	WATER-PIPE	1662	0.4104	-	7470A	PPB	16 Mar 2022 14:48:07	U	SMPL
N1955-03 - 1	1335	-15	-0.0552	-	7470A	PPB	16 Mar 2022 14:50:23	U	SMPL
N1978-02 - 1	TT-RW7A-IDWGW-20220315	8	-0.0489	-	7470A	PPB	16 Mar 2022 14:55:48	U	SMPL
PB143402BL - 1	PBW	74	-0.0305	-	7470A	PPB	16 Mar 2022 14:58:05	U	SMPL
PB143402BS - 1	LCSW	16656	4.5740	-	7470A	PPB	16 Mar 2022 15:00:22	U	SMPL
N1923-01 - 1	33590	-43	-0.0630	-	7470A	PPB	16 Mar 2022 15:02:38	U	SMPL
CCV64 - 1	CCV64	18707	5.1435	-	7470A	PPB	16 Mar 2022 15:04:55	U	SMPL
CCB64 - 1	CCB64	15	-0.0469	-	7470A	PPB	16 Mar 2022 15:07:11	U	SMPL
N1923-02 - 1	33591	35199	9.7231	-	7470A	PPB	16 Mar 2022 15:09:28	U	SMPL
N1923-03 - 1	33592	257	0.0203	-	7470A	PPB	16 Mar 2022 15:11:44	U	SMPL
N1923-04 - 1	33593	187	0.0008	-	7470A	PPB	16 Mar 2022 15:14:03	U	SMPL
N1923-05 - 1	33594	3457	0.9089	-	7470A	PPB	16 Mar 2022 15:16:19	U	SMPL
N1923-06 - 1	33595	749	0.1569	-	7470A	PPB	16 Mar 2022 15:18:34	U	SMPL
N1923-07 - 1	33596	705	0.1447	-	7470A	PPB	16 Mar 2022 15:20:51	U	SMPL
N1923-08 - 1	33597	708	0.1455	-	7470A	PPB	16 Mar 2022 15:23:08	U	SMPL
N1923-09 - 1	33598	776	0.1644	-	7470A	PPB	16 Mar 2022 15:25:25	U	SMPL
N1923-10 - 1	33599	1750	0.4349	-	7470A	PPB	16 Mar 2022 15:27:42	U	SMPL
N1923-11 - 1	33600	581	0.1103	-	7470A	PPB	16 Mar 2022 15:29:58	U	SMPL
CCV65 - 1	CCV65	19113	5.2563	-	7470A	PPB	16 Mar 2022 15:38:29	U	SMPL
CCB65 - 1	CCB65	88	-0.0266	-	7470A	PPB	16 Mar 2022 15:43:21	U	SMPL
N1923-12 - 1	33603	498	0.0872	-	7470A	PPB	16 Mar 2022 15:45:37	U	SMPL
N1923-12DUP - 1	33603DUP	505	0.0891	-	7470A	PPB	16 Mar 2022 15:47:53	U	SMPL
N1923-12MS - 1	33603MS	15241	4.1811	-	7470A	PPB	16 Mar 2022 15:50:08	U	SMPL
N1923-12MSD - 1	33603MSD	15108	4.1441	-	7470A	PPB	16 Mar 2022 15:52:24	U	SMPL
PB143307TB - 1	PB143307TB	-75	-0.0719	-	7470A	PPB	16 Mar 2022 15:54:42	U	SMPL
N1950-01LX5 - 1		199	0.0042	-	7470A	PPB	16 Mar 2022 15:56:59	U	SMPL
N1950-01A - 1		12817	3.5080	-	7470A	PPB	16 Mar 2022 15:59:16	U	SMPL
N1923-12LX5 - 1		56	-0.0355	-	7470A	PPB	16 Mar 2022 16:01:33	U	SMPL
N1923-12A - 1		17054	4.6845	-	7470A	PPB	16 Mar 2022 16:03:51	U	SMPL
CCV66 - 1	CCV66	17900	4.9194	-	7470A	PPB	16 Mar 2022 16:06:07	U	SMPL
CCB66 - 1	CCB66	85	-0.0275	-	7470A	PPB	16 Mar 2022 16:11:03	U	SMPL