



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : mohammad

Run Number: LB119190

Slope : 98.7

BalanceID: WC SC-6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	03/16/2022	15:15
2	CAL2	1	Water	NA	NA	20.3	7.00	03/16/2022	15:16
3	CAL3	1	Water	NA	NA	20.3	10.03	03/16/2022	15:19
4	ICV	1	Water	NA	NA	20.4	7.02	03/16/2022	15:20
5	CCV1	1	Water	NA	NA	20.3	2.01	03/16/2022	15:22
6	N1985-02	1	Solid	20.03	20	20.3	6.93	03/16/2022	15:25
7	N1985-02DUP	1	Solid	20.04	20	20.1	6.94	03/16/2022	15:26
8	N1985-05	1	Solid	20.02	20	20.1	7.70	03/16/2022	15:30
9	CCV2	1	Water	NA	NA	20.4	12.03	03/16/2022	15:22

18 119190

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY N1985 WorkList ID : 157790 Department : Wet-Chemistry Date : 03-16-2022 15:07:00

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/14/2022	Solid	N1985-02	Corrosivity	Cool 4 deg C	NOB103	N33	VACLOT-SOIL-1	03/14/2022	9045D
03/14/2022	Solid	N1985-05	Corrosivity	Cool 4 deg C	NOB103	N33	VACLOT-SOIL-2	03/14/2022	9045D

Date/Time 03/16/22 15:10
Raw Sample Received by: JH (SW)
Raw Sample Relinquished by: JH (SW)

Date/Time 03-16-22 16:30
Raw Sample Received by: JH (SW)
Raw Sample Relinquished by: JH (SW)