

LB 119198


245.1

INSTRUMENT ID: CVK

Linear

 μ Abs.:

34865



A= 0.0000e+000

B= 2.8863e-004

C= -2.4529e-003

Rho= 0.9998598

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	1/2D
0.0	0.000	0.002	0.002	14	0.000	14					
0.05	0.050	0.052	0.002	187	0.0 %	187	0				4
0.2	0.200	0.205	0.005	719	0.0 %	719					2
2.5	2.500	2.485	-0.015	8618	0.0 %	8618					1
5.0	5.000	5.077	0.077	17600	0.0 %	17600					2
7.5	7.500	7.369	-0.131	25540	0.0 %	25540					2
10.0	10.000	10.060	0.060	34865	0.0 %	34865					1

LB119198 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 17 Mar 2022 10:30:59

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	14	-	0.0000	245.1	PPB	17 Mar 2022 10:32:55	S	Std
0.05 - 1	51.5	187	-	0.0500	245.1	PPB	17 Mar 2022 10:35:12	S	Std
0.2 - 1	50.2	719	-	0.2000	245.1	PPB	17 Mar 2022 10:37:30	S	Std
2.5 - 1	52.5	8618	-	2.5000	245.1	PPB	17 Mar 2022 10:39:47	S	Std
5.0 - 1	55	17600	-	5.0000	245.1	PPB	17 Mar 2022 10:42:04	S	Std
7.5 - 1	57.5	25540	-	7.5000	245.1	PPB	17 Mar 2022 10:44:21	S	Std
10.0 - 1	510	34865	-	10.0000	245.1	PPB	17 Mar 2022 10:56:58	S	Std
ICV35 - 1	ICV35	13571	3.9145	-245.1		PPB	17 Mar 2022 11:00:26	U	SMPL
ICB35 - 1	ICB35	-56	-0.0186	-245.1		PPB	17 Mar 2022 11:02:41	U	SMPL
CCV67 - 1	CCV67	18276	5.2725	-245.1		PPB	17 Mar 2022 11:04:57	U	SMPL
CCB67 - 1	CCB67	-79	-0.0253	-245.1		PPB	17 Mar 2022 11:07:16	U	SMPL
CRA - 1	CRA	187	0.0515	-245.1		PPB	17 Mar 2022 11:09:32	U	SMPL
HighStd - 1	HighStd	34836	10.0521	-245.1		PPB	17 Mar 2022 11:11:47	U	SMPL
ChkStd - 1	ChkStd	25408	7.3310	-245.1		PPB	17 Mar 2022 11:14:03	U	SMPL
PB143424BL - 1	PBW	-103	-0.0322	-245.1		PPB	17 Mar 2022 11:16:21	U	SMPL
PB143424BS - 1	LCSW	13437	3.8758	-245.1		PPB	17 Mar 2022 11:18:39	U	SMPL
N1824-02 - 1	SPRING-2022	397	0.1121	-245.1		PPB	17 Mar 2022 11:21:00	U	SMPL
N1877-01 - 1	15220012-7	6	-0.0007	-245.1		PPB	17 Mar 2022 11:23:18	U	SMPL
N1877-01DUP - 1	15220012-7DUP	52	0.0126	-245.1		PPB	17 Mar 2022 11:25:35	U	SMPL
N1877-01MS - 1	15220012-7MS	12152	3.5049	-245.1		PPB	17 Mar 2022 11:27:54	U	SMPL
N1877-01MSD - 1	15220012-7MSD	14121	4.0732	-245.1		PPB	17 Mar 2022 11:30:09	U	SMPL
CCV68 - 1	CCV68	18085	5.2173	-245.1		PPB	17 Mar 2022 11:32:27	U	SMPL
CCB68 - 1	CCB68	-91	-0.0287	-245.1		PPB	17 Mar 2022 11:34:45	U	SMPL
N1881-01 - 1	32220008-3	7	-0.0004	-245.1		PPB	17 Mar 2022 11:37:02	U	SMPL
N1935-02 - 1	EFF-WW	-6	-0.0042	-245.1		PPB	17 Mar 2022 11:39:17	U	SMPL
N1937-02 - 1	OB-722	5	-0.0010	-245.1		PPB	17 Mar 2022 11:41:33	U	SMPL
N1938-02 - 1	TI-633	42	0.0097	-245.1		PPB	17 Mar 2022 11:43:49	U	SMPL
N1877-01LX5 - 1		3	-0.0016	-245.1		PPB	17 Mar 2022 11:46:05	U	SMPL
N1877-01A - 1		13634	3.9327	-245.1		PPB	17 Mar 2022 11:48:22	U	SMPL
CCV69 - 1	CCV69	17536	5.0589	-245.1		PPB	17 Mar 2022 11:50:39	U	SMPL
CCB69 - 1	CCB69	-55	-0.0183	-245.1		PPB	17 Mar 2022 11:52:58	U	SMPL