


LB119205

7471B

INSTRUMENT ID: CV1

 μ Abs.:

47592



Linear

A= 0.0000e+000

B= 2.0659e-004

C= -3.6242e-002

Rho= 0.9993627

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/d
0.0	0.000	-0.036	-0.036	3	0.000	3					1
0.2	0.200	0.194	-0.006	1115	0.0 %	1115					3
2.5	2.500	2.450	-0.050	12034	0.0 %	12034					2
5.0	5.000	5.062	0.062	24677	0.0 %	24677					1
7.5	7.500	7.734	0.234	37612	0.0 %	37612					3
10.0	10.000	9.796	-0.204	47592	0.0 %	47592					2

LB119205 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 17 Mar 2022 12:41:49

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SO	3	-	0.00007471B	PPB		17 Mar 2022 13:24:38	S	Std
0.2 - 1	SO.2	1115	-	0.20007471B	PPB		17 Mar 2022 13:26:55	S	Std
2.5 - 1	SO.5	12034	-	2.50007471B	PPB		17 Mar 2022 13:29:13	S	Std
5.0 - 1	SO	24677	-	5.00007471B	PPB		17 Mar 2022 13:31:30	S	Std
7.5 - 1	SO.5	37612	-	7.50007471B	PPB		17 Mar 2022 13:33:49	S	Std
10.0 - 1	SO	47592	-	10.00007471B	PPB		17 Mar 2022 13:38:44	S	Std
ICV36 - 1	ICV36	19241	3.9388	-7471B	PPB		17 Mar 2022 13:43:39	U	SMPL
ICB36 - 1	ICB36	-126	-0.0623	-7471B	PPB		17 Mar 2022 13:45:55	U	SMPL
CCV70 - 1	CCV70	25431	5.2176	-7471B	PPB		17 Mar 2022 13:48:13	U	SMPL
CCB70 - 1	CCB70	-164	-0.0701	-7471B	PPB		17 Mar 2022 13:50:28	U	SMPL
CRA - 1	CRA	1129	0.1970	-7471B	PPB		17 Mar 2022 13:52:46	U	SMPL
HighStd - 1	HighStd	49359	10.1609	-7471B	PPB		17 Mar 2022 13:55:02	U	SMPL
ChkStd - 1	ChkStd	34841	7.1616	-7471B	PPB		17 Mar 2022 13:57:17	U	SMPL
PB143405BL - 1	PBS	-167	-0.0707	-7471B	PPB		17 Mar 2022 13:59:40	U	SMPL
PB143405BS - 1	LCSS	19022	3.8935	-7471B	PPB		17 Mar 2022 14:02:02	U	SMPL
N1923-01DLX10 - 1	33590	366	0.0394	-7471B	PPB		17 Mar 2022 14:04:18	U	SMPL
N1923-02DLX100 - 1	33591	783147	HIGH	-7471B	PPB		17 Mar 2022 14:09:21	X	
N1923-03DLX10 - 1	33592	6290	1.2632	-7471B	PPB		17 Mar 2022 14:11:13	U	SMPL
N1923-04DLX10 - 1	33593	1918	0.3600	-7471B	PPB		17 Mar 2022 14:14:06	U	SMPL
N1923-05DLX10 - 1	33594	84192	17.3570	-7471B	PPB		17 Mar 2022 14:16:24	U	SMPL
CCV71 - 1	CCV71	25492	5.2302	-7471B	PPB		17 Mar 2022 14:18:39	U	SMPL
CCB71 - 1	CCB71	-232	-0.0842	-7471B	PPB		17 Mar 2022 14:21:05	U	SMPL
N1923-06DLX10 - 1	33595	13980	2.8519	-7471B	PPB		17 Mar 2022 14:23:24	U	SMPL
N1923-07DLX10 - 1	33596	1881	0.3524	-7471B	PPB		17 Mar 2022 14:25:39	U	SMPL
N1923-08DLX10 - 1	33597	6105	1.2250	-7471B	PPB		17 Mar 2022 14:27:58	U	SMPL
N1923-09DLX10 - 1	33598	1852	0.3464	-7471B	PPB		17 Mar 2022 14:30:14	U	SMPL
N1923-10DLX10 - 1	33599	15510	3.1680	-7471B	PPB		17 Mar 2022 14:32:31	U	SMPL
N1923-11DLX10 - 1	33600	1282	0.2286	-7471B	PPB		17 Mar 2022 14:34:48	U	SMPL
N1923-12DLX10 - 1	33603	2653	0.5118	-7471B	PPB		17 Mar 2022 14:37:07	U	SMPL
N1954-01 - 1	CL-02-031522	2608	0.5025	-7471B	PPB		17 Mar 2022 14:39:24	U	SMPL
N1954-01DUP - 1	CL-02-031522DUP	2548	0.4902	-7471B	PPB		17 Mar 2022 14:41:42	U	SMPL
N1954-01MS - 1	CL-02-031522MS	21629	4.4321	-7471B	PPB		17 Mar 2022 14:43:58	U	SMPL
CCV72 - 1	CCV72	25368	5.2045	-7471B	PPB		17 Mar 2022 14:46:15	U	SMPL
CCB72 - 1	CCB72	-160	-0.0693	-7471B	PPB		17 Mar 2022 14:48:34	U	SMPL
N1954-01MSD - 1	CL-02-031522MSD	23511	4.8209	-7471B	PPB		17 Mar 2022 14:50:53	U	SMPL
N1956-01 - 1	TPK-1	873	0.1441	-7471B	PPB		17 Mar 2022 14:53:09	U	SMPL
N1956-02 - 1	TPK-2	2224	0.4232	-7471B	PPB		17 Mar 2022 14:55:28	U	SMPL
N1956-03 - 1	TPK-3	1021	0.1747	-7471B	PPB		17 Mar 2022 14:57:44	U	SMPL
N1956-04 - 1	TPK-4	1712	0.3174	-7471B	PPB		17 Mar 2022 15:00:00	U	SMPL
N1954-01LX5 - 1		488	0.0646	-7471B	PPB		17 Mar 2022 15:02:16	U	SMPL
N1954-01A - 1		22330	4.5769	-7471B	PPB		17 Mar 2022 15:04:33	U	SMPL
N1923-02DL2X1000 - 1		195258	40.3022	-7471B	PPB		17 Mar 2022 15:06:50	U	SMPL
N1923-02DL2X10000 - 1		22944	4.7038	-7471B	PPB		17 Mar 2022 15:10:21	U	SMPL
N1923-05DL2X50 - 1		16384	3.3485	-7471B	PPB		17 Mar 2022 15:12:36	U	SMPL
CCV73 - 1	CCV73	25668	5.2665	-7471B	PPB		17 Mar 2022 15:14:55	U	SMPL
CCB73 - 1	CCB73	-151	-0.0674	-7471B	PPB		17 Mar 2022 15:17:13	U	SMPL