



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB119210

Slope : 98.8

BalanceID: WC SC-6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	03/17/2022	13:40
2	CAL2	1	Water	NA	NA	20.3	7.00	03/17/2022	13:41
3	CAL3	1	Water	NA	NA	20.4	10.03	03/17/2022	13:42
4	ICV	1	Water	NA	NA	20.5	7.00	03/17/2022	13:45
5	CCV1	1	Water	NA	NA	20.6	2.01	03/17/2022	13:46
6	N2007-01	1	Solid	20.03	20	20.7	7.36	03/17/2022	13:47
7	N2007-01DUP	1	Solid	20.04	20	20.8	7.37	03/17/2022	13:48
8	N2007-02	1	Solid	20.01	20	20.6	8.76	03/17/2022	13:50
9	N2007-03	1	Solid	20.03	20	20.5	7.80	03/17/2022	13:52
10	N2009-01	1	Solid	20.03	20	20.1	5.03	03/17/2022	14:00
11	CCV2	1	Water	NA	NA	20.6	12.03	03/17/2022	14:01

WORKLIST(Hardcopy Internal Chain)

VB19210

WorkList Name : PH N2007 WorkList ID : 157834 Department : Wet-Chemistry Date : 03-17-2022 13:33:52

Due Date	Matrix	Sample	Test	Raw Sample			Collect Date	Method
				Preservative	Customer	Storage Location		
03/17/2022	Solid	N2007-01	pH	Cool 4 deg C	DASK01	E11 SP1	03/17/2022	9045D
03/17/2022	Solid	N2007-02	pH	Cool 4 deg C	DASK01	E11 SP2	03/17/2022	9045D
03/17/2022	Solid	N2007-03	pH	Cool 4 deg C	DASK01	E11 SP3	03/17/2022	9045D
03/17/2022	Solid	N2009-01	pH	Cool 4 deg C	PSEG03	E11 13563	03/17/2022	9045D

Date/Time 03-17-22 13:35
 Raw Sample Received by: JH
 Raw Sample Relinquished by: JH

Date/Time 03-17-22 17:00
 Raw Sample Received by: JH
 Raw Sample Relinquished by: JH