



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB119211

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	03/17/2022	14:30
2	CAL2	1	Water	NA	NA	20.3	7.00	03/17/2022	14:31
3	CAL3	1	Water	NA	NA	20.3	10.03	03/17/2022	14:32
4	ICV	1	Water	NA	NA	20.4	6.99	03/17/2022	14:35
5	CCV1	1	Water	NA	NA	20.4	2.01	03/17/2022	14:36
6	N1974-13	1	Water	NA	NA	20.1	7.53	03/17/2022	14:37
7	N1992-01	1	Water	NA	NA	20.2	11.71	03/17/2022	14:40
8	N2009-03	1	Water	NA	NA	20.6	5.78	03/17/2022	14:42
9	N2009-05	1	Water	NA	NA	20.4	6.83	03/17/2022	14:45
10	N2009-06	1	Water	NA	NA	20.1	9.08	03/17/2022	14:47
11	N2013-01	1	Water	NA	NA	20.1	7.26	03/17/2022	14:50
12	N2013-01DUP	1	Water	NA	NA	20.2	7.27	03/17/2022	14:51
13	CCV2	1	Water	NA	NA	20.6	12.03	03/17/2022	15:00

119211

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N2031 WorkList ID : 157837 Department : Wet-Chemistry Date : 03-17-2022 15:06:35

Due Date	Matrix	Sample	Test	Raw Sample			Collect Date	Method
				Preservative	Customer	Storage Location		
03/15/2022	Water	N1974-13	pH	Cool 4 deg C	URSC01	E11 B-3	03/15/2022	9040C
03/16/2022	Water	N1992-01	pH	Cool 4 deg C	PSEG03	M11 N-37853	03/16/2022	9040C
03/17/2022	Water	N2009-03	pH	Cool 4 deg C	PSEG03	E11 31722	03/17/2022	9040C
03/17/2022	Water	N2009-05	pH	Cool 4 deg C	PSEG03	E11 31405	03/17/2022	9040C
03/17/2022	Water	N2009-06	pH	Cool 4 deg C	PSEG03	E11 40453	03/17/2022	9040C
03/15/2022	Water	N2013-01	pH	Cool 4 deg C	STER01	N11 POOL-WATER	03/15/2022	9040C

Date/Time 03-17-22 13:35
Raw Sample Received by: JO (WC)
Raw Sample Relinquished by: [Signature]

Date/Time 03-17-22 17:10
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: JO (WC)