

LB119216


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

49033



A= 0.0000e+000

B= 2.0291e-004

C= -5.6187e-002

Rho= 0.9992064

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	9/21
0.0	0.000	-0.041	-0.041	77	0.000	77					-18
0.2	0.200	0.165	-0.035	1092	0.0 %	1092					-3
2.5	2.500	2.422	-0.078	12212	0.0 %	12212					6
5.0	5.000	5.324	0.324	26516	0.0 %	26516					-1
7.5	7.500	7.436	-0.064	36922	0.0 %	36922					-1
10.0	10.000	9.893	-0.107	49033	0.0 %	49033					

LB119216 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 18 Mar 2022 09:20:24

Sample ID	Extended ID	µ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	77	-	0.00007471B	PPB	18 Mar 2022 09:45:16	S	Std
0.2 - 1	50.2	1092	-	0.20007471B	PPB	18 Mar 2022 09:47:34	S	Std
2.5 - 1	52.5	12212	-	2.50007471B	PPB	18 Mar 2022 09:49:51	S	Std
5.0 - 1	55	26516	-	5.00007471B	PPB	18 Mar 2022 09:52:08	S	Std
7.5 - 1	59.1	36922	-	7.50007471B	PPB	18 Mar 2022 09:54:26	S	Std
10.0 - 1	310	49033	-	10.00007471B	PPB	18 Mar 2022 10:02:06	S	Std
ICV37 - 1	ICV37	18877	3.7742	-7471B	PPB	18 Mar 2022 10:04:59	U	SMPL
ICB37 - 1	ICB37	64	-0.0432	-7471B	PPB	18 Mar 2022 10:07:14	U	SMPL
CCV74 - 1	CCV74	25589	5.1362	-7471B	PPB	18 Mar 2022 10:09:33	U	SMPL
CCB74 - 1	CCB74	38	-0.0485	-7471B	PPB	18 Mar 2022 10:11:51	U	SMPL
CRA - 1	CRA	1153	0.1778	-7471B	PPB	18 Mar 2022 10:14:09	U	SMPL
HighStd - 1	HighStd	50121	10.1141	-7471B	PPB	18 Mar 2022 10:16:25	U	SMPL
ChkStd - 1	ChkStd	34530	6.9504	-7471B	PPB	18 Mar 2022 10:18:41	U	SMPL
PB143427BL - 1	PBS	189	-0.0178	-7471B	PPB	18 Mar 2022 10:23:43	U	SMPL
PB143427BS - 1	LCSS	19102	3.8199	-7471B	PPB	18 Mar 2022 10:25:58	U	SMPL
N1958-01 - 1	C-1-0-1	470528	HIGH	-7471B	PPB	18 Mar 2022 10:28:15	X	
N1958-02 - 1	C-2-1-2	86112	17.4171	-7471B	PPB	18 Mar 2022 10:30:08	U	SMPL
N1958-03 - 1	C-3-2-3	216498	43.8743	-7471B	PPB	18 Mar 2022 10:32:37	U	SMPL
N1958-04 - 1	C-4-3-4	117572	23.8008	-7471B	PPB	18 Mar 2022 10:35:05	U	SMPL
N1958-05 - 1	C-5-4-5	197661	40.0520	-7471B	PPB	18 Mar 2022 10:37:41	U	SMPL
CCV75 - 1	CCV75	23423	4.6967	-7471B	PPB	18 Mar 2022 10:40:11	U	SMPL
CCB75 - 1	CCB75	-146	-0.0858	-7471B	PPB	18 Mar 2022 10:42:43	U	SMPL
N1958-06 - 1	C-6-5-6	325680	66.0288	-7471B	PPB	18 Mar 2022 10:45:03	U	SMPL
N1958-07 - 1	C-7-6-7	172323	34.9105	-7471B	PPB	18 Mar 2022 10:47:19	U	SMPL
N1958-08 - 1	C-8-7-8	127434	25.8019	-7471B	PPB	18 Mar 2022 10:49:58	U	SMPL
N1958-09 - 1	C-9-8-9	142745	28.9088	-7471B	PPB	18 Mar 2022 10:52:32	U	SMPL
N1958-10 - 1	C-10-9-10	237964	48.2300	-7471B	PPB	18 Mar 2022 10:55:03	U	SMPL
N1958-12 - 1	C-12-11-12	32273	6.4925	-7471B	PPB	18 Mar 2022 10:57:36	U	SMPL
N1958-13 - 1	C-13-12-13	34779	7.0010	-7471B	PPB	18 Mar 2022 11:00:15	U	SMPL
N1958-14 - 1	C-14-13-14	42101	8.4867	-7471B	PPB	18 Mar 2022 11:02:36	U	SMPL
N1958-15 - 1	C-15-14-15	17294	3.4530	-7471B	PPB	18 Mar 2022 11:04:58	U	SMPL
N1977-01 - 1	SB-12-20220315-16.0-17.0	10248	2.0233	-7471B	PPB	18 Mar 2022 11:07:20	U	SMPL
CCV76 - 1	CCV76	25299	5.0773	-7471B	PPB	18 Mar 2022 11:09:38	U	SMPL
CCB76 - 1	CCB76	-4	-0.0570	-7471B	PPB	18 Mar 2022 11:11:56	U	SMPL
N1977-01DUP - 1	SB-12-20220315-16.0-17.0DUP	10886	2.1527	-7471B	PPB	18 Mar 2022 11:14:15	U	SMPL
N1977-02 - 1	N1977-01MS	30410	6.1144	-7471B	PPB	18 Mar 2022 11:19:08	U	SMPL
N1977-03 - 1	N1977-01MSD	31524	6.3405	-7471B	PPB	18 Mar 2022 11:21:25	U	SMPL
N1977-04 - 1	SB-12-20220315-16.0-17.0-A	27661	5.5566	-7471B	PPB	18 Mar 2022 11:23:44	U	SMPL
N1977-05 - 1	SB-12-20220315-19.0-20.0	11530	2.2834	-7471B	PPB	18 Mar 2022 11:26:03	U	SMPL
N1977-06 - 1	SB-14-20220315-15.0-16.0	511148	HIGH	-7471B	PPB	18 Mar 2022 11:28:23	X	
N1977-07 - 1	SB-14-20220315-18.5-19.5	36812	7.4135	-7471B	PPB	18 Mar 2022 11:30:15	U	SMPL
PB143428BL - 1	PBS	-81	-0.0726	-7471B	PPB	18 Mar 2022 11:32:49	U	SMPL
PB143428BS - 1	LCSS	19865	3.9747	-7471B	PPB	18 Mar 2022 11:35:11	U	SMPL
N1962-01 - 1	C-16-15-16	37218	7.4959	-7471B	PPB	18 Mar 2022 11:37:27	U	SMPL
CCV77 - 1	CCV77	26884	5.3990	-7471B	PPB	18 Mar 2022 11:39:46	U	SMPL
CCB77 - 1	CCB77	-38	-0.0639	-7471B	PPB	18 Mar 2022 11:42:05	U	SMPL
N1962-02 - 1	C-17-16-17	38657	7.7879	-7471B	PPB	18 Mar 2022 11:44:24	U	SMPL
N1962-03 - 1	C-18-17-18	40386	8.1387	-7471B	PPB	18 Mar 2022 11:46:40	U	SMPL
N1962-04 - 1	C-19-18-19	50316	10.1536	-7471B	PPB	18 Mar 2022 11:49:00	U	SMPL
N1962-07 - 1	C-20-19-21	34097	6.8626	-7471B	PPB	18 Mar 2022 11:51:22	U	SMPL
N1962-08 - 1	C-21-21-23	25386	5.0950	-7471B	PPB	18 Mar 2022 11:53:45	U	SMPL
N1962-09 - 1	C-22-23-25	66287	13.3944	-7471B	PPB	18 Mar 2022 11:56:08	U	SMPL
N1962-10 - 1	C-23-25-28	86852	17.5673	-7471B	PPB	18 Mar 2022 11:58:27	U	SMPL
N1979-01 - 1	RT4743	1142	0.1755	-7471B	PPB	18 Mar 2022 12:00:54	U	SMPL
N1981-01 - 1	OR-3-031622	2323	0.4152	-7471B	PPB	18 Mar 2022 12:03:22	U	SMPL
N1982-01 - 1	CONCRETE-A	2815	0.5150	-7471B	PPB	18 Mar 2022 12:05:38	U	SMPL
CCV78 - 1	CCV78	27179	5.4588	-7471B	PPB	18 Mar 2022 12:07:53	U	SMPL
CCB78 - 1	CCB78	-1	-0.0564	-7471B	PPB	18 Mar 2022 12:10:10	U	SMPL
N1982-03 - 1	CONCRETE-B	732	0.0923	-7471B	PPB	18 Mar 2022 12:12:29	U	SMPL
N1983-01 - 1	124035	2469	0.4448	-7471B	PPB	18 Mar 2022 12:14:45	U	SMPL
N1983-03 - 1	124042	1482	0.2445	-7471B	PPB	18 Mar 2022 12:17:00	U	SMPL
N1983-05 - 1	124019	2364	0.4235	-7471B	PPB	18 Mar 2022 12:19:16	U	SMPL
N1983-07 - 1	124055	1062	0.1593	-7471B	PPB	18 Mar 2022 12:21:32	U	SMPL
N1993-01 - 1	TR-04-03162022	569	0.0593	-7471B	PPB	18 Mar 2022 12:23:49	U	SMPL
N1993-01DUP - 1	TR-04-03162022DUP	453	0.0357	-7471B	PPB	18 Mar 2022 12:26:05	U	SMPL
N1993-01MS - 1	TR-04-03162022MS	19984	3.9988	-7471B	PPB	18 Mar 2022 12:28:23	U	SMPL
N1993-01MSD - 1	TR-04-03162022MSD	20563	4.1163	-7471B	PPB	18 Mar 2022 12:30:40	U	SMPL
PB143435BL - 1	PBS	179	-0.0199	-7471B	PPB	18 Mar 2022 12:32:59	U	SMPL
CCV79 - 1	CCV79	26519	5.3249	-7471B	PPB	18 Mar 2022 12:35:18	U	SMPL
CCB79 - 1	CCB79	-1	-0.0564	-7471B	PPB	18 Mar 2022 12:37:34	U	SMPL
PB143435BS - 1	LCSS	19435	3.8874	-7471B	PPB	18 Mar 2022 12:39:53	U	SMPL
N1960-15 - 1	SB-13-20220316-16.0-17.0	29382	5.9058	-7471B	PPB	18 Mar 2022 12:42:09	U	SMPL
N1960-16 - 1	SB-13-20220316-18.0-19.0	18628	3.7237	-7471B	PPB	18 Mar 2022 12:44:28	U	SMPL
N1994-02 - 1	SB-01-COMP	1108	0.1686	-7471B	PPB	18 Mar 2022 12:46:48	U	SMPL
N1994-05 - 1	SB-02-COMP	454	0.0359	-7471B	PPB	18 Mar 2022 12:49:06	U	SMPL
N1994-08 - 1	SB-03-COMP	821	0.1104	-7471B	PPB	18 Mar 2022 12:51:23	U	SMPL
N1994-11 - 1	SB-04-COMP	11709	2.3197	-7471B	PPB	18 Mar 2022 12:53:39	U	SMPL
N1994-14 - 1	SB-05-COMP	5479	1.0556	-7471B	PPB	18 Mar 2022 12:55:56	U	SMPL
N1995-02 - 1	SB-06-COMP	5062	0.9710	-7471B	PPB	18 Mar 2022 12:58:16	U	SMPL
N1995-05 - 1	SB-07-COMP	47778	9.6386	-7471B	PPB	18 Mar 2022 13:00:34	U	SMPL
CCV80 - 1	CCV80	25442	5.1063	-7471B	PPB	18 Mar 2022 13:05:33	U	SMPL
CCB80 - 1	CCB80	20	-0.0521	-7471B	PPB	18 Mar 2022 13:07:50	U	SMPL
N1995-08 - 1	SB-06-COMP	33063	6.6528	-7471B	PPB	18 Mar 2022 13:10:10	U	SMPL
N2001-01 - 1	D-14A	737	0.0934	-7471B	PPB	18 Mar 2022 13:12:26	U	SMPL
N2001-01MS - 1	D-14AMS	20676	4.1393	-7471B	PPB	18 Mar 2022 13:17:02	U	SMPL
N2001-01MSD - 1	D-14AMSD	21122	4.2298	-7471B	PPB	18 Mar 2022 13:19:19	U	SMPL
N2007-01 - 1	SP1	5143	0.9874	-7471B	PPB	18 Mar 2022 13:21:38	U	SMPL
N2007-02 - 1	SP2	2373	0.4253	-7471B	PPB	18 Mar 2022 13:23:58	U	SMPL
N2007-03 - 1	SP3	20302	4.0634	-7471B	PPB	18 Mar 2022 13:26:17	U	SMPL
N2001-01DUP - 1	D-14ADUP	804	0.1070	-7471B	PPB	18 Mar 2022 13:28:54	U	SMPL
N1977-01LX5 - 1		2513	0.4537	-7471B	PPB	18 Mar 2022 13:31:33	U	SMPL
N1977-01A - 1		31292	6.2934	-7471B	PPB	18 Mar 2022 13:36:41	U	SMPL
CCV81 - 1	CCV81	24627	4.9410	-7471B	PPB	18 Mar 2022 13:38:57	U	SMPL

LB119216 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 18 Mar 2022 09:20:24

Sample ID	Extended ID	μ Abs.	Conc.	Std Cond	Method	Units	Date	Type	Type
CCB81 - 1	CCB81	40	-0.0481		-7471B	PPB	18 Mar 2022 13:41:17	U	SMPL
N1993-01LX5 - 1		82	-0.0395		-7471B	PPB	18 Mar 2022 13:43:37	U	SMPL
N1993-01A - 1		20252	4.0532		-7471B	PPB	18 Mar 2022 13:45:53	U	SMPL
N2001-01LX5 - 1		-1	-0.0564		-7471B	PPB	18 Mar 2022 13:48:09	U	SMPL
N2001-01A - 1		20978	4.2005		-7471B	PPB	18 Mar 2022 13:50:28	U	SMPL
N1958-01DLX10 - 1		81247	16.4300		-7471B	PPB	18 Mar 2022 13:52:45	U	SMPL
N1958-02DLX5 - 1		19074	3.8142		-7471B	PPB	18 Mar 2022 13:55:04	U	SMPL
N1958-03DLX10 - 1		24816	4.9793		-7471B	PPB	18 Mar 2022 13:57:30	U	SMPL
N1958-04DLX5 - 1		26341	5.2888		-7471B	PPB	18 Mar 2022 13:59:50	U	SMPL
N1958-05DLX10 - 1		21998	4.4075		-7471B	PPB	18 Mar 2022 14:02:10	U	SMPL
N1958-06DLX10 - 1		43539	8.7785		-7471B	PPB	18 Mar 2022 14:04:30	U	SMPL
CCV82 - 1	CCV82	25824	5.1839		-7471B	PPB	18 Mar 2022 14:11:14	U	SMPL
CCB82 - 1	CCB82	34	-0.0493		-7471B	PPB	18 Mar 2022 14:13:30	U	SMPL
N1958-07DLX10 - 1		18308	3.6588		-7471B	PPB	18 Mar 2022 14:15:50	U	SMPL
N1958-08DLX5 - 1		28249	5.6759		-7471B	PPB	18 Mar 2022 14:18:07	U	SMPL
N1958-09DLX5 - 1		32192	6.4760		-7471B	PPB	18 Mar 2022 14:20:26	U	SMPL
N1958-10DLX10 - 1		28151	5.6560		-7471B	PPB	18 Mar 2022 14:22:46	U	SMPL
N1977-06DLX10 - 1		86344	17.4642		-7471B	PPB	18 Mar 2022 14:25:05	U	SMPL
N1962-04DLX2 - 1		26984	5.4192		-7471B	PPB	18 Mar 2022 14:27:25	U	SMPL
N1962-09DLX5 - 1		14840	2.9551		-7471B	PPB	18 Mar 2022 14:29:53	U	SMPL
N1962-10DLX5 - 1		18513	3.7004		-7471B	PPB	18 Mar 2022 14:32:13	U	SMPL
N1958-01DL2X50 - 1		16845	3.3619		-7471B	PPB	18 Mar 2022 14:34:32	U	SMPL
N1977-06DL2X50 - 1		18145	3.6257		-7471B	PPB	18 Mar 2022 14:36:52	U	SMPL
CCV83 - 1	CCV83	24867	4.9897		-7471B	PPB	18 Mar 2022 14:39:11	U	SMPL
CCB83 - 1	CCB83	71	-0.0418		-7471B	PPB	18 Mar 2022 14:41:30	U	SMPL