



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : mohammad

Run Number: LB119221

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
ZOBELLS SOLUTION ORP STD 500ML	W2859
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	03/18/2022	12:00
2	CAL2	1	Water	NA	NA	20.2	7.00	03/18/2022	12:01
3	CAL3	1	Water	NA	NA	20.2	10.03	03/18/2022	12:02
4	ICV	1	Water	NA	NA	20.3	6.99	03/18/2022	12:05
5	CCV1	1	Water	NA	NA	20.4	2.01	03/18/2022	12:06
6	N2014-01	1	Water	NA	NA	20.1	9.12	03/18/2022	12:10
7	N2014-01DUP	1	Water	NA	NA	20.2	9.13	03/18/2022	12:11
8	CCV2	1	Water	NA	NA	20.4	12.03	03/18/2022	12:15

VB 11922)

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N1936 W WorkList ID : 157860 Department : Wet-Chemistry Date : 03-18-2022 11:45:33

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/17/2022	Water	N2014-01	pH	Cool 4 deg C	ENV127	N22	DEWATERING-EFFLUENT-031722	03/17/2022	9040C

Date/Time 03/18/22 11:50
Raw Sample Received by: J.P. (SM)
Raw Sample Relinquished by: J.P. (SM)

Date/Time 03/18/22 13:00
Raw Sample Received by: J.P. (SM)
Raw Sample Relinquished by: J.P. (SM)