



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB119225

Slope : 98.1

BalanceID: WC SC-6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
ZOBELLS SOLUTION ORP STD 500ML	W2859
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	03/18/2022	12:35
2	CAL2	1	Water	NA	NA	20.3	7.00	03/18/2022	12:36
3	CAL3	1	Water	NA	NA	20.3	10.03	03/18/2022	12:37
4	ICV	1	Water	NA	NA	20.4	6.98	03/18/2022	12:40
5	CCV1	1	Water	NA	NA	20.4	2.01	03/18/2022	12:44
6	N2010-01	1	Solid	20.04	20	20.4	8.93	03/18/2022	14:15
7	N2010-01DUP	1	Solid	20.03	20	20.3	8.94	03/18/2022	14:16
8	CCV2	1	Water	NA	NA	20.4	12.03	03/18/2022	14:20

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH S N2010 WorkList ID : 157864 Department : Wet-Chemistry Date : 03-18-2022 12:27:23

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/16/2022	Solid	N2010-01	pH	Cool 4 deg C	AMOE01	N11	WC1-220316	03/16/2022	9045D

Date/Time 03-18-22 12:30
 Raw Sample Received by: JN(CDC)
 Raw Sample Relinquished by: JN(CDC)

Date/Time 03-18-22 14:30
 Raw Sample Received by: JN(CDC)
 Raw Sample Relinquished by: JN(CDC)

VB 169225