



PERCENT SOLID

Supervisor: mohammad

Analyst: JIGNESH

Date: 3/21/2022

OVENTEMP IN Celsius(°C): 103

Time IN: 14:00

In Date: 03/19/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 08:00

Out Date: 03/20/2022

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB119243

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
N2034-01	BGZJ3	1	1.14	8.75	9.89	5.88	54.2	
N2034-02	BGZJ5	2	1.15	8.83	9.98	8.05	78.1	
N2034-03	BGZK6	3	1.15	8.79	9.94	3.83	30.5	
N2034-04	BGZK7	4	1.12	8.84	9.96	5.26	46.8	
N2034-05	BGZL2	5	1.17	8.59	9.76	7.48	73.5	
N2034-06	BGZM0	6	1.15	8.82	9.97	6.22	57.5	
N2034-07	BGZM1	7	1.18	8.48	9.66	7.43	73.7	
N2034-08	BGZM2	8	1.19	8.79	9.98	5.94	54.0	
N2034-09	BGZM3	9	1.18	8.67	9.85	5.53	50.2	
N2034-10	BGZN6	10	1.16	8.72	9.88	4.98	43.8	
N2034-11	BGZS0	11	1.18	8.66	9.84	5.96	55.2	
N2034-12	BGZS2	12	1.18	8.73	9.91	5.32	47.4	
N2034-13	BGZS4	13	1.16	8.80	9.96	9.12	90.5	
N2034-14	BGZS5	14	1.14	8.84	9.98	5.32	47.3	
N2034-15	BGZS6	15	1.14	8.38	9.52	6.47	63.6	
N2034-16	BGZT2	16	1.12	8.46	9.58	7.29	72.9	
N2034-17	BGZT3	17	1.19	8.49	9.68	7.42	73.4	
N2034-18	BGZW2	18	1.18	8.79	9.97	5.77	52.2	
N2034-19	BGZW3	19	1.1	8.88	9.98	5.27	47.0	
N2034-20	BH047	20	1.17	8.80	9.97	7.85	75.9	
N2034-21	BH047D	21	1.17	8.80	9.97	7.85	75.9	
N2034-22	BH047S	22	1.17	8.80	9.97	7.85	75.9	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$