



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: apatel

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB119267

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	6	27.00	80.6	80.3	03/21/2022	12:40
2	N1950-01		1	11	100.00	>212.0	>212.0	03/21/2022	12:55
3	N1950-02		1	11	100.00	>212.0	>212.0	03/21/2022	13:15
4	N1950-03		1	10	100.00	>212.0	>212.0	03/21/2022	13:35
5	N1950-04		1	10	100.00	>212.0	>212.0	03/21/2022	13:57
6	N1950-05		1	9	100.00	>212.0	>212.0	03/21/2022	14:16
7	N1950-06		1	10	100.00	>212.0	>212.0	03/21/2022	14:38
8	N1950-07		1	9	100.00	>212.0	>212.0	03/21/2022	14:55
9	N1950-08		1	10	100.00	>212.0	>212.0	03/21/2022	15:16
10	N1950-10		1	11	100.00	>212.0	>212.0	03/21/2022	15:35
11	N1992-01		1	10	100.00	>212.0	>212.0	03/21/2022	15:54
12	N1992-01DUP		1	9	100.00	>212.0	>212.0	03/21/2022	16:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FB032122 WorkList ID : 157924 Department : Wet-Chemistry Date : 03-21-2022 12:20:25

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample		Collect Date	Method
						Storage	Location		
03/21/2022	Water	N1950-01	Flash Point	Cool 4 deg C	PSEG03	E11	31339	03/14/2022	1010B
03/21/2022	Water	N1950-02	Flash Point	Cool 4 deg C	PSEG03	E11	31340	03/14/2022	1010B
03/21/2022	Water	N1950-03	Flash Point	Cool 4 deg C	PSEG03	E11	31406	03/14/2022	1010B
03/21/2022	Water	N1950-04	Flash Point	Cool 4 deg C	PSEG03	E11	31407	03/14/2022	1010B
03/21/2022	Water	N1950-05	Flash Point	Cool 4 deg C	PSEG03	E11	31408	03/14/2022	1010B
03/21/2022	Water	N1950-06	Flash Point	Cool 4 deg C	PSEG03	E11	31432	03/14/2022	1010B
03/21/2022	Water	N1950-07	Flash Point	Cool 4 deg C	PSEG03	E11	31433	03/14/2022	1010B
03/21/2022	Water	N1950-08	Flash Point	Cool 4 deg C	PSEG03	E11	31424	03/14/2022	1010B
03/21/2022	Water	N1950-10	Flash Point	Cool 4 deg C	PSEG03	E11	3222	03/14/2022	1010B
03/21/2022	Water	N1992-01	Flash Point	Cool 4 deg C	PSEG03	M11	N-37853	03/16/2022	1010B

Date/Time 3/21/22 12:25
 Raw Sample Received by: AL (W)
 Raw Sample Relinquished by: AL (W)

Date/Time 3/24/22 12:12
 Raw Sample Received by: AL (W)
 Raw Sample Relinquished by: AL (W)