



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB119270

Slope : 98.5

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
ZOBELLS SOLUTION ORP STD 500ML	W2859
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	03/21/2022	15:30
2	CAL2	1	Water	NA	NA	20.4	7.00	03/21/2022	16:15
3	CAL3	1	Water	NA	NA	20.5	10.03	03/21/2022	16:16
4	ICV	1	Water	NA	NA	20.4	6.99	03/21/2022	16:19
5	CCV1	1	Water	NA	NA	20.5	2.01	03/21/2022	16:20
6	N2025-01	1	Solid	20.01	20	20.1	5.62	03/21/2022	16:22
7	N2058-02	1	Solid	20.03	20	21.3	9.41	03/21/2022	16:24
8	N2058-02DUP	1	Solid	20.04	20	21.4	9.42	03/21/2022	16:25
9	CCV2	1	Water	NA	NA	20.3	12.03	03/21/2022	16:26

LB 119270

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY N2058 WorkList ID : 157933 Department : Wet-Chemistry Date : 03-21-2022 15:20:41

Due Date	Matrix	Sample	Test	Raw Sample			Collect Date	Method
				Preservative	Customer	Storage Location		
03/18/2022	Solid	N2025-01	Corrosivity	Cool 4 deg C	PSEG03	E11 D-154	03/18/2022	9045D
03/21/2022	Solid	N2058-02	Corrosivity	Cool 4 deg C	PSEG03	E12 VR-TP1	03/21/2022	9045D

Date/Time 03/21/22 15:25
Raw Sample Received by: JD (WCI)
Raw Sample Relinquished by: JD (WCI)

Date/Time 03/21/22 17:00
Raw Sample Received by: JD (WCI)
Raw Sample Relinquished by: JD (WCI)