


LB119277

7470A

INSTRUMENT ID: CV1

 μ Abs.:

3804E



Linear

A= 0.0000e+000

B= 2.6157e-004

C= 4.9648e-002

Rho= 0.9999017

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.060	0.060	40	0.000	40					1
0.2	0.200	0.217	0.017	638	0.0 %	638					8
2.5	2.500	2.405	-0.095	9005	0.0 %	9005					-4
5.0	5.000	4.969	-0.031	18809	0.0 %	18809					-1
7.5	7.500	7.547	0.047	28664	0.0 %	28664					1
10.0	10.000	10.002	0.002	38048	0.0 %	38048					0

LB119277 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 22 Mar 2022 11:12:08

Sample ID	Extended ID	µ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	SP	40	-	0.00007470A	PPB	22 Mar 2022 11:22:10	S	Std
0.2 - 1	SP2	638	-	0.20007470A	PPB	22 Mar 2022 11:24:28	S	Std
2.5 - 1	SP2.5	9005	-	2.50007470A	PPB	22 Mar 2022 11:26:45	S	Std
5.0 - 1	SP5	18809	-	5.00007470A	PPB	22 Mar 2022 11:29:03	S	Std
7.5 - 1	SP7.5	28664	-	7.50007470A	PPB	22 Mar 2022 11:34:00	S	Std
10.0 - 1	SP10	38048	-	10.00007470A	PPB	22 Mar 2022 11:41:27	S	Std
ICV39 - 1	ICV39	13759	3.6485	-7470A	PPB	22 Mar 2022 11:44:06	U	SMPL
ICB39 - 1	ICB39	-58	0.0345	-7470A	PPB	22 Mar 2022 11:46:21	U	SMPL
CCV91 - 1	CCV91	20321	5.3649	-7470A	PPB	22 Mar 2022 11:48:37	U	SMPL
CCB91 - 1	CCB91	-51	0.0363	-7470A	PPB	22 Mar 2022 11:50:53	U	SMPL
CRA - 1	CRA	712	0.2359	-7470A	PPB	22 Mar 2022 11:53:10	U	SMPL
HighStd - 1	HighStd	39150	10.2899	-7470A	PPB	22 Mar 2022 11:57:36	U	SMPL
ChkStd - 1	ChkStd	27258	7.1794	-7470A	PPB	22 Mar 2022 11:59:52	U	SMPL
PB143490BL - 1	PBW	-86	0.0272	-7470A	PPB	22 Mar 2022 12:02:10	U	SMPL
PB143490BS - 1	LCSW	16148	4.2734	-7470A	PPB	22 Mar 2022 12:04:28	U	SMPL
N1958-16 - 1	C-1-0-1	-8	0.0476	-7470A	PPB	22 Mar 2022 12:06:44	U	SMPL
N1958-18 - 1	C-3-2-3	97	0.0750	-7470A	PPB	22 Mar 2022 12:09:02	U	SMPL
N1958-19 - 1	C-4-3-4	32	0.0580	-7470A	PPB	22 Mar 2022 12:11:20	U	SMPL
N1958-20 - 1	C-5-4-5	36	0.0591	-7470A	PPB	22 Mar 2022 12:13:36	U	SMPL
N1958-22 - 1	C-7-6-7	385	0.1504	-7470A	PPB	22 Mar 2022 12:15:53	U	SMPL
CCV92 - 1	CCV92	20631	5.4460	-7470A	PPB	22 Mar 2022 12:18:08	U	SMPL
CCB92 - 1	CCB92	-40	0.0392	-7470A	PPB	22 Mar 2022 12:20:24	U	SMPL
N1958-23 - 1	C-8-7-8	14	0.0533	-7470A	PPB	22 Mar 2022 12:22:40	U	SMPL
N1958-24 - 1	C-9-8-9	31	0.0578	-7470A	PPB	22 Mar 2022 12:24:55	U	SMPL
N1958-25 - 1	C-10-9-10	352	0.1417	-7470A	PPB	22 Mar 2022 12:27:11	U	SMPL
N1962-05 - 1	C-11-10-12	27	0.0567	-7470A	PPB	22 Mar 2022 12:29:28	U	SMPL
N1962-06 - 1	C-12-12-14	44	0.0612	-7470A	PPB	22 Mar 2022 12:31:45	U	SMPL
N1962-11 - 1	C-17-19-22	66	0.0669	-7470A	PPB	22 Mar 2022 12:34:02	U	SMPL
N1962-12 - 1	C-18-22-25	42	0.0606	-7470A	PPB	22 Mar 2022 12:36:19	U	SMPL
N1962-13 - 1	C-16-15-16	50	0.0627	-7470A	PPB	22 Mar 2022 12:38:36	U	SMPL
N1962-14 - 1	C-17-16-17	68	0.0674	-7470A	PPB	22 Mar 2022 12:40:53	U	SMPL
N1962-15 - 1	C-18-17-18	66	0.0669	-7470A	PPB	22 Mar 2022 12:43:09	U	SMPL
CCV93 - 1	CCV93	20755	5.4784	-7470A	PPB	22 Mar 2022 12:45:25	U	SMPL
CCB93 - 1	CCB93	-39	0.0394	-7470A	PPB	22 Mar 2022 12:47:41	U	SMPL
N1962-16 - 1	C-19-18-19	30	0.0575	-7470A	PPB	22 Mar 2022 12:49:58	U	SMPL
N1962-20 - 1	C-23-25-28	23	0.0557	-7470A	PPB	22 Mar 2022 12:52:14	U	SMPL
N1985-02 - 1	VACLOT-SOIL-1	48	0.0622	-7470A	PPB	22 Mar 2022 12:54:30	U	SMPL
N1985-05 - 1	VACLOT-SOIL-2	21	0.0551	-7470A	PPB	22 Mar 2022 12:56:45	U	SMPL
N2010-02 - 1	WC1-220316	23	0.0557	-7470A	PPB	22 Mar 2022 12:59:02	U	SMPL
N2010-02DUP - 1	WC1-220316DUP	31	0.0578	-7470A	PPB	22 Mar 2022 13:01:19	U	SMPL
N2010-02MS - 1	WC1-220316MS	16226	4.2938	-7470A	PPB	22 Mar 2022 13:03:35	U	SMPL
N2010-02MSD - 1	WC1-220316MSD	15460	4.0935	-7470A	PPB	22 Mar 2022 13:05:52	U	SMPL
PB143491BL - 1	PBW	-13	0.0462	-7470A	PPB	22 Mar 2022 13:08:11	U	SMPL
PB143491BS - 1	LCSW	15369	4.0697	-7470A	PPB	22 Mar 2022 13:10:27	U	SMPL
CCV94 - 1	CCV94	20555	5.4261	-7470A	PPB	22 Mar 2022 13:12:43	U	SMPL
CCB94 - 1	CCB94	-67	0.0321	-7470A	PPB	22 Mar 2022 13:15:00	U	SMPL
N2021-05 - 1	2281	92	0.0737	-7470A	PPB	22 Mar 2022 13:17:17	U	SMPL
N2021-05DUP - 1	2281DUP	82	0.0711	-7470A	PPB	22 Mar 2022 13:19:34	U	SMPL
N2021-05MS - 1	2281MS	16675	4.4113	-7470A	PPB	22 Mar 2022 13:21:51	U	SMPL
N2021-05MSD - 1	2281MSD	16095	4.2595	-7470A	PPB	22 Mar 2022 13:24:08	U	SMPL
N2033-05 - 1	D-2377	-58	0.0345	-7470A	PPB	22 Mar 2022 13:26:25	U	SMPL
PB143470TB - 1	PB143470TB	-38	0.0397	-7470A	PPB	22 Mar 2022 13:28:42	U	SMPL
PB143475TB - 1	PB143475TB	-2	0.0491	-7470A	PPB	22 Mar 2022 13:30:59	U	SMPL
CCV95 - 1	CCV95	19879	5.2493	-7470A	PPB	22 Mar 2022 13:33:16	U	SMPL
CCB95 - 1	CCB95	-47	0.0374	-7470A	PPB	22 Mar 2022 13:35:33	U	SMPL
N2010-02LX5 - 1		17	0.0541	-7470A	PPB	22 Mar 2022 13:37:50	U	SMPL
N2010-02A - 1		15272	4.0443	-7470A	PPB	22 Mar 2022 13:40:07	U	SMPL
N2021-05LX5 - 1		-71	0.0311	-7470A	PPB	22 Mar 2022 13:42:24	U	SMPL
N2021-05A - 1		16334	4.3221	-7470A	PPB	22 Mar 2022 13:44:41	U	SMPL
CCV96 - 1	CCV96	19593	5.1745	-7470A	PPB	22 Mar 2022 13:46:58	U	SMPL
CCB96 - 1	CCB96	-53	0.0358	-7470A	PPB	22 Mar 2022 13:49:15	U	SMPL