



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: apatel

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB119292

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	10	27.00	80.6	80.6	03/22/2022	15:10
2	N2021-01		1	6	23.00	73.4	73.4	03/22/2022	15:20
3	N2033-01		1	9	50.00	122.0	122.0	03/22/2022	15:35
4	N2033-01DUP		1	10	51.00	123.8	123.8	03/22/2022	15:52

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FB-032122-2 WorkList ID : 157927 Department : Wet-Chemistry Date : 03-21-2022 14:08:39

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/23/2022	Solid	N2021-01	Flash Point	Cool 4 deg C	PSEG03	E11	2279	03/18/2022	1010B
03/23/2022	Solid	N2033-01	Flash Point	Cool 4 deg C	PSEG03	E11	D-2378	03/18/2022	1010B

Date/Time 3/22/22 8:30
 Raw Sample Received by: AD (WOL)
 Raw Sample Relinquished by: AD (WOL)

Date/Time 3/22/22 12:00
 Raw Sample Received by: AD (WOL)
 Raw Sample Relinquished by: AD (WOL)