



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 3/24/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 17:30
In Date: 03/23/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:33
Out Date: 03/24/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB119300

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
N2063-01	WC-1-(10-10.5)	1	1.18	8.57	9.75	8.65	87.2	
N2063-02	WC-2-(10-10.5)	2	1.18	8.62	9.8	9.44	95.8	
N2063-03	WC-3-(10-10.5)	3	1.18	8.43	9.61	9.24	95.6	
N2063-04	WC-4-1-(10-10.5)	4	1.14	8.80	9.94	8.88	88.0	
N2063-05	WC-4-2-(10-10.5)	5	1.18	8.46	9.64	8.45	85.9	
N2063-06	WC-5-(10-10.5)	6	1.18	8.38	9.56	8.72	90.0	
N2063-07	WC-7-1-(10-10.5)	7	1.18	8.78	9.96	8.85	87.4	
N2063-08	WC-7-2-(10-10.5)	8	1.16	8.49	9.65	9.3	95.9	
N2063-09	WC-8-(10-10.5)	9	1.15	8.80	9.95	9.22	91.7	
N2063-10	WC-9-(10-10.5)	10	1.17	8.75	9.92	8.51	83.9	
N2063-10D	WC-9-(10-10.5) DUP	11	1.17	8.77	9.94	8.54	84.0	
N2077-01	WC-14A-4-(2.5-3.0)	12	1.13	8.76	9.89	8.63	85.6	
N2077-03	WC-14A-4	13	1.15	8.80	9.95	8.94	88.5	
N2077-04	WC-14A-4	14	1.15	8.80	9.95	8.94	88.5	
N2077-05	WC-14A-3-1-(3.0-3.5)	15	1.18	8.56	9.74	8.6	86.7	
N2077-07	WC-14A-3-1	16	1.18	8.79	9.97	8.28	80.8	
N2077-08	WC-14A-3-1	17	1.18	8.79	9.97	8.28	80.8	
N2077-09	WC-14A-3-2-(3.0-3.5)	18	1.12	8.57	9.69	8.38	84.7	
N2077-11	WC-14A-3-2	19	1.15	8.50	9.65	7.81	78.4	
N2077-12	WC-14A-3-2	20	1.15	8.50	9.65	7.81	78.4	
N2077-13	WC-14A-3-3-(2.5-3.0)	21	1.17	8.79	9.96	8.43	82.6	
N2077-15	WC-14A-3-3	22	1.17	8.77	9.94	8.26	80.8	
N2077-16	WC-14A-3-3	23	1.17	8.77	9.94	8.26	80.8	
N2084-01	BLDG-3	24	1.14	8.80	9.94	9.32	93.0	
N2089-01	CL-01-032322	25	1.17	8.43	9.6	9.03	93.2	
N2089-02	CL-01-032322-EPH-2	26	1.18	8.58	9.76	9.15	92.9	
N2091-01	MHB	27	1.13	8.85	9.98	8.84	87.1	
N2091-03	MHB-EPH-2	28	1.14	8.82	9.96	8.95	88.5	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$