



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB119307

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
ZOBELLS SOLUTION ORP STD 500ML	W2859
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	03/23/2022	14:00
2	CAL2	1	Water	NA	NA	20.3	7.00	03/23/2022	14:01
3	CAL3	1	Water	NA	NA	20.4	10.03	03/23/2022	14:02
4	ICV	1	Water	NA	NA	20.4	6.98	03/23/2022	14:05
5	CCV1	1	Water	NA	NA	20.3	2.01	03/23/2022	14:06
6	N2076-01	1	Water	NA	NA	20.2	6.98	03/23/2022	14:10
7	N2076-01DUP	1	Water	NA	NA	20.3	6.99	03/23/2022	14:11
8	CCV2	1	Water	NA	NA	20.4	12.03	03/23/2022	14:15

LB 119307

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY N1746 W WorkList ID : 157988 Department : Wet-Chemistry Date : 03-23-2022 13:43:07

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/22/2022	Water	N2076-01	Corrosivity - F	Cool 4 deg C	PARS02	E12	WC-20220322	03/22/2022	9040C

Date/Time 03/23/22 13:50
Raw Sample Received by: TD (WC)
Raw Sample Relinquished by: CF SM

Date/Time 03/23/22 15:00
Raw Sample Received by: CP SM
Raw Sample Relinquished by: TD (WC)