

LB11930P


7470A

INSTRUMENT ID: CVL

Linear

μ Abs.:

41465



A= 0.0000e+000

B= 2.4146e-004

C= -3.3129e-002

Rho= 0.9998508

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.022	-0.022	47	0.000	47				
0.2	0.200	0.227	0.027	1079	0.0 %	1079				
2.5	2.500	2.424	-0.076	10177	0.0 %	10177				
5.0	5.000	5.126	0.126	21365	0.0 %	21365				
7.5	7.500	7.466	-0.034	31056	0.0 %	31056				
10.0	10.000	9.979	-0.021	41465	0.0 %	41465				

1/01
14
-3
3
0
0

LB119308 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 23 Mar 2022 14:08:31				
Sample ID	Extended ID	µ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	47	-	0.00007470A	PPB	23 Mar 2022 14:12:04	S	Std
0.2 - 1	51.2	1079	-	0.20007470A	PPB	23 Mar 2022 14:14:22	S	Std
2.5 - 1	52.5	10177	-	2.50007470A	PPB	23 Mar 2022 14:16:39	S	Std
5.0 - 1	55	21365	-	5.00007470A	PPB	23 Mar 2022 14:18:56	S	Std
7.5 - 1	57.5	31056	-	7.50007470A	PPB	23 Mar 2022 14:26:10	S	Std
10.0 - 1	511	41465	-	10.00007470A	PPB	23 Mar 2022 14:35:31	S	Std
ICV42 - 1	ICV42	15224	3.6428	-7470A	PPB	23 Mar 2022 14:38:14	U	SMPL
ICB42 - 1	ICB42	-54	-0.0462	-7470A	PPB	23 Mar 2022 14:40:30	U	SMPL
CCV05 - 1	CCV05	20416	4.8965	-7470A	PPB	23 Mar 2022 14:42:45	U	SMPL
CCB05 - 1	CCB05	-65	-0.0488	-7470A	PPB	23 Mar 2022 14:45:01	U	SMPL
CRA - 1	CRA	1007	0.2100	-7470A	PPB	23 Mar 2022 14:47:17	U	SMPL
HighStd - 1	HighStd	39651	9.5409	-7470A	PPB	23 Mar 2022 14:49:33	U	SMPL
ChkStd - 1	ChkStd	29304	7.0426	-7470A	PPB	23 Mar 2022 14:51:48	U	SMPL
PB143545BL - 1	PBW	299	0.0391	-7470A	PPB	23 Mar 2022 14:54:06	U	SMPL
PB143545BS - 1	LCSW	15609	3.7358	-7470A	PPB	23 Mar 2022 14:58:42	U	SMPL
N2031-08 - 1	FIELD-BLANK	-85	-0.0537	-7470A	PPB	23 Mar 2022 15:14:29	U	SMPL
N2053-03 - 1	AUD-1267	2739	0.6282	-7470A	PPB	23 Mar 2022 15:21:42	U	SMPL
N2053-03DUP - 1	AUD-1267DUP	2710	0.6212	-7470A	PPB	23 Mar 2022 15:23:59	U	SMPL
N2053-03MS - 1	AUD-1267MS	19376	4.6454	-7470A	PPB	23 Mar 2022 15:28:44	U	SMPL
CCV06 - 1	CCV06	19433	4.6591	-7470A	PPB	23 Mar 2022 15:30:59	U	SMPL
CCB06 - 1	CCB06	79	-0.0141	-7470A	PPB	23 Mar 2022 15:36:06	U	SMPL
N2053-03MSD - 1	AUD-1267MSD	18498	4.4334	-7470A	PPB	23 Mar 2022 15:38:22	U	SMPL
N2053-04 - 1	AUD-1271	258	0.0292	-7470A	PPB	23 Mar 2022 15:40:37	U	SMPL
N2061-03 - 1	EB-20220321	363	0.0545	-7470A	PPB	23 Mar 2022 15:46:23	U	SMPL
N2064-01 - 1	20220280-SITE-WATER	391	0.0613	-7470A	PPB	23 Mar 2022 15:48:38	U	SMPL
N2064-02 - 1	20220281-FIELD-BLANK	216	0.0190	-7470A	PPB	23 Mar 2022 15:50:54	U	SMPL
CCV07 - 1	CCV07	20775	4.9832	-7470A	PPB	23 Mar 2022 16:12:15	U	SMPL
CCB07 - 1	CCB07	77	-0.0145	-7470A	PPB	23 Mar 2022 16:17:22	U	SMPL
N1872-07 - 1	PT-HG-WP	27247	6.5459	-7470A	PPB	23 Mar 2022 16:29:04	U	SMPL
N2053-03LX5 - 1		643	0.1221	-7470A	PPB	23 Mar 2022 16:35:44	U	SMPL
N2053-03A - 1		19755	4.7369	-7470A	PPB	23 Mar 2022 16:38:01	U	SMPL
CCV08 - 1	CCV08	20073	4.8137	-7470A	PPB	23 Mar 2022 16:44:53	U	SMPL
CCB08 - 1	CCB08	56	-0.0196	-7470A	PPB	23 Mar 2022 16:47:09	U	SMPL