



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB119316

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	03/23/2022	17:20
2	CAL2	1	Water	NA	NA	20.7	7.00	03/23/2022	17:21
3	CAL3	1	Water	NA	NA	20.7	10.03	03/23/2022	17:22
4	ICV	1	Water	NA	NA	20.6	7.00	03/23/2022	17:25
5	CCV1	1	Water	NA	NA	20.6	2.01	03/23/2022	17:30
6	N2060-01	1	Solid	20.03	20	20.1	8.91	03/23/2022	17:32
7	N2077-04	1	Solid	20.04	20	20.2	8.64	03/23/2022	17:35
8	N2077-08	1	Solid	20.02	20	20.2	8.50	03/23/2022	17:36
9	N2077-12	1	Solid	20.03	20	20.5	8.76	03/23/2022	17:38
10	N2077-16	1	Solid	20.04	20	20.8	8.33	03/23/2022	17:40
11	N2084-02	1	Solid	20.02	20	20.4	9.10	03/23/2022	17:43
12	N2091-02	1	Solid	20.03	20	20.1	8.15	03/23/2022	17:45
13	N2091-02DUP	1	Solid	20.04	20	20.2	8.16	03/23/2022	17:46
14	CCV2	1	Water	NA	NA	20.3	12.03	03/23/2022	17:47

VP 119316

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY N2077 WorkList ID : 157991 Department : Wet-Chemistry Date : 03-23-2022 15:08:25

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/21/2022	Solid	N2060-01	Corrosivity	Cool 4 deg C	PARS02	E12	WC-20220321	03/21/2022	9045D
03/22/2022	Solid	N2077-04	Corrosivity	Cool 4 deg C	ROUX02	M51	WC-14A-4	03/22/2022	9045D
03/22/2022	Solid	N2077-08	Corrosivity	Cool 4 deg C	ROUX02	M51	WC-14A-3-1	03/22/2022	9045D
03/22/2022	Solid	N2077-12	Corrosivity	Cool 4 deg C	ROUX02	M51	WC-14A-3-2	03/22/2022	9045D
03/22/2022	Solid	N2077-16	Corrosivity	Cool 4 deg C	ROUX02	M51	WC-14A-3-3	03/22/2022	9045D
03/23/2022	Solid	N2084-02	Corrosivity	Cool 4 deg C	ENTA05	E12	BLDG-3	03/23/2022	9045D
03/23/2022	Solid	N2091-02	Corrosivity	Cool 4 deg C	PSEG03	E12	MHB	03/23/2022	9045D

Date/Time 03/23/22 17:15
 Raw Sample Received by: JD (CSC)
 Raw Sample Relinquished by: JD (CSC)

Date/Time 03/23/22 18:00
 Raw Sample Received by: JD (CSC)
 Raw Sample Relinquished by: JD (CSC)