



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB119319

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	03/24/2022	08:30
2	CAL2	1	Water	NA	NA	20.4	7.00	03/24/2022	08:31
3	CAL3	1	Water	NA	NA	20.4	10.03	03/23/2022	08:32
4	ICV	1	Water	NA	NA	20.4	6.99	03/24/2022	08:33
5	CCV1	1	Water	NA	NA	20.4	2.02	03/24/2022	08:34
6	N1873-21	1	Water	NA	NA	22.2	6.06	03/24/2022	08:35
7	N2085-02	1	Water	NA	NA	20.2	7.52	03/24/2022	08:36
8	N2085-03	1	Water	NA	NA	20.1	7.95	03/24/2022	08:39
9	N2085-04	1	Water	NA	NA	20.3	7.60	03/24/2022	08:40
10	N2085-05	1	Water	NA	NA	20.2	7.95	03/24/2022	08:45
11	N2085-05DUP	1	Water	NA	NA	20.3	7.96	03/24/2022	08:46
12	CCV2	1	Water	NA	NA	20.3	12.03	03/24/2022	08:49

114319

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph n2085 WorkList ID : 157997 Department : Wet-Chemistry Date : 03-24-2022 08:23:06

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample		Collect Date	Method
						Storage Location	Customer Sample		
03/23/2022	Water	N2085-02	pH	Cool 4 deg C	PSEG03	E12	GW-1	03/23/2022	9040C
03/23/2022	Water	N2085-03	pH	Cool 4 deg C	PSEG03	E12	GW-2	03/23/2022	9040C
03/23/2022	Water	N2085-04	pH	Cool 4 deg C	PSEG03	E12	GW-3	03/23/2022	9040C
03/23/2022	Water	N2085-05	pH	Cool 4 deg C	PSEG03	E12	GW-4	03/23/2022	9040C

Date/Time 03/24/22 08:23
Raw Sample Received by: JD (WC)
Raw Sample Relinquished by: JD (WC)

Date/Time 03/24/22 10:00
Raw Sample Received by: JD (WC)
Raw Sample Relinquished by: JD (WC)