



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB119320

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
ZOBELLS SOLUTION ORP STD 500ML	W2859
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	03/24/2022	08:30
2	CAL2	1	Water	NA	NA	20.4	7.00	03/24/2022	08:31
3	CAL3	1	Water	NA	NA	20.4	10.03	03/24/2022	08:32
4	ICV	1	Water	NA	NA	20.4	6.99	03/24/2022	08:34
5	CCV1	1	Water	NA	NA	20.4	2.02	03/24/2022	08:35
6	N1873-20	1	Water	NA	NA	22.2	6.06	03/24/2022	08:36
7	CCV2	1	Water	NA	NA	20.03	12.03	03/24/2022	08:49