



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Iwona

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB119324

Reagent/Standard	Lot/Log #
Residual chlorine LOD, 0.05PPM	WP95933
Residual chlorine ICV-LCS, 0.4PPM	WP95932
Chlorine Calibration std, 0.1ppm	WP95927
Chlorine Calibration std, 0.2ppm	WP95928
Chlorine Calibration std, 0.8ppm	WP95930
Chlorine Calibration std, 0.0ppm	WP95926
Chlorine Calibration std, 1.6ppm	WP95931
Residual Chlorine LOQ 0.1ppm	WP95934
Residual Chlorine Calibration and CCV std, 0	WP95929
Total Chlorine Powder Pillows	W2759

Intercept: 0.0077

Slope: 1.0044

Regression: 0.999955

Seq	Lab ID	True Val (mg/l)	DF	Initial Reading	Final Reading	Difference	Result (mg/l)	%D	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	-0.01		03/24/2022	11:10
2	CAL2	0.1	1	0.000	0.110	0.110	0.10	-2	03/24/2022	11:13
3	CAL3	0.2	1	0.000	0.210	0.210	0.20	-0.5	03/24/2022	11:16
4	CAL4	0.4	1	0.000	0.410	0.410	0.40	-0.3	03/24/2022	11:20
5	CAL5	0.8	1	0.000	0.820	0.820	0.81	-1.1	03/24/2022	11:23
6	CAL6	1.6	1	0.000	1.610	1.610	1.60	0.3	03/24/2022	11:26



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Iwona

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB119324

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.0000	0.4200	0.4200	0.4100	03/24/2022	11:30
2	ICB				1	0.0000	0.0000	0.0000	-0.0080	03/24/2022	11:33
3	CCV1			0.4	1	0.0000	0.4100	0.4100	0.4010	03/24/2022	11:36
4	CCB1				1	0.0000	0.0100	0.0100	0.0020	03/24/2022	11:39
5	LB119324BL	50	50		1	0.0000	0.0100	0.0100	0.0020	03/24/2022	11:42
6	LB119324BS	50	50	0.4	1	0.0000	0.4000	0.4000	0.3910	03/24/2022	11:45
7	N1645-07	50	50		1	0.0000	0.0600	0.0600	0.0520	03/24/2022	11:48
8	N1645-08	50	50		1	0.0000	0.1100	0.1100	0.1020	03/24/2022	11:51
9	N1873-19	50	50		1	0.0000	0.6500	0.6500	0.6390	03/24/2022	11:54
10	CCV2			0.4	1	0.0000	0.4100	0.4100	0.4010	03/24/2022	11:57
11	CCB2				1	0.0000	0.0000	0.0000	-0.0080	03/24/2022	12:00