


LB119325

7471B

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.1317e-004

C= -6.3776e-002

Rho= 0.9996609

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/01
0.0	0.000	-0.010	-0.010	250	0.000	250					4
0.2	0.200	0.208	0.008	1274	0.0 %	1274					1
2.5	2.500	2.529	0.029	12163	0.0 %	12163					1
5.0	5.000	5.072	0.072	24094	0.0 %	24094					3
7.5	7.500	7.302	-0.198	34553	0.0 %	34553					1
10.0	10.000	10.099	0.099	47677	0.0 %	47677					

LB119325 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 24 Mar 2022 12:42:29					
Sample ID	Extended ID	µ Abs.	Conc.	Std. Conc.	Method	Units	Date	Type	Type
0.0 - 1	30	250	-	0.0000	7471B	PPB	24 Mar 2022 13:04:26	S	Std
0.2 - 1	50.2	1274	-	0.2000	7471B	PPB	24 Mar 2022 13:06:44	S	Std
2.5 - 1	52.5	12163	-	2.5000	7471B	PPB	24 Mar 2022 13:09:01	S	Std
5.0 - 1	53	24094	-	5.0000	7471B	PPB	24 Mar 2022 13:11:19	S	Std
7.5 - 1	53.5	34553	-	7.5000	7471B	PPB	24 Mar 2022 13:13:38	S	Std
10.0 - 1	53.5	47677	-	10.0000	7471B	PPB	24 Mar 2022 13:23:50	S	Std
ICV44 - 1	ICV44	19828	4.1629	-	7471B	PPB	24 Mar 2022 13:26:33	U	SMPL
ICB44 - 1	ICB44	131	-0.0359	-	7471B	PPB	24 Mar 2022 13:28:49	U	SMPL
CCV13 - 1	CCV13	24391	5.1356	-	7471B	PPB	24 Mar 2022 13:31:07	U	SMPL
CCB13 - 1	CCB13	124	-0.0373	-	7471B	PPB	24 Mar 2022 13:33:22	U	SMPL
CRA - 1	CRA	1257	0.2042	-	7471B	PPB	24 Mar 2022 13:35:41	U	SMPL
HighStd - 1	HighStd	48365	10.2462	-	7471B	PPB	24 Mar 2022 13:37:56	U	SMPL
ChkStd - 1	ChkStd	32055	6.7694	-	7471B	PPB	24 Mar 2022 13:40:12	U	SMPL
PB143560BL - 1	PBS	41	-0.0550	-	7471B	PPB	24 Mar 2022 13:42:33	U	SMPL
PB143560BS - 1	LCSS	16732	3.5030	-	7471B	PPB	24 Mar 2022 13:44:52	U	SMPL
N2061-01 - 1	RW-1-20220321-16.0-17.0	21128	4.4401	-	7471B	PPB	24 Mar 2022 13:47:09	U	SMPL
N2061-02 - 1	RW-1-20220321-19.0-20.0	29565	6.2386	-	7471B	PPB	24 Mar 2022 13:49:28	U	SMPL
N2063-01 - 1	WC-1-(10-10.5)	9087	1.8733	-	7471B	PPB	24 Mar 2022 13:51:48	U	SMPL
N2063-01DUP - 1	WC-1-(10-10.5)DUP	10291	2.1299	-	7471B	PPB	24 Mar 2022 13:54:08	U	SMPL
N2063-01MS - 1	WC-1-(10-10.5)MS	30520	6.4421	-	7471B	PPB	24 Mar 2022 14:00:09	U	SMPL
CCV14 - 1	CCV14	24392	5.1358	-	7471B	PPB	24 Mar 2022 14:02:24	U	SMPL
CCB14 - 1	CCB14	106	-0.0412	-	7471B	PPB	24 Mar 2022 14:04:43	U	SMPL
N2063-01MSD - 1	WC-1-(10-10.5)MSD	31128	6.5718	-	7471B	PPB	24 Mar 2022 14:07:01	U	SMPL
N2063-02 - 1	WC-2-(10-10.5)	249	-0.0107	-	7471B	PPB	24 Mar 2022 14:09:17	U	SMPL
N2063-03 - 1	WC-3-(10-10.5)	376	0.0164	-	7471B	PPB	24 Mar 2022 14:11:35	U	SMPL
N2063-04 - 1	WC-4-1-(10-10.5)	24507	5.1604	-	7471B	PPB	24 Mar 2022 14:13:51	U	SMPL
N2063-05 - 1	WC-4-2-(10-10.5)	8336	1.7132	-	7471B	PPB	24 Mar 2022 14:16:07	U	SMPL
N2063-06 - 1	WC-5-(10-10.5)	469	0.0362	-	7471B	PPB	24 Mar 2022 14:18:26	U	SMPL
N2063-07 - 1	WC-7-1-(10-10.5)	1990	0.3604	-	7471B	PPB	24 Mar 2022 14:20:45	U	SMPL
N2063-08 - 1	WC-7-2-(10-10.5)	1176	0.1869	-	7471B	PPB	24 Mar 2022 14:23:02	U	SMPL
N2063-09 - 1	WC-8-(10-10.5)	500	0.0428	-	7471B	PPB	24 Mar 2022 14:25:20	U	SMPL
N2063-10 - 1	WC-9-(10-10.5)	1980	0.3583	-	7471B	PPB	24 Mar 2022 14:27:36	U	SMPL
CCV15 - 1	CCV15	25142	5.2957	-	7471B	PPB	24 Mar 2022 14:29:51	U	SMPL
CCB15 - 1	CCB15	120	-0.0382	-	7471B	PPB	24 Mar 2022 14:32:07	U	SMPL
N2070-01 - 1	285	2071	0.3777	-	7471B	PPB	24 Mar 2022 14:34:26	U	SMPL
N2070-03 - 1	VNJ-212	2188	0.4026	-	7471B	PPB	24 Mar 2022 14:36:42	U	SMPL
N2071-01 - 1	UNDER-PILE	101952	21.6693	-	7471B	PPB	24 Mar 2022 14:38:57	U	SMPL
N2071-04 - 1	PILE	23647	4.9770	-	7471B	PPB	24 Mar 2022 14:41:13	U	SMPL
N2072-01 - 1	EO-03-32222	1654	0.2888	-	7471B	PPB	24 Mar 2022 14:43:40	U	SMPL
PB143587BL - 1	PBS	197	-0.0218	-	7471B	PPB	24 Mar 2022 14:45:59	U	SMPL
PB143587BS - 1	LCSS	17919	3.7560	-	7471B	PPB	24 Mar 2022 14:48:16	U	SMPL
N2077-03 - 1	WC-14A-4	4755	0.9498	-	7471B	PPB	24 Mar 2022 14:50:33	U	SMPL
N2077-07 - 1	WC-14A-3-1	30741	6.4893	-	7471B	PPB	24 Mar 2022 14:52:52	U	SMPL
N2077-11 - 1	WC-14A-3-2	59636	12.6488	-	7471B	PPB	24 Mar 2022 14:55:09	U	SMPL
CCV16 - 1	CCV16	24816	5.2262	-	7471B	PPB	24 Mar 2022 14:57:31	U	SMPL
CCB16 - 1	CCB16	114	-0.0395	-	7471B	PPB	24 Mar 2022 14:59:56	U	SMPL
N2077-15 - 1	WC-14A-3-3	13062	2.7206	-	7471B	PPB	24 Mar 2022 15:02:15	U	SMPL
N2084-01 - 1	BLDG-3	335595	71.4748	-	7471B	PPB	24 Mar 2022 15:04:31	U	SMPL
N2089-01 - 1	CL-01-032322	2809	0.5350	-	7471B	PPB	24 Mar 2022 15:18:30	U	SMPL
N2089-01DUP - 1	CL-01-032322DUP	2590	0.4883	-	7471B	PPB	24 Mar 2022 15:20:45	U	SMPL
N2089-01MS - 1	CL-01-032322MS	22007	4.6274	-	7471B	PPB	24 Mar 2022 15:23:02	U	SMPL
N2089-01MSD - 1	CL-01-032322MSD	21687	4.5592	-	7471B	PPB	24 Mar 2022 15:25:19	U	SMPL
N2091-01 - 1	MHB	223	-0.0162	-	7471B	PPB	24 Mar 2022 15:27:38	U	SMPL
CCV17 - 1	CCV17	24587	5.1774	-	7471B	PPB	24 Mar 2022 15:29:57	U	SMPL
CCB17 - 1	CCB17	156	-0.0305	-	7471B	PPB	24 Mar 2022 15:32:15	U	SMPL
N2063-01LX5 - 1		1822	0.3246	-	7471B	PPB	24 Mar 2022 15:34:33	U	SMPL
N2063-01A - 1		29023	6.1230	-	7471B	PPB	24 Mar 2022 15:36:49	U	SMPL
N2089-01LX5 - 1		348	0.0104	-	7471B	PPB	24 Mar 2022 15:39:05	U	SMPL
N2089-01A - 1		20916	4.3949	-	7471B	PPB	24 Mar 2022 15:41:23	U	SMPL
N2071-01DLX5 - 1		21966	4.6187	-	7471B	PPB	24 Mar 2022 15:43:40	U	SMPL
N2077-11DLX2 - 1		29807	6.2902	-	7471B	PPB	24 Mar 2022 15:45:58	U	SMPL
N2084-01DLX10 - 1		38921	8.2330	-	7471B	PPB	24 Mar 2022 15:48:16	U	SMPL
CCV18 - 1	CCV18	24768	5.2160	-	7471B	PPB	24 Mar 2022 15:52:18	U	SMPL
CCB18 - 1	CCB18	121	-0.0380	-	7471B	PPB	24 Mar 2022 15:54:34	U	SMPL