



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB119341

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	03/24/2022	16:10
2	CAL2	1	Water	NA	NA	20.3	7.00	03/24/2022	16:11
3	CAL3	1	Water	NA	NA	20.3	10.03	03/24/2022	16:12
4	ICV	1	Water	NA	NA	20.4	6.99	03/24/2022	16:15
5	CCV1	1	Water	NA	NA	20.4	2.01	03/24/2022	16:19
6	N2092-01	1	Water	NA	NA	20.5	7.30	03/24/2022	16:35
7	N2099-01	1	Water	NA	NA	20.1	6.69	03/24/2022	16:38
8	N2104-01	1	Water	NA	NA	20.2	7.31	03/24/2022	16:40
9	N2104-01DUP	1	Water	NA	NA	20.2	7.32	03/24/2022	16:41
10	N2106-12	1	Water	NA	NA	20.2	7.34	03/24/2022	16:45
11	CCV2	1	Water	NA	NA	20.1	12.02	03/24/2022	16:47

WORKLIST(Hardcopy Internal Chain)

28 11934)

WorkList Name : PH N2106 WorkList ID : 158034 Department : Wet-Chemistry Date : 03-24-2022 15:35:03

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample		Collect Date	Method
						Storage	Location		
03/24/2022	Water	N2092-01	pH	Cool 4 deg C	PSEG03	E12	FMI120	03/24/2022	9040C
03/24/2022	Water	N2099-01	pH	Cool 4 deg C	PSEG03	E12	T-2	03/24/2022	9040C
03/24/2022	Water	N2104-01	pH	Cool 4 deg C	PSEG03	E12	FRAC-TANK-A229	03/24/2022	9040C
03/24/2022	Water	N2106-12	pH	Cool 4 deg C	PSEG03	E21	OILY-WATER-DRUM-1-2	03/24/2022	9040C

Date/Time 03/24/22 16:10
 Raw Sample Received by: AD GGC
 Raw Sample Relinquished by: J. C. (Sant)

Date/Time 03/24/22 18:00
 Raw Sample Received by: J. C. (Sant)
 Raw Sample Relinquished by: J. C. (Sant)