



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Iwona

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB119345

Reagent/Standard	Lot/Log #
Residual chlorine LOD, 0.05PPM	WP95988
Residual chlorine ICV-LCS, 0.4PPM	WP95987
Chlorine Calibration std, 0.1ppm	WP95982
Chlorine Calibration std, 0.2ppm	WP95983
Chlorine Calibration std, 0.8ppm	WP95985
Chlorine Calibration std, 0.0ppm	WP95981
Chlorine Calibration std, 1.6ppm	WP95986
Residual Chlorine Calibration and CCV std, 0	WP95984
Total Chlorine Powder Pillows	W2759

Intercept: 0.0046

Slope: 1.0137

Regression: 0.999908

Seq	Lab ID	True Val (mg/l)	DF	Initial Reading	Final Reading	Difference	Result (mg/l)	%D	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	-0.01		03/25/2022	09:30
2	CAL2	0.1	1	0.000	0.110	0.110	0.10	-4	03/25/2022	09:33
3	CAL3	0.2	1	0.000	0.200	0.200	0.19	3.5	03/25/2022	09:36
4	CAL4	0.4	1	0.000	0.410	0.410	0.40	0	03/25/2022	09:39
5	CAL5	0.8	1	0.000	0.830	0.830	0.81	-1.7	03/25/2022	09:42
6	CAL6	1.6	1	0.000	1.620	1.620	1.59	0.4	03/25/2022	09:45



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Iwona

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB119345

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.0000	0.4200	0.4200	0.4100	03/25/2022	09:48
2	ICB				1	0.0000	0.0100	0.0100	0.0060	03/25/2022	09:51
3	CCV1			0.4	1	0.0000	0.3900	0.3900	0.3810	03/25/2022	09:54
4	CCB1				1	0.0000	0.0000	0.0000	-0.0040	03/25/2022	09:57
5	LB119345BL	50	50		1	0.0000	0.0100	0.0100	0.0060	03/25/2022	10:00
6	LB119345BS	50	50	0.4	1	0.0000	0.4100	0.4100	0.4000	03/25/2022	10:03
7	N1645-09	50	50		1	0.0000	0.0500	0.0500	0.0450	03/25/2022	10:06
8	CCV2			0.4	1	0.0000	0.4000	0.4000	0.3910	03/25/2022	10:09
9	CCB2				1	0.0000	0.0000	0.0000	-0.0040	03/25/2022	10:12