

LB119346


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

51138



A= 0.0000e+000

B= 1.9572e-004

C= 1.0694e-002

Rho= 0.9998359

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.022	0.022	58	0.000	58					
0.2	0.200	0.239	0.039	1164	0.0 %	1164					20
2.5	2.500	2.381	-0.119	12113	0.0 %	12113					-5
5.0	5.000	5.089	0.089	25945	0.0 %	25945					2
7.5	7.500	7.450	-0.050	38007	0.0 %	38007					-1
10.0	10.000	10.020	0.020	51138	0.0 %	51138					0

LB119346 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 25 Mar 2022 11:15:43

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	58	-	0.00007471B	PPB		25 Mar 2022 11:23:16	S	Std
0.2 - 1	50.2	1164	-	0.20007471B	PPB		25 Mar 2022 11:25:34	S	Std
2.5 - 1	52.5	12113	-	2.50007471B	PPB		25 Mar 2022 11:27:51	S	Std
5.0 - 1	55	25945	-	5.00007471B	PPB		25 Mar 2022 11:30:08	S	Std
7.5 - 1	57.5	38007	-	7.50007471B	PPB		25 Mar 2022 11:32:27	S	Std
10.0 - 1	510	51138	-	10.00007471B	PPB		25 Mar 2022 11:37:25	S	Std
ICV45 - 1	ICV45	19855	3.8968	-7471B	PPB		25 Mar 2022 11:40:11	U	SMPL
ICB45 - 1	ICB45	-92	-0.0073	-7471B	PPB		25 Mar 2022 11:42:26	U	SMPL
CCV19 - 1	CCV19	25978	5.0952	-7471B	PPB		25 Mar 2022 11:44:44	U	SMPL
CCB19 - 1	CCB19	-108	-0.0104	-7471B	PPB		25 Mar 2022 11:47:00	U	SMPL
CRA - 1	CRA	1148	0.2354	-7471B	PPB		25 Mar 2022 11:49:19	U	SMPL
HighStd - 1	HighStd	51760	10.1414	-7471B	PPB		25 Mar 2022 11:51:34	U	SMPL
ChkStd - 1	ChkStd	35640	6.9863	-7471B	PPB		25 Mar 2022 11:53:50	U	SMPL
PB143595BL - 1	PBS	-143	-0.0173	-7471B	PPB		25 Mar 2022 11:56:12	U	SMPL
PB143595BS - 1	LCSS	19863	3.8984	-7471B	PPB		25 Mar 2022 12:05:37	U	SMPL
N2094-01 - 1	WC-2	20887	4.0988	-7471B	PPB		25 Mar 2022 12:07:55	U	SMPL
N2094-01DUP - 1	WC-2DUP	20931	4.1074	-7471B	PPB		25 Mar 2022 12:10:14	U	SMPL
N2094-01MS - 1	WC-2MS	40024	7.8443	-7471B	PPB		25 Mar 2022 12:15:38	U	SMPL
N2094-01MSD - 1	WC-2MSD	38096	7.4670	-7471B	PPB		25 Mar 2022 12:17:56	U	SMPL
N2102-01 - 1	HD-02-32422	553	0.1189	-7471B	PPB		25 Mar 2022 12:20:15	U	SMPL
CCV20 - 1	CCV20	24652	4.8357	-7471B	PPB		25 Mar 2022 12:22:35	U	SMPL
CCB20 - 1	CCB20	-99	-0.0087	-7471B	PPB		25 Mar 2022 12:24:50	U	SMPL
N2103-01 - 1	CHRT-2001	36586	7.1714	-7471B	PPB		25 Mar 2022 12:27:09	U	SMPL
N2103-03 - 1	RBR-200055	26712	5.2389	-7471B	PPB		25 Mar 2022 12:29:24	U	SMPL
N2103-05 - 1	RBR-200031	20733	4.0686	-7471B	PPB		25 Mar 2022 12:31:47	U	SMPL
N2103-07 - 1	RBR-200049	10865	2.1372	-7471B	PPB		25 Mar 2022 12:34:06	U	SMPL
N2103-09 - 1	CHRT-20019	8588	1.6916	-7471B	PPB		25 Mar 2022 12:36:25	U	SMPL
N2106-01 - 1	OILY-DEBRIS-DRUM-1	246	0.0588	-7471B	PPB		25 Mar 2022 12:38:44	U	SMPL
N2108-01 - 1	R206	1066	0.2193	-7471B	PPB		25 Mar 2022 12:41:02	U	SMPL
N2109-01 - 1	TP-1	2105	0.4227	-7471B	PPB		25 Mar 2022 12:43:20	U	SMPL
N2094-01LX5 - 1		4447	0.8811	-7471B	PPB		25 Mar 2022 12:45:37	U	SMPL
N2094-01A - 1		39345	7.7114	-7471B	PPB		25 Mar 2022 12:57:43	U	SMPL
CCV21 - 1	CCV21	23840	4.6768	-7471B	PPB		25 Mar 2022 12:59:59	U	SMPL
CCB21 - 1	CCB21	-162	-0.0210	-7471B	PPB		25 Mar 2022 13:02:19	U	SMPL