



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB119352

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	03/25/2022	11:33
2	CAL2	1	Water	NA	NA	20.4	7.00	03/25/2022	11:34
3	CAL3	1	Water	NA	NA	20.4	10.03	03/25/2022	11:35
4	ICV	1	Water	NA	NA	20.3	6.99	03/25/2022	11:36
5	CCV1	1	Water	NA	NA	20.3	2.01	03/25/2022	11:38
6	N2114-01	1	Water	NA	NA	20.3	6.40	03/25/2022	11:40
7	N2114-01DUP	1	Water	NA	NA	20.3	6.41	03/25/2022	11:41
8	CCV2	1	Water	NA	NA	20.3	12.03	03/25/2022	11:45

119352

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N2114 WorkList ID : 158061 Department : Wet-Chemistry Date : 03-25-2022 11:25:44

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
03/23/2022	Water	N2114-01	pH	Cool 4 deg C	STER14	N22	JS10000	03/23/2022	9040C

Date/Time 03/25/22 11:30
Raw Sample Received by: JP (WCC)
Raw Sample Relinquished by: CR SM

Date/Time 03/25/22 12:30
Raw Sample Received by: JP SM
Raw Sample Relinquished by: JP (WCC)