

LB119357


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

37146



A= 0.0000e+000

B= 2.6602e-004 slope

C= -2.1396e-002 y-intercept

Rho= 0.9996032

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%DI
0.00	0.000	0.003	0.003	92	0.000	92					7
0.05	0.050					185	0				3
0.20	0.200	0.205	0.005	852	0.0 %	852					-5
2.50	2.500	2.387	-0.113	9053	0.0 %	9053					2
5.00	5.000	5.088	0.088	19206	0.0 %	19206					2
7.50	7.500	7.657	0.157	28865	0.0 %	28865					-1
10.0	10.000	9.860	-0.140	37146	0.0 %	37146					

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	92	-		0 SFAM01.1	PPB	3/25/2022 14:18	Std
	0.2 S01	852	-		0.2 SFAM01.1	PPB	3/25/2022 14:20	Std
	2.5 S02	9053	-		2.5 SFAM01.1	PPB	3/25/2022 14:22	Std
	5 S03	19206	-		5 SFAM01.1	PPB	3/25/2022 14:24	Std
	7.5 S04	28865	-		7.5 SFAM01.1	PPB	3/25/2022 14:27	Std
	10 S05	37146	-		10 SFAM01.1	PPB	3/25/2022 14:29	Std
ICV046	ICV046	15244	4.0337	-	SFAM01.1	PPB	3/25/2022 14:35	SMPL
ICB046	ICB046	25	-0.0147	-	SFAM01.1	PPB	3/25/2022 14:37	SMPL
CCV022	CCV022	18678	4.9472	-	SFAM01.1	PPB	3/25/2022 14:39	SMPL
CCB022	CCB022	3	-0.0206	-	SFAM01.1	PPB	3/25/2022 14:41	SMPL
PB143623BL	PBW623	63	-0.0046	-	SFAM01.1	PPB	3/25/2022 14:44	SMPL
N1842-01	YAYB4	209858	55.804	-	SFAM01.1	PPB	3/25/2022 14:46	SMPL
N1842-02	YAYB8	-663	-0.1978	-	SFAM01.1	PPB	3/25/2022 14:48	SMPL
N1842-03	YAYC6	32844	8.7156	-	SFAM01.1	PPB	3/25/2022 14:51	SMPL
N1842-04	YAYD2	33670	8.9353	-	SFAM01.1	PPB	3/25/2022 14:53	SMPL
N1842-05	YAYF0	196653	52.2913	-	SFAM01.1	PPB	3/25/2022 14:55	SMPL
N1842-06	YAYF2	-81	-0.0429	-	SFAM01.1	PPB	3/25/2022 15:00	SMPL
N1842-07	YAY98	17724	4.6935	-	SFAM01.1	PPB	3/25/2022 15:03	SMPL
N1842-08	YAYA4	266623	70.9044	-	SFAM01.1	PPB	3/25/2022 15:05	SMPL
N1842-10	YAYB0	29666	7.8702	-	SFAM01.1	PPB	3/25/2022 15:10	SMPL
N1842-11	YAYB0D	30108	7.9878	-	SFAM01.1	PPB	3/25/2022 15:12	SMPL
N1842-13	YAYB2	9	-0.019	-	SFAM01.1	PPB	3/25/2022 15:17	SMPL
N1842-14	YAYG5	246	0.044	-	SFAM01.1	PPB	3/25/2022 15:19	SMPL
N1842-15	YAYG7	56	-0.0065	-	SFAM01.1	PPB	3/25/2022 15:21	SMPL
N1842-16	YAYH3	196	0.0307	-	SFAM01.1	PPB	3/25/2022 15:23	SMPL
N1842-17	YAYJ1	293	0.0565	-	SFAM01.1	PPB	3/25/2022 15:26	SMPL
N1842-18	YAYJ3	308	0.0605	-	SFAM01.1	PPB	3/25/2022 15:28	SMPL
N1842-19	YAYJ5	1870	0.4761	-	SFAM01.1	PPB	3/25/2022 15:30	SMPL
CCV023	CCV023	19058	5.0483	-	SFAM01.1	PPB	3/25/2022 15:32	SMPL
CCB023	CCB023	22	-0.0155	-	SFAM01.1	PPB	3/25/2022 15:35	SMPL
PB143624BL	PBW624	187	0.0283	-	SFAM01.1	PPB	3/25/2022 15:37	SMPL
N1850-01	YAYB1	474	0.1047	-	SFAM01.1	PPB	3/25/2022 15:39	SMPL
N1850-02	YAYB3	241886	64.324	-	SFAM01.1	PPB	3/25/2022 15:42	SMPL
N1850-03	YAYG4	5335	1.3978	-	SFAM01.1	PPB	3/25/2022 15:44	SMPL
N1850-04	YAYG6	46	-0.0092	-	SFAM01.1	PPB	3/25/2022 15:46	SMPL
N1850-05	YAYH2	321	0.064	-	SFAM01.1	PPB	3/25/2022 15:49	SMPL
N1850-06	YAYJ0	372	0.0776	-	SFAM01.1	PPB	3/25/2022 15:51	SMPL
N1850-07	YAYJ2	10473	2.7646	-	SFAM01.1	PPB	3/25/2022 15:53	SMPL
N1850-08	YAYJ4	2000	0.5106	-	SFAM01.1	PPB	3/25/2022 15:55	SMPL
N1850-09	YAYG3	342	0.0696	-	SFAM01.1	PPB	3/25/2022 15:58	SMPL
N1850-10	YAYH0	220	0.0371	-	SFAM01.1	PPB	3/25/2022 16:00	SMPL
N1850-11	YAYH4	239	0.0422	-	SFAM01.1	PPB	3/25/2022 16:02	SMPL
N1850-12	YAYH6	481	0.1066	-	SFAM01.1	PPB	3/25/2022 16:04	SMPL
N1850-13	YAYH8	280	0.0531	-	SFAM01.1	PPB	3/25/2022 16:07	SMPL
N1850-14	YAYJ6	552	0.1254	-	SFAM01.1	PPB	3/25/2022 16:09	SMPL
N1850-15	YAYG8	156	0.0201	-	SFAM01.1	PPB	3/25/2022 16:11	SMPL

LB119357

INSTRUMENT ID : CV1

N1850-16	YAYG8D	129	0.0129 -	SFAM01.1	PPB	3/25/2022 16:14 SMPL
PB143625BL	PBW625	62	-0.0049 -	SFAM01.1	PPB	3/25/2022 16:18 SMPL
N1940-01	YAYH7	153	0.0193 -	SFAM01.1	PPB	3/25/2022 16:21 SMPL
CCV024	CCV024	19340	5.1233 -	SFAM01.1	PPB	3/25/2022 16:23 SMPL
CCB024	CCB024	30	-0.0134 -	SFAM01.1	PPB	3/25/2022 16:25 SMPL
N1940-02	YAYG9	143	0.0166 -	SFAM01.1	PPB	3/25/2022 16:28 SMPL
N1940-03	YAYG9D	116	0.0095 -	SFAM01.1	PPB	3/25/2022 16:30 SMPL
N1940-04	YAYG9S	3409	0.8855 -	SFAM01.1	PPB	3/25/2022 16:35 SMPL
N1940-05	YAYH1	91	0.0028 -	SFAM01.1	PPB	3/25/2022 16:37 SMPL
N1940-06	YAYH5	63	-0.0046 -	SFAM01.1	PPB	3/25/2022 16:39 SMPL
N1940-07	YAYH9	175	0.0252 -	SFAM01.1	PPB	3/25/2022 16:41 SMPL
N1940-08	YAYJ7	119	0.0103 -	SFAM01.1	PPB	3/25/2022 16:44 SMPL
N1842-09	YAYA6	123	0.0113 -	SFAM01.1	PPB	3/25/2022 16:46 SMPL
N1842-01DLX10	YAYB4	24709	6.5516 -	SFAM01.1	PPB	3/25/2022 16:48 SMPL
N1842-05DLX10	YAYF0	22298	5.9102 -	SFAM01.1	PPB	3/25/2022 16:51 SMPL
N1842-08DLX25	YAYA4	12384	3.2729 -	SFAM01.1	PPB	3/25/2022 16:53 SMPL
N1842-12	YAYB0S	33616	8.921 -	SFAM01.1	PPB	3/25/2022 16:58 SMPL
N1850-02DLX10	YAYB3	25546	6.7742 -	SFAM01.1	PPB	3/25/2022 17:00 SMPL
N1850-17	YAYG8S	3918	1.0209 -	SFAM01.1	PPB	3/25/2022 17:09 SMPL
CCV025	CCV025	19545	5.1779 -	SFAM01.1	PPB	3/25/2022 17:11 SMPL
CCB025	CCB25	31	-0.0131 -	SFAM01.1	PPB	3/25/2022 17:13 SMPL