



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB119362

Slope : 98.7

BalanceID: WC SC-6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	03/26/2022	07:45
2	CAL2	1	Water	NA	NA	20.3	7.00	03/26/2022	07:46
3	CAL3	1	Water	NA	NA	20.3	10.03	03/26/2022	07:47
4	ICV	1	Water	NA	NA	20.4	7.00	03/26/2022	07:49
5	CCV1	1	Water	NA	NA	20.4	2.01	03/26/2022	07:50
6	N2133-01	1	Solid	20.03	20	20.6	7.36	03/26/2022	07:52
7	N2133-02	1	Solid	20.04	20	20.5	6.81	03/26/2022	07:55
8	N2133-03	1	Solid	20.04	20	20.4	3.26	03/26/2022	08:00
9	N2133-04	1	Solid	20.02	20	20.1	5.61	03/26/2022	08:02
10	N2134-01	1	Solid	20.03	20	21.6	8.70	03/26/2022	08:05
11	N2140-02	1	Solid	20.04	20	21.0	9.40	03/26/2022	08:09
12	N2140-02DUP	1	Solid	20.03	20	21.1	9.41	03/26/2022	08:10
13	CCV2	1	Water	NA	NA	20.7	12.03	03/26/2022	08:11

VB 119362

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY N2133 s WorkList ID : 158097 Department : Wet-Chemistry Date : 03-26-2022 07:37:01

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample		Collect Date	Method
						Storage Location	Customer Sample		
03/25/2022	Solid	N2133-01	Corrosivity	Cool 4 deg C	PSEG03	E21	COMP-1	03/25/2022	9045D
03/25/2022	Solid	N2133-02	Corrosivity	Cool 4 deg C	PSEG03	E21	COMP-2	03/25/2022	9045D
03/25/2022	Solid	N2133-03	Corrosivity	Cool 4 deg C	PSEG03	E21	COMP-3	03/25/2022	9045D
03/25/2022	Solid	N2133-04	Corrosivity	Cool 4 deg C	PSEG03	E21	1934	03/25/2022	9045D
03/25/2022	Solid	N2134-01	Corrosivity	Cool 4 deg C	TULL02	E31	PIT-11-STOCKPILE	03/25/2022	9045D
03/25/2022	Solid	N2140-02	Corrosivity	Cool 4 deg C	PSEG03	E11	MHC	03/25/2022	9045D

Date/Time 03/26/22 07:42
 Raw Sample Received by: JO (WCC)
 Raw Sample Relinquished by: CP SM

Date/Time 03/26/22 09:20
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: JO WCC