



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB119363

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	03/26/2022	07:45
2	CAL2	1	Water	NA	NA	20.3	7.00	03/26/2022	07:46
3	CAL3	1	Water	NA	NA	20.3	10.03	03/26/2022	07:47
4	ICV	1	Water	NA	NA	20.4	7.00	03/26/2022	07:49
5	CCV1	1	Water	NA	NA	20.4	2.01	03/26/2022	07:50
6	N2133-05	1	Water	NA	NA	20.3	7.61	03/26/2022	08:15
7	N2133-06	1	Water	NA	NA	20.1	6.81	03/26/2022	08:16
8	N2133-06DUP	1	Water	NA	NA	20.2	6.82	03/26/2022	08:17
9	N2133-07	1	Water	NA	NA	20.5	5.53	03/26/2022	08:19
10	CCV2	1	Water	NA	NA	20.4	12.03	03/26/2022	08:22

215 119363

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY N2133 w WorkList ID : 158098 Department : Wet-Chemistry Date : 03-26-2022 07:37:57

Due Date	Matrix	Sample	Test	Raw Sample			Collect Date	Method
				Preservative	Customer	Raw Sample Storage Location		
03/25/2022	Water	N2133-05	Corrosivity	Cool 4 deg C	PSEG03	E21 COMP-4	03/25/2022	9040C
03/25/2022	Water	N2133-06	Corrosivity	Cool 4 deg C	PSEG03	E21 1933	03/25/2022	9040C
03/25/2022	Water	N2133-07	Corrosivity	Cool 4 deg C	PSEG03	E21 1932	03/25/2022	9040C

Date/Time 03/26/22 07:42
Raw Sample Received by: 206001
Raw Sample Relinquished by: CP SM

Date/Time 03/26/22 09:30
Raw Sample Received by: CP SM
Raw Sample Relinquished by: 206001