

LB119365


SFAM01.1

INSTRUMENT ID: CV4

Linear

 μ Abs.:

38704



A= 0.0000e+000

B= 2.5603e-004 slope

C= -1.9501e-003 intercept

Rho= 0.9997084

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.021	0.021	89	0.000	89					=
0.05	0.050					185	0				17
0.20	0.200	0.233	0.033	916	0.0 %	916					-4
2.50	2.500	2.388	-0.112	9334	0.0 %	9334					0
5.00	5.000	4.993	-0.007	19510	0.0 %	19510					2
7.50	7.500	7.658	0.158	29918	0.0 %	29918					-1
10.0	10.000	9.908	-0.092	38704	0.0 %	38704					

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	89	-		0 SFAM01.1	PPB	3/26/2022 8:56	Std
	0.2 S01	916	-		0.2 SFAM01.1	PPB	3/26/2022 8:58	Std
	2.5 S02	9334	-		2.5 SFAM01.1	PPB	3/26/2022 9:00	Std
	5 S03	19510	-		5 SFAM01.1	PPB	3/26/2022 9:03	Std
	7.5 S04	29918	-		7.5 SFAM01.1	PPB	3/26/2022 9:05	Std
	10 S05	38704	-		10 SFAM01.1	PPB	3/26/2022 9:07	Std
ICV047	ICV047	15607	3.9939 -		SFAM01.1	PPB	3/26/2022 9:10	SMPL
ICB047	ICB047	7	-0.0002 -		SFAM01.1	PPB	3/26/2022 9:13	SMPL
CCV026	CCV026	19445	4.9766 -		SFAM01.1	PPB	3/26/2022 9:15	SMPL
CCB026	CCB026	0	-0.002 -		SFAM01.1	PPB	3/26/2022 9:17	SMPL
PB143649BL	PBW649	12	0.0011 -		SFAM01.1	PPB	3/26/2022 9:19	SMPL
N1841-01	YAYB7	1039	0.2641 -		SFAM01.1	PPB	3/26/2022 9:22	SMPL
N1841-02	YAYC5	39899	10.2135 -		SFAM01.1	PPB	3/26/2022 9:24	SMPL
N1841-03	YAYD1	138469	35.4505 -		SFAM01.1	PPB	3/26/2022 9:26	SMPL
N1841-04	YAYE3	926	0.2351 -		SFAM01.1	PPB	3/26/2022 9:28	SMPL
N1841-05	YAYE4	116	0.0277 -		SFAM01.1	PPB	3/26/2022 9:31	SMPL
N1841-06	YAYE5	512	0.1291 -		SFAM01.1	PPB	3/26/2022 9:33	SMPL
N1841-07	YAYE6	32	0.0062 -		SFAM01.1	PPB	3/26/2022 9:35	SMPL
N1841-08	YAYE9	252526	64.6527 -		SFAM01.1	PPB	3/26/2022 9:38	SMPL
N1841-09	YAYF1	69	0.0157 -		SFAM01.1	PPB	3/26/2022 9:43	SMPL
N1841-10	YAYF3	1159	0.2948 -		SFAM01.1	PPB	3/26/2022 9:45	SMPL
N1841-11	YAYF4	146	0.0354 -		SFAM01.1	PPB	3/26/2022 9:47	SMPL
N1841-12	YAYF8	46	0.0098 -		SFAM01.1	PPB	3/26/2022 9:49	SMPL
N1841-13	YAYF9	68	0.0155 -		SFAM01.1	PPB	3/26/2022 9:52	SMPL
N1841-14	YAYG0	597	0.1509 -		SFAM01.1	PPB	3/26/2022 9:54	SMPL
N1841-15	YAYG1	294	0.0733 -		SFAM01.1	PPB	3/26/2022 9:56	SMPL
N1841-16	YAYG2	33	0.0065 -		SFAM01.1	PPB	3/26/2022 9:58	SMPL
N1841-19	YAY97	608346	HIGH -		SFAM01.1	PPB	3/26/2022 10:01	
N1841-20	YAYA3	290387	74.3463 -		SFAM01.1	PPB	3/26/2022 10:03	SMPL
CCV027	CCV027	19362	4.9553 -		SFAM01.1	PPB	3/26/2022 10:07	SMPL
CCB027	CCB027	-36	-0.0112 -		SFAM01.1	PPB	3/26/2022 10:10	SMPL
N1841-21	YAYA5	166	0.0406 -		SFAM01.1	PPB	3/26/2022 10:12	SMPL
N1841-22	YAYA9	33478	8.5695 -		SFAM01.1	PPB	3/26/2022 10:14	SMPL
N1841-23	YAYA9D	34082	8.7241 -		SFAM01.1	PPB	3/26/2022 10:17	SMPL
N1841-24	YAYA9S	37669	9.6425 -		SFAM01.1	PPB	3/26/2022 10:24	SMPL
N1841-02DLX2	YAYC5	19898	5.0926 -		SFAM01.1	PPB	3/26/2022 10:27	SMPL
N1841-03DLX10	YAYD1	13808	3.5333 -		SFAM01.1	PPB	3/26/2022 10:29	SMPL
N1841-08DLX10	YAYE9	27715	7.094 -		SFAM01.1	PPB	3/26/2022 10:34	SMPL
N1841-19DLX10	YAY97	104762	26.8204 -		SFAM01.1	PPB	3/26/2022 10:39	SMPL
N1841-20DLX10	YAYA3	28103	7.1933 -		SFAM01.1	PPB	3/26/2022 10:41	SMPL
N1841-19DL2X50	YAY97	18392	4.707 -		SFAM01.1	PPB	3/26/2022 10:43	SMPL
CCV028	CCV028	19171	4.9064 -		SFAM01.1	PPB	3/26/2022 10:46	SMPL
CCB028	CCB028	-69	-0.0196 -		SFAM01.1	PPB	3/26/2022 10:49	SMPL