

LB119366


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

40869



A= 0.0000e+000

B= 2.4868e-004 slope

C= -5.1652e-002 y-intercept

Rho= 0.9993514

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.060	-0.060	-34	0.000	-34					1
0.05	0.050					185	0				-8
0.20	0.200	0.185	-0.015	951	0.0 %	951					9
2.50	2.500	2.719	0.219	11140	0.0 %	11140					1
5.00	5.000	4.941	-0.059	20078	0.0 %	20078					-3
7.50	7.500	7.304	-0.196	29579	0.0 %	29579					1
10.0	10.000	10.111	0.111	40869	0.0 %	40869					

LB119366

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	-34	-		0 SFAM01.1	PPB	3/26/2022 13:57	Std
	0.2 S01	951	-		0.2 SFAM01.1	PPB	3/26/2022 13:59	Std
	2.5 S02	11140	-		2.5 SFAM01.1	PPB	3/26/2022 14:04	Std
	5 S03	20078	-		5 SFAM01.1	PPB	3/26/2022 14:06	Std
	7.5 S04	29579	-		7.5 SFAM01.1	PPB	3/26/2022 14:09	Std
	10 S05	40869	-		10 SFAM01.1	PPB	3/26/2022 14:14	Std
ICV048	ICV048	16480	4.0465 -		SFAM01.1	PPB	3/26/2022 14:18	SMPL
ICB048	ICB048	50	-0.0392 -		SFAM01.1	PPB	3/26/2022 14:20	SMPL
CCV029	CCV029	20289	4.9937 -		SFAM01.1	PPB	3/26/2022 14:22	SMPL
CCB029	CCB029	44	-0.0407 -		SFAM01.1	PPB	3/26/2022 14:25	SMPL
PB143657BL	PBS657	26	-0.0452 -		SFAM01.1	PPB	3/26/2022 14:27	SMPL
N1844-04	DBSE7	265	0.0142 -		SFAM01.1	PPB	3/26/2022 14:36	SMPL
N1844-05	DBSE8	254	0.0115 -		SFAM01.1	PPB	3/26/2022 14:38	SMPL
N1844-06	DBSE9	301	0.0232 -		SFAM01.1	PPB	3/26/2022 14:40	SMPL
N1844-07	DBSF0	368	0.0399 -		SFAM01.1	PPB	3/26/2022 14:43	SMPL
N1844-08	DBSF1	593	0.0958 -		SFAM01.1	PPB	3/26/2022 14:45	SMPL
N1844-09	DBSF2	574	0.0911 -		SFAM01.1	PPB	3/26/2022 14:47	SMPL
N1844-10	DBSF3	462	0.0632 -		SFAM01.1	PPB	3/26/2022 14:49	SMPL
N1844-11	DBSF4	467	0.0645 -		SFAM01.1	PPB	3/26/2022 14:52	SMPL
N1844-12	DBSF4D	499	0.0724 -		SFAM01.1	PPB	3/26/2022 14:54	SMPL
N1844-13	DBSF4S	12014	2.9359 -		SFAM01.1	PPB	3/26/2022 14:56	SMPL
N1844-14	DBSF5	559	0.0874 -		SFAM01.1	PPB	3/26/2022 14:59	SMPL
N1844-15	DBSF6	339	0.0326 -		SFAM01.1	PPB	3/26/2022 15:01	SMPL
N1844-16	DBSF7	646	0.109 -		SFAM01.1	PPB	3/26/2022 15:03	SMPL
N1844-17	DBSF8	469	0.065 -		SFAM01.1	PPB	3/26/2022 15:05	SMPL
N1844-18	DBSF9	1302	0.2721 -		SFAM01.1	PPB	3/26/2022 15:08	SMPL
CCV030	CCV030	20423	5.0271 -		SFAM01.1	PPB	3/26/2022 15:10	SMPL
CCB030	CCB030	34	-0.0432 -		SFAM01.1	PPB	3/26/2022 15:12	SMPL
N1844-01	DBRQ7	8728	2.1188 -		SFAM01.1	PPB	3/26/2022 15:19	SMPL
N1844-02	DBRQ8	407180	HIGH -		SFAM01.1	PPB	3/26/2022 15:22	
N1844-02DLX10	DBRQ8	58772	14.5635 -		SFAM01.1	PPB	3/26/2022 15:30	SMPL
N1844-02DL2X50	DBRQ8	12133	2.9655 -		SFAM01.1	PPB	3/26/2022 15:37	SMPL
CCV031	CCV031	20022	4.9273 -		SFAM01.1	PPB	3/26/2022 15:39	SMPL
CCB031	CCB031	32	-0.0437 -		SFAM01.1	PPB	3/26/2022 15:41	SMPL