

L3119371


7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

35548



A= 0.0000e+000

B= 2.7680e-004

C= -3.5277e-002

Rho= 0.9992686

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	-0.039	-0.039	-15	0.000	-15					1/b
0.2	0.200	0.218	0.018	915	0.0 %	915					9
2.5	2.500	2.479	-0.021	9084	0.0 %	9084					-1
5.0	5.000	4.960	-0.040	18045	0.0 %	18045					-1
7.5	7.500	7.778	0.278	28227	0.0 %	28227					4
10.0	10.000	9.805	-0.195	35548	0.0 %	35548					-2

LB119371 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 28 Mar 2022 09:43:17					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-15	-	0.0007470A	PPB		28 Mar 2022 09:46:29	S	Std
0.2 - 1	50.2	915	-	0.20007470A	PPB		28 Mar 2022 09:51:42	S	Std
2.5 - 1	52.5	9084	-	2.50007470A	PPB		28 Mar 2022 09:53:59	S	Std
5.0 - 1	55	18045	-	5.00007470A	PPB		28 Mar 2022 09:56:16	S	Std
7.5 - 1	57.5	28227	-	7.50007470A	PPB		28 Mar 2022 09:58:33	S	Std
10.0 - 1	511	35548	-	10.00007470A	PPB		28 Mar 2022 10:00:50	S	Std
ICV49 - 1	ICV49	14980	4.1112	-7470A	PPB		28 Mar 2022 10:04:01	U	SMPL
ICB49 - 1	ICB49	-54	-0.0502	-7470A	PPB		28 Mar 2022 10:06:16	U	SMPL
CCV32 - 1	CCV32	18012	4.9505	-7470A	PPB		28 Mar 2022 10:11:32	U	SMPL
CCB32 - 1	CCB32	-63	-0.0527	-7470A	PPB		28 Mar 2022 10:13:47	U	SMPL
CRA - 1	CRA	935	0.2235	-7470A	PPB		28 Mar 2022 10:16:03	U	SMPL
HighStd - 1	HighStd	34051	9.3902	-7470A	PPB		28 Mar 2022 10:18:18	U	SMPL
ChkStd - 1	ChkStd	26642	7.3393	-7470A	PPB		28 Mar 2022 10:20:34	U	SMPL
PB143630BL - 1	PBW	-104	-0.0641	-7470A	PPB		28 Mar 2022 10:22:52	U	SMPL
PB143630BS - 1	LCSW	14088	3.8643	-7470A	PPB		28 Mar 2022 10:28:00	U	SMPL
N2085-01 - 1	GW-3-23-2022	41	-0.0239	-7470A	PPB		28 Mar 2022 10:30:16	U	SMPL
N2085-01DUP - 1	GW-3-23-2022DUP	70	-0.0159	-7470A	PPB		28 Mar 2022 10:32:33	U	SMPL
N2085-01MS - 1	GW-3-23-2022MS	14920	4.0946	-7470A	PPB		28 Mar 2022 10:34:50	U	SMPL
N2085-01MSD - 1	GW-3-23-2022MSD	14806	4.0631	-7470A	PPB		28 Mar 2022 10:37:08	U	SMPL
N2092-01 - 1	FMI120	3002	0.7957	-7470A	PPB		28 Mar 2022 10:39:24	U	SMPL
CCV33 - 1	CCV33	17672	4.8564	-7470A	PPB		28 Mar 2022 10:41:39	U	SMPL
CCB33 - 1	CCB33	-61	-0.0522	-7470A	PPB		28 Mar 2022 10:43:56	U	SMPL
N2099-01 - 1	T-2	2310	0.6041	-7470A	PPB		28 Mar 2022 10:59:36	U	SMPL
N2104-01 - 1	FRAC-TANK-A229	4386	1.1788	-7470A	PPB		28 Mar 2022 11:01:52	U	SMPL
N2106-12 - 1	OILY-WATER-DRUM-1-2	5120	1.3820	-7470A	PPB		28 Mar 2022 11:04:07	U	SMPL
N2123-04 - 1	20220344-MODIFIED-ELUTRIATE	3799	1.0163	-7470A	PPB		28 Mar 2022 11:06:25	U	SMPL
N2123-05 - 1	DISSOLVED-FRACTION-CS-1-MODIFIED-ELUTRIATE	529	0.1112	-7470A	PPB		28 Mar 2022 11:08:41	U	SMPL
PB143633BL - 1	PBW	47	-0.0223	-7470A	PPB		28 Mar 2022 11:10:57	U	SMPL
PB143633BS - 1	LCSW	14947	4.1021	-7470A	PPB		28 Mar 2022 11:16:13	U	SMPL
N2060-01 - 1	WC-20220321	30	-0.0270	-7470A	PPB		28 Mar 2022 11:18:30	U	SMPL
N2077-04 - 1	WC-14A-4	66	-0.0170	-7470A	PPB		28 Mar 2022 11:20:47	U	SMPL
N2077-08 - 1	WC-14A-3-1	63	-0.0178	-7470A	PPB		28 Mar 2022 11:23:03	U	SMPL
CCV34 - 1	CCV34	17014	4.6743	-7470A	PPB		28 Mar 2022 11:25:19	U	SMPL
CCB34 - 1	CCB34	-58	-0.0513	-7470A	PPB		28 Mar 2022 11:27:35	U	SMPL
N2077-12 - 1	WC-14A-3-2	70	-0.0159	-7470A	PPB		28 Mar 2022 11:29:52	U	SMPL
N2077-16 - 1	WC-14A-3-3	66	-0.0170	-7470A	PPB		28 Mar 2022 11:32:07	U	SMPL
N2084-02 - 1	BLDG-3	403	0.0763	-7470A	PPB		28 Mar 2022 11:34:23	U	SMPL
N2091-02 - 1	MHB	56	-0.0198	-7470A	PPB		28 Mar 2022 11:36:39	U	SMPL
N2091-02DUP - 1	MHBDUP	70	-0.0159	-7470A	PPB		28 Mar 2022 11:38:56	U	SMPL
N2091-02MS - 1	MHBMS	15916	4.3703	-7470A	PPB		28 Mar 2022 11:43:35	U	SMPL
N2091-02MSD - 1	MHBMSD	14496	3.9773	-7470A	PPB		28 Mar 2022 11:45:51	U	SMPL
PB143658BL - 1	PBW	-49	-0.0488	-7470A	PPB		28 Mar 2022 11:48:11	U	SMPL
PB143658BS - 1	LCSW	15902	4.3665	-7470A	PPB		28 Mar 2022 11:50:29	U	SMPL
N2123-03 - 1	SPLP-LEACHATE-AMENDED-COMP-CS-1	1453	0.3669	-7470A	PPB		28 Mar 2022 11:52:45	U	SMPL
CCV35 - 1	CCV35	19412	5.3380	-7470A	PPB		28 Mar 2022 11:55:01	U	SMPL
CCB35 - 1	CCB35	-43	-0.0472	-7470A	PPB		28 Mar 2022 11:57:17	U	SMPL
N2123-03DUP - 1	SPLP-LEACHATE-AMENDED-COMP-CS-1DUP	1487	0.3763	-7470A	PPB		28 Mar 2022 11:59:33	U	SMPL
N2123-03MS - 1	SPLP-LEACHATE-AMENDED-COMP-CS-1MS	15744	4.3227	-7470A	PPB		28 Mar 2022 12:01:49	U	SMPL
N2123-03MSD - 1	SPLP-LEACHATE-AMENDED-COMP-CS-1MSD	15711	4.3136	-7470A	PPB		28 Mar 2022 12:04:05	U	SMPL
PB143659BL - 1	PBW	-33	-0.0444	-7470A	PPB		28 Mar 2022 12:06:22	U	SMPL
PB143659BS - 1	LCSW	15846	4.3510	-7470A	PPB		28 Mar 2022 12:08:39	U	SMPL
N2125-02 - 1	RO-124030-EPH-2	48	-0.0220	-7470A	PPB		28 Mar 2022 12:10:55	U	SMPL
N2125-04 - 1	164718-EPH-2	79	-0.0134	-7470A	PPB		28 Mar 2022 12:13:12	U	SMPL
N2133-01 - 1	COMP-1	81	-0.0129	-7470A	PPB		28 Mar 2022 12:15:29	U	SMPL
N2133-02 - 1	COMP-2	111	-0.0046	-7470A	PPB		28 Mar 2022 12:17:46	U	SMPL
N2133-03 - 1	COMP-3	84	-0.0120	-7470A	PPB		28 Mar 2022 12:20:02	U	SMPL
CCV36 - 1	CCV36	18915	5.2005	-7470A	PPB		28 Mar 2022 12:22:19	U	SMPL
CCB36 - 1	CCB36	-47	-0.0483	-7470A	PPB		28 Mar 2022 12:24:36	U	SMPL
N2133-04 - 1	1934	102	-0.0070	-7470A	PPB		28 Mar 2022 12:26:52	U	SMPL
N2134-01 - 1	PIT-11-STOCKPILE	225	0.0270	-7470A	PPB		28 Mar 2022 12:29:08	U	SMPL
N2140-02 - 1	MHC	284	0.0433	-7470A	PPB		28 Mar 2022 12:31:24	U	SMPL
N2140-02DUP - 1	MHCDUP	217	0.0248	-7470A	PPB		28 Mar 2022 12:33:40	U	SMPL
N2140-02MS - 1	MHCM	15605	4.2842	-7470A	PPB		28 Mar 2022 12:35:57	U	SMPL
N2140-02MSD - 1	MHCMMSD	15498	4.2546	-7470A	PPB		28 Mar 2022 12:38:13	U	SMPL
PB143660BL - 1	PBW	-46	-0.0480	-7470A	PPB		28 Mar 2022 12:40:31	U	SMPL
PB143660BS - 1	LCSW	15257	4.1879	-7470A	PPB		28 Mar 2022 12:42:48	U	SMPL
N2133-05 - 1	COMP-4	33	-0.0261	-7470A	PPB		28 Mar 2022 12:45:06	U	SMPL
N2133-06 - 1	1933	48	-0.0220	-7470A	PPB		28 Mar 2022 12:47:22	U	SMPL
CCV37 - 1	CCV37	19473	5.3549	-7470A	PPB		28 Mar 2022 12:49:38	U	SMPL
CCB37 - 1	CCB37	-47	-0.0483	-7470A	PPB		28 Mar 2022 12:51:55	U	SMPL
N2133-06DUP - 1	1933DUP	112	-0.0043	-7470A	PPB		28 Mar 2022 12:54:11	U	SMPL
N2133-06MS - 1	1933MS	14466	3.9690	-7470A	PPB		28 Mar 2022 12:59:12	U	SMPL
N2133-06MSD - 1	1933MSD	15062	4.1339	-7470A	PPB		28 Mar 2022 13:01:28	U	SMPL
N2133-07 - 1	1932	48	-0.0220	-7470A	PPB		28 Mar 2022 13:03:45	U	SMPL
PB143558TB - 1	PB143558TB	40	-0.0242	-7470A	PPB		28 Mar 2022 13:06:02	U	SMPL
PB143658TB - 1	PB143632TB	28	-0.0275	-7470A	PPB		28 Mar 2022 13:08:19	U	SMPL
PB143634TB - 1	PB143634TB	70	-0.0159	-7470A	PPB		28 Mar 2022 13:10:35	U	SMPL
PB143641TB - 1	PB143641TB	82	-0.0126	-7470A	PPB		28 Mar 2022 13:12:53	U	SMPL
CCV38 - 1	CCV38	19358	5.3231	-7470A	PPB		28 Mar 2022 13:18:57	U	SMPL
CCB38 - 1	CCB38	-49	-0.0488	-7470A	PPB		28 Mar 2022 13:21:13	U	SMPL
N2085-01LX5 - 1		-3	-0.0361	-7470A	PPB		28 Mar 2022 13:23:31	U	SMPL
N2085-01A - 1		15788	4.3349	-7470A	PPB		28 Mar 2022 13:25:47	U	SMPL
N2091-02LX5 - 1		-46	-0.0480	-7470A	PPB		28 Mar 2022 13:28:04	U	SMPL
N2091-02A - 1		15733	4.3197	-7470A	PPB		28 Mar 2022 13:30:21	U	SMPL
N2123-03LX5 - 1		263	0.0375	-7470A	PPB		28 Mar 2022 13:32:37	U	SMPL
N2123-03A - 1		15444	4.2397	-7470A	PPB		28 Mar 2022 13:34:54	U	SMPL
N2140-02LX5 - 1		0	-0.0353	-7470A	PPB		28 Mar 2022 13:37:11	U	SMPL
N2140-02A - 1		15241	4.1835	-7470A	PPB		28 Mar 2022 13:39:28	U	SMPL
N2133-06LX5 - 1		-37	-0.0455	-7470A	PPB		28 Mar 2022 13:41:45	U	SMPL
N2133-06A - 1		15473	4.2477	-7470A	PPB		28 Mar 2022 13:44:02	U	SMPL
CCV39 - 1	CCV39	18431	5.0665	-7470A	PPB		28 Mar 2022 13:46:20	U	SMPL
CCB39 - 1	CCB39	-47	-0.0483	-7470A	PPB		28 Mar 2022 13:48:36	U	SMPL