



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: apatel

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB119377

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
SODIUM THIOSULFATE, 0.025N, 4LITRE (ICV)	W2843

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.9	03/28/2022	08:45
2	N2076-01		1	9	100.00	>212.0	>212.0	03/28/2022	09:15
3	N2085-01		1	8	100.00	>212.0	>212.0	03/28/2022	09:31
4	N2092-01		1	8	100.00	>212.0	>212.0	03/28/2022	09:48
5	N2099-01		1	9	100.00	>212.0	>212.0	03/28/2022	10:05
6	N2104-01		1	10	100.00	>212.0	>212.0	03/28/2022	10:20
7	N2106-12		1	10	100.00	>212.0	>212.0	03/28/2022	10:38
8	N2133-05		1	9	100.00	>212.0	>212.0	03/28/2022	10:54
9	N2133-05DUP		1	9	100.00	>212.0	>212.0	03/28/2022	11:12
10	N2133-06		1	10	100.00	>212.0	>212.0	03/28/2022	11:29
11	N2133-07		1	9	100.00	>212.0	>212.0	03/28/2022	11:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP-032822 WorkList ID : 158126 Department : Wet-Chemistry Date : 03-28-2022 08:07:04

Due Date	Matrix	Sample	Test	Raw Sample			Customer Sample	Collect Date	Method
				Preservative	Customer	Storage Location			
03/30/2022	Water	N2076-01	Flash Point	Cool 4 deg C	PARS02	E12	WC-20220322	03/22/2022	1010B
03/28/2022	Water	N2085-01	Flash Point	Cool 4 deg C	PSEG03	E12	GW-3-23-2022	03/23/2022	1010B
03/31/2022	Water	N2092-01	Flash Point	Cool 4 deg C	PSEG03	E12	FMI120	03/24/2022	1010B
03/31/2022	Water	N2099-01	Flash Point	Cool 4 deg C	PSEG03	E12	T-2	03/24/2022	1010B
03/28/2022	Water	N2104-01	Flash Point	Cool 4 deg C	PSEG03	E12	FRAC-TANK-A229	03/24/2022	1010B
03/31/2022	Water	N2106-12	Flash Point	Cool 4 deg C	PSEG03	E21	OILY-WATER-DRUM-1-2	03/24/2022	1010B
04/01/2022	Water	N2133-05	Flash Point	Cool 4 deg C	PSEG03	E21	COMP-4	03/25/2022	1010B
04/01/2022	Water	N2133-06	Flash Point	Cool 4 deg C	PSEG03	E21	1933	03/25/2022	1010B
04/01/2022	Water	N2133-07	Flash Point	Cool 4 deg C	PSEG03	E21	1932	03/25/2022	1010B

Date/Time 3/28/22 8:30
Raw Sample Received by: AL (VOL)
Raw Sample Relinquished by: CR SM

Date/Time 3/28/22 12:15
Raw Sample Received by: CR SM
Raw Sample Relinquished by: AL (VOL)