



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB119383

Slope : 98.7

BalanceID: WC SC-6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	03/28/2022	14:40
2	CAL2	1	Water	NA	NA	20.3	7.00	03/28/2022	14:41
3	CAL3	1	Water	NA	NA	20.3	10.03	03/28/2022	14:42
4	ICV	1	Water	NA	NA	20.4	6.98	03/28/2022	14:45
5	CCV1	1	Water	NA	NA	20.4	2.01	03/28/2022	14:46
6	N2155-01	1	Solid	20.03	20	22.5	11.97	03/28/2022	14:50
7	N2155-02	1	Solid	20.04	20	22.3	11.02	03/28/2022	15:00
8	N2155-03	1	Solid	20.03	20	22.1	11.83	03/28/2022	15:01
9	N2155-04	1	Solid	20.04	20	21.8	11.93	03/28/2022	15:05
10	N2155-05	1	Solid	20.02	20	22.1	11.74	03/28/2022	15:07
11	N2155-06	1	Solid	20.01	20	21.6	12.01	03/28/2022	15:08
12	N2155-07	1	Solid	20.03	20	21.4	12.03	03/28/2022	15:15
13	N2155-08	1	Solid	20.04	20	21.3	12.07	03/28/2022	15:19
14	N2155-09	1	Solid	20.02	20	21.2	12.16	03/28/2022	15:22
15	N2155-09DUP	1	Solid	20.03	20	21.3	12.17	03/28/2022	15:23
16	CCV2	1	Water	NA	NA	20.4	12.03	03/28/2022	15:25

119383

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH N2155 WorkList ID : 158137 Department : Wet-Chemistry Date : 03-28-2022 14:33:41

Due Date	Matrix	Sample	Test	Raw Sample			Customer Sample	Collect Date	Method
				Preservative	Customer	Storage Location			
03/28/2022	Solid	N2155-01	pH	Cool 4 deg C	ENV181	N21	22L076-A	03/28/2022	9045D
03/28/2022	Solid	N2155-02	pH	Cool 4 deg C	ENV181	N21	22L076-B	03/28/2022	9045D
03/28/2022	Solid	N2155-03	pH	Cool 4 deg C	ENV181	N21	22L076-C	03/28/2022	9045D
03/28/2022	Solid	N2155-04	pH	Cool 4 deg C	ENV181	N21	22L076-D	03/28/2022	9045D
03/28/2022	Solid	N2155-05	pH	Cool 4 deg C	ENV181	N21	22L076-E	03/28/2022	9045D
03/28/2022	Solid	N2155-06	pH	Cool 4 deg C	ENV181	N21	22L076-F	03/28/2022	9045D
03/28/2022	Solid	N2155-07	pH	Cool 4 deg C	ENV181	N21	22L076-G	03/28/2022	9045D
03/28/2022	Solid	N2155-08	pH	Cool 4 deg C	ENV181	N21	22L076-H	03/28/2022	9045D
03/28/2022	Solid	N2155-09	pH	Cool 4 deg C	ENV181	N21	22L076-I	03/28/2022	9045D

Date/Time 03/28/22 14:35
 Raw Sample Received by: JH (WC)
 Raw Sample Relinquished by: JH (SM)

Date/Time 03/28/22 15:30
 Raw Sample Received by: JH (SM)
 Raw Sample Relinquished by: JH (WC)