



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB119389

Slope : 98.7

BalanceID: WC SC-6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2886
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.5	4.01	03/28/2022	17:00
2	CAL2	1	Water	NA	NA	20.5	7.01	03/28/2022	17:01
3	CAL3	1	Water	NA	NA	20.5	10.03	03/28/2022	17:03
4	ICV	1	Water	NA	NA	20.4	6.99	03/28/2022	17:05
5	CCV1	1	Water	NA	NA	20.4	2.01	03/28/2022	17:06
6	N2157-02	1	Solid	20.03	20	20.3	9.04	03/28/2022	17:10
7	N2157-05	1	Solid	20.04	20	20.5	8.93	03/28/2022	17:12
8	N2157-05DUP	1	Solid	20.03	20	20.6	8.94	03/28/2022	17:13
9	CCV2	1	Water	NA	NA	20.4	12.03	03/28/2022	17:15

VB119389

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY N2157 WorkList ID : 158143 Department : Wet-Chemistry Date : 03-28-2022 16:19:31

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample		Collect Date	Method
						Storage	Location		
03/28/2022	Solid	N2157-02	Corrosivity	Cool 4 deg C	PSEG03	N31	MHD	03/28/2022	9045D
03/28/2022	Solid	N2157-05	Corrosivity	Cool 4 deg C	PSEG03	N31	TP2	03/28/2022	9045D

Date/Time 03/28/22 16:20
Raw Sample Received by: JP (QPC)
Raw Sample Relinquished by: L-C (Sow)

Date/Time 03/28/22 17:30
Raw Sample Received by: L-C (Sow)
Raw Sample Relinquished by: JP (QPC)