



Analytical Summary Report

Analysis Method: SM2510 B

ANALYST: Iwona

Parameter: Conductance

SUPERVISOR REVIEW BY: apatel

Run Number: LB119392

Meter ID: WC CONDUCTANCE METER-1

Thermometer ID: WC PH METER-1

Reagent/Standard	Lot/Log #
conductivity standard	WP95377

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB119392BL		1	21.8	0.68	0.72	03/29/2022	09:30
2	LB119392BS	1413	1	21.9	1347.00	1431.78	03/29/2022	09:35
3	N1873-07		1	22.0	939.00	996.08	03/29/2022	09:40

Note : Conductivity result is corrected to 25°C

Result = $\frac{\text{(Reading)}}{1 + 0.0191 (\text{Temperature} - 25)}$